



HAL
open science

Computational modeling of radiofrequency ablation for the planning and guidance of abdominal tumor treatment

Chloé Audigier

► **To cite this version:**

Chloé Audigier. Computational modeling of radiofrequency ablation for the planning and guidance of abdominal tumor treatment. Other. Université Nice Sophia Antipolis, 2015. English. NNT : 2015NICE4071 . tel-01256010

HAL Id: tel-01256010

<https://theses.hal.science/tel-01256010>

Submitted on 14 Jan 2016

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

UNIVERSITÉ DE NICE - SOPHIA ANTIPOLIS
ÉCOLE DOCTORALE STIC
SCIENCES ET TECHNOLOGIES DE L'INFORMATION
ET DE LA COMMUNICATION

THÈSE

pour obtenir le titre de

Docteur en Sciences

de l'Université de Nice - Sophia Antipolis

Mention : **AUTOMATIQUE, TRAITEMENT DU SIGNAL ET
DES IMAGES**

Présentée et soutenue par

Chloé AUDIGIER

Modélisation de l'ablation radiofréquence pour la planification de la résection de tumeurs abdominales

Thèse dirigée par Hervé DELINGETTE et Nicholas AYACHE

préparée à Inria Sophia Antipolis, projet ASCLEPIOS et à SIEMENS Corporate Research, Princeton

soutenue le 14 Octobre 2015

Jury :

<i>Rapporteurs :</i>	Sir Michael BRADY	-	University of Oxford
	David HAWKES	-	University College London
<i>Directeur :</i>	Hervé DELINGETTE	-	Inria (Asclepios)
<i>Co-Directeur :</i>	Nicholas AYACHE	-	Inria (Asclepios)
<i>Président :</i>	Stéphane COTIN	-	Inria (Mimesis)
<i>Examineurs :</i>	Tommaso MANSI	-	Siemens Corporate Research (Princeton)
	Michele DIANA	-	IHU (Strasbourg)
<i>Invités :</i>	Luc SOLER	-	IRCAD (Strasbourg)
	Dorin COMANICIU	-	Siemens Corporate Research (Princeton)

UNIVERSITY OF NICE - SOPHIA ANTIPOLIS
DOCTORAL SCHOOL STIC
INFORMATION AND COMMUNICATION TECHNOLOGIES AND SCIENCES

PHD THESIS

to obtain the title of

PhD of Science

of the University of Nice - Sophia Antipolis

Specialty : CONTROL, SIGNAL AND IMAGE PROCESSING

Defended by

Chloé AUDIGIER

Computational Modeling of Radiofrequency Ablation for the Planning and Guidance of Abdominal Tumor Treatment

Thesis Advisor: Hervé DELINGETTE and Nicholas AYACHE

prepared at Inria Sophia Antipolis, ASCLEPIOS Team and at Siemens Corporate Research,
Princeton

defended on October 14, 2015

Jury :

<i>Reviewers :</i>	Sir Michael BRADY	-	University of Oxford
	David HAWKES	-	University College London
<i>Advisor :</i>	Hervé DELINGETTE	-	Inria (Asclepios)
<i>Co-Advisor :</i>	Nicholas AYACHE	-	Inria (Asclepios)
<i>President :</i>	Stéphane COTIN	-	Inria (Mimesis)
<i>Examinators :</i>	Tommaso MANSI	-	Siemens Corporate Research (Princeton)
	Michele DIANA	-	IHU (Strasbourg)
<i>Invited :</i>	Dorin COMANICIU	-	Siemens Corporate Research (Princeton)
	Luc SOLER	-	IRCAD (Strasbourg)

Résumé

L'ablation par radiofréquence (ARF) de tumeurs abdominales est rendue difficile par l'absence de planification spécifique à un patient donné. En particulier, l'influence des vaisseaux sanguins et les variations spatiales et temporelles de la conductivité thermique compliquent la prédiction de l'étendue de l'ablation. Cela peut se traduire par des traitements incomplets et l'augmentation des risques de récurrences compromettant le résultat final. En fournissant des outils prédictifs, les modèles biophysiques peuvent aider les cliniciens à planifier et guider efficacement la procédure. Pour que ces modèles soient utiles en routine clinique, le calcul doit être rapide et précis.

Nous présentons d'abord un modèle mathématique détaillé des mécanismes biophysiques impliqués dans l'ARF des tumeurs hépatiques tels que la diffusion de la chaleur, la nécrose cellulaire et le flux sanguin hépatique. Ce modèle simule l'étendue de l'ablation à partir de l'emplacement de la sonde et de certains paramètres biologiques tels que la conductivité ou la capacité calorifique des tissus. Le modèle est basé sur des images médicales, à partir desquelles des modèles personnalisés du foie, des vaisseaux visibles et des tumeurs sont reconstruits. Une nouvelle approche pour résoudre ces équations aux dérivées partielles basée sur la méthode de Lattice Boltzmann est introduite. Le modèle est mis en oeuvre avec une accélération matérielle liée à la carte graphique qui permet d'atteindre un temps de calcul presque égal au temps réel. Les simulations obtenues sont vérifiées précisément avec une solution analytique.

Dans un deuxième temps, nous voulons valider le modèle. Il est d'abord évalué sur des données cliniques de patients qui ont subi une ARF de tumeurs du foie. L'importance de la prise en compte du flux sanguin et de l'estimation des paramètres sont mises en évidence. Ensuite, un protocole expérimental complet combinant des images multi-modales, anatomiques et fonctionnelles, pré et post-opératoires, ainsi que le suivi de la température et de la puissance délivrée pendant l'intervention est présenté. Il permet une validation pré-clinique totale qui considère des données les plus complètes possible pour la validation du modèle.

Enfin, nous estimons automatiquement des paramètres personnalisés en utilisant des algorithmes de résolution de problèmes inverses, afin de mieux prédire l'ablation. Cette stratégie de personnalisation a été testée et validée sur sept ablations dans trois cas cliniques. A partir de l'étude pré-clinique, nous pouvons aller plus loin dans la personnalisation en comparant la température et la puissance délivrée simulées avec la température et la puissance délivrée mesurées pendant la procédure. Cette seconde stratégie de personnalisation a été testée et validée sur douze ablations dans cinq cas pré-cliniques. Ces contributions ont abouti à des résultats prometteurs, et ouvrent de nouvelles perspectives pour planifier et guider l'ARF.

MOTS CLÉS: Modélisation d'ARF; Foie; Personnalisation; Diffusion de la Chaleur; Nécrose Cellulaire; Mécanique des Fluides; Modèle Informatique; Méthode de Lattice Boltzmann; Estimation de Paramètres; Etude Pré-Clinique; Imagerie Médicale

Abstract

The outcome of radiofrequency ablation (RFA) treatment for abdominal tumors is challenged by lack of patient-specific planning. In particular, the presence of blood vessels, time- and spatial-varying thermal conductivity make the prediction of the extent of ablated tissue difficult. This results in possible incomplete treatments and increased risks of recurrence and may compromise the overall outcome. By providing predictive tools, biophysical models can help clinicians to plan and guide the procedure for an effective treatment. For such models to be useful in clinical routine, the computation has to be fast and accurate.

We first present a detailed computational model of the biophysical mechanisms involved in RFA of hepatic tumors such as heat diffusion, cellular necrosis and hepatic blood flow. This model simulates the extent of ablated tissue given the probe location and some biological parameters such as tissue conductivity or heat capacity. The model is based on medical images, from which patient-specific models of the liver, visible vessels and tumors are built. A new approach for solving these partial differential equations based upon the Lattice Boltzmann Method is introduced. The model is implemented with a material speed-up thanks to the graphics card to reach near real-time computation. The resulting simulations are thoroughly verified against an analytical solution.

In a second step, we aim at validating the model. It is first evaluated against clinical data of patients who underwent RFA of liver tumors. The importance of taking into account the blood flow and of estimating the parameters is highlighted. Then, a comprehensive pre-clinical experiment combining multi-modal, pre- and post-operative anatomical and functional images, as well as the interventional monitoring of the temperature and delivered power is presented. This enables an end-to-end pre-clinical evaluation framework that considers the most comprehensive data set for model validation.

Then, we automatically estimate patient-specific parameters using inverse problem algorithms, to better predict the ablated tissue. This personalization strategy has been tested and evaluated on seven ablations from three clinical cases. From the pre-clinical study, we can go further in the personalization by comparing the simulated temperature and delivered power with the actual temperature and delivered power measured during the procedure. This second personalization strategy has been tested and evaluated on twelve ablations from five pre-clinical cases. These contributions have led to promising results, and open new perspectives in RFA guidance and planning.

KEYWORDS: RFA Modeling; Liver; Patient-Specific; Heat Transfer; Cellular Necrosis; Computational Fluid Dynamics; Computer Model; Lattice Boltzmann Method; Parameter Estimation; Pre-clinical Study; Medical Imaging

Acknowledgments

First of all, I would like to warmly thank my advisors, Tommaso Mansi and Hervé Delingette for their guidance and support all along my Ph.D. period. You were always there to answer my questions and to help me during my research work. I have learned a lot from your deep scientific knowledge. Without your availability and your generosity, this work would not have been possible. Tommaso, it was a pleasure to work with you; your enthusiasm, your passion for the field were highly communicative. I am happy that you motivated me, pushed me, and taught me to never give up even when nothing is working. Hervé, thank you for your continuous support, your scientific insight and your valuable advice. I really appreciated working with you, and having those long discussions about the model. I would also like to thank Nicholas Ayache for accepting me in the Asclepios team. I appreciated your great faculty to present scientific results, and your vision of the future. Thank you for creating a friendly research environment.

I am extremely grateful to Prof. Sir. Michael Brady and Prof. David Hawkes for accepting to be my reviewers and for spending their precious time to read and correct this manuscript. I warmly thank them for their sharp and constructive comments on my work and for their encouraging compliments. I am also thankful to Michele Diana and Stéphane Cotin, who accepted to be member of my jury and to come at my defense. Dear committee, thank you, it has been a great honor for me to have such an outstanding jury.

I also thank Dorin Comaniciu for giving me the chance to do three internships in his prestigious team. Tommaso, thank you for organizing my stays three times, it is also fun to work close to you. I thank my colleagues in Princeton who made my three internships wonderful. I thank Viorel Mihalef, Saikaran Rapaka, Tiziano Passerini for their help about the CFD solver, Marie-Pierre Joly for her help about the analysis of the Phase-Contrast MRI, Dominik Neumann for the support during the nights before the MICCAI deadlines, Puneet Sharma and Ali Kamen, it was a great honor to work with you. And of course, I thank all the Siemens interns I met from all around the world, especially Tanguy Abel for helping me to process the pre-clinical data. Sorry guys, I cannot give all your names here, but I am thinking about each of you from the years 2012, 2013, 2014 and 2015.

My sincere thanks to all my Asclepios colleagues. Without such a great team, this Ph. D. would not have been such a pleasure. I thank Matthieu Le and Nicolas Cordier, it was great to share the office with you, thank you for supporting me ! I thank Federico Spadoni, Brina Goyette and Stéphanie Marchesseau for answering all my questions about Sofa. I warmly thank Vikash Gupta for all his deep thoughts and his Gaussian theory about life. I thank Marco Lorenzi, Christoph Seiler, Marine Breuilly, Alan Garny, Adityo Prakosa, Ezequiel Germia for the warm welcome when I arrived. I thank Ján Margeta, Rocio Cabrera Lozoya, Thomas Demarcy, Marc-Michel Rohe, Sophie Giffard for the good time we have spent together in and also outside the lab. Thanks Bisheh Khanal for inviting us to Nepal, and thanks to Irina Vidal, Kristin McLeod, Florian Vichot and Hakim Fadil for deciding within few days

to join for such an adventure ! They did not spend enough time with us, but I want to thank Hugo Talbot, Thomas Benseghir, Pietro Gori, Aziz Ouerdani and Florent Collot for the fun time we had especially in Auron. I will not forget the engineers, who were always there to answer all my questions about CUDA or Windows, thanks Loic Cadour and Merzieh Kohandani Tafreshi. I also thank Mehdi Hadj-Hamou, Nina Miolane, Roch Mollero, Nicolas Duchateau, Aurelie Degletane, Heloise Bleton, Raphael Sivera, Mike Buckingham, Hervé Lombaert, Maxime Sermesant and Xavier Pennec. Thank you for the great atmosphere in the Asclepios team! And of course, I thank Isabelle Strobant, for her support, her patience and her kind help when organizing my different trips.

I thank the IHU of Strasbourg for allowing me to set up the SimulAB project. I thank Luc Soler for his advice, without him this project would not have been possible. I thank Frank Blindauer, Gael Fourré, Mourad Bouhadjar and Rodrigo Cararo for their warm welcome in Strasbourg. I thank Raoul Pop and Michele Diana for their clinical expertise, for taking the time to answer all my questions about the liver, hepatic tumor and RFA. Finally, I thank the 5 pigs, who helped us to understand a bit more RFA of liver tumors.

At last but not least, I thank my family and my friends who always believed in me. I thank my mother and my father for their endless support and understanding during my long years of education, my two brothers for telling me the same joke over and over again ¹. And of course, I thank Camille, for being with me, supporting me even when it was not easy.

MERCI !!

¹Des chercheurs qui cherchent, on en trouve; mais des chercheurs qui trouvent, on en cherche.
Charles de Gaulle

Contents

1	Introduction	1
1.1	Motivation	1
1.1.1	Clinical Context	1
1.1.2	Role of Medical Imaging	2
1.1.3	Objectives	3
1.2	Main Contributions and Manuscript Organization	3
1.2.1	Main Contributions	3
1.2.2	Organization of the Thesis	4
2	State of the Art on RFA of Hepatic Tumors	9
2.1	Structure and Function of the Liver	10
2.2	Radiofrequency Ablation	11
2.2.1	Radiofrequency Ablation Procedure	11
2.2.2	Monitoring of RFA based on MR Thermometry	13
2.3	Percutaneous Tumor Ablation Techniques	14
2.3.1	High Intensity Focused Ultrasound	15
2.3.2	Microwave Ablation	15
2.3.3	Cryotherapy	15
2.3.4	Laser Ablation	16
2.3.5	Electroporation	16
2.4	Technical Background	17
2.4.1	Models of Electrical Heating	17
2.4.2	Biophysical Models of Heat Transfer in Tissue	18
2.4.3	Models of Cellular Necrosis	20
2.5	State of the Art of Existing Radiofrequency Ablation Models	22
2.5.1	Computational Models of RFA	22
2.5.2	Model Simplification	23
2.5.3	Model Validation	24
2.5.4	Model Personalization	25
2.5.5	Surrogate Model for Optimization of the Probe Placement	26
2.6	MR Thermometry	27
2.6.1	Thermal Tissue Properties Acquisition during HIFU based on MR Thermometry	27
2.6.2	MR Thermometry Filtered by Bioheat Equation	27
3	Modeling of Radiofrequency Ablation	29
3.1	Introduction	29
3.2	Method	30
3.2.1	Model of the Patient Hepatic Blood Circulation System	30

3.2.2	Fully Coupled Model of Heat Transfer in Liver Tissue and Cellular Necrosis	38
3.2.3	Estimation of Anatomical Model from Patient Images	45
3.3	Heating Power and Cooling Temperature Computation	46
3.3.1	Heating Stage	46
3.3.2	Cooling Stage	47
3.4	Model Verification	47
3.4.1	Quantitative Verification of the Heat Transfer Model	47
3.4.2	Computational Efficiency	49
3.5	Discussion	50
3.5.1	Summary	50
3.5.2	Model Limitations	50
3.6	Conclusion	51
4	Evaluation on Clinical Data	53
4.1	Introduction	53
4.2	Available Clinical Data	54
4.3	Patient-Specific RFA Computation	55
4.3.1	Experimental Protocol	55
4.3.2	Computation Time	55
4.4	Evaluation on Patient Data	56
4.4.1	Systematic Study	56
4.4.2	Effect of Advection	58
4.4.3	Effect of the Probe Position	60
4.4.4	Effect of the Segmentation	61
4.4.5	Effect of Biophysical Parameters	63
4.5	Discussion	65
4.5.1	Effect of Large Vessels	65
4.5.2	Effect of Advection	65
4.5.3	Effect of the Probe Position	66
4.5.4	Effect of the Segmentation	66
4.5.5	Towards Personalization	66
4.6	Conclusion	67
5	Parameter Estimation For Personalization of RFA	69
5.1	Introduction	69
5.2	Available Clinical Data	70
5.3	Mathematical Model of RFA Simulation	70
5.3.1	Forward Model	70
5.3.2	Parameter Estimation	71
5.4	Experiments and Results	72
5.4.1	Evaluation of the CFD Solver on a Synthetic Case	72
5.4.2	Sensitivity Analysis	73
5.4.3	Verification of the Optimization Framework	74

5.4.4	Evaluation on Patient Data	74
5.5	Discussion	79
6	Pre-Clinical Study: Protocol and Data Analysis	81
6.1	Introduction	81
6.2	Available Pre-clinical Data	82
6.3	Pre-clinical Experiment	82
6.3.1	A Pre-clinical Study for RFA Model Validation	82
6.3.2	Experimental Set-up	84
6.4	Computational Experiment	85
6.4.1	Overview	85
6.4.2	RFA Modeling	86
6.4.3	Numerical Computation using LBM	88
6.4.4	Parameter Estimation from Probe Measurements	88
6.5	Data Pre-processing	90
6.5.1	3D Modeling of the RFA Probe	91
6.5.2	Segmentation	92
6.5.3	CFD Computational Domain	92
6.5.4	Post-To-Pre Registration	95
6.6	Conclusion	96
7	Pre-Clinical Study: Results and Discussion	99
7.1	Introduction	100
7.2	Evaluation of the Pre-processing	101
7.2.1	Evaluation of the Registration	101
7.2.2	Evaluation of the Smooth Vessels Trees	103
7.3	Verification of the Parameter Estimation Framework	104
7.3.1	Synthetic Data Generation	104
7.3.2	Evaluation	105
7.4	Sensitivity Analysis of the Model Parameters	106
7.5	Evaluation on Synthetic Data	108
7.6	Evaluation on Swine Data	110
7.6.1	Systematic Study	110
7.6.2	Towards Personalization	116
7.7	Sensitivity Analysis on Swine Data	119
7.7.1	Effect of the Advection	119
7.7.2	Effect of the Registration	122
7.7.3	Effect of the Probe Position	123
7.7.4	Cell Death Model Evaluation	125
7.8	Discussion	127
7.8.1	Model Limitations	127
7.8.2	Effect of the Registration	128
7.8.3	Towards Personalization	128
7.9	Conclusion	129

8	Summary and Perspectives	137
8.1	Main Contributions	137
8.1.1	Modeling of Radiofrequency Ablation with LBM	137
8.1.2	Model Evaluation on Clinical Data	138
8.1.3	Towards Model Personalization	138
8.1.4	Pre-clinical Model Evaluation	139
8.2	Perspectives	140
8.2.1	Model Improvements	140
8.2.2	Model Validation	140
8.2.3	Model Personalization	141
8.3	Conclusion	141
A	The Lattice Boltzmann Method	147
A.1	The Boltzmann Equation	147
A.1.1	Boltzmann Transport Equation	147
A.1.2	The BGKW Approximation	148
A.1.3	Equilibrium Distribution Function	149
A.2	Advection-Diffusion Equation	151
A.2.1	1D Formulation	151
A.2.2	3D Formulation	155
A.3	Implementation	159
A.4	Advection-Diffusion-Reaction Equation	160
A.5	Boundary Condition	164
	Bibliography	165

List of Figures

1.1	Diagram showing the chapter architecture.	4
2.1	Anatomy of the human liver	10
2.2	Diagram showing the RF current through the patient body	11
2.3	Heating with RFA probe	12
2.4	RFA-induced lesion	13
2.5	A CT slice with RFA probe	14
2.6	Modeling steps of RFA modeling	17
2.7	RITA probe representation	18
3.1	Pipeline of the modeling method	31
3.2	Pipeline of the modeling method using the Decoupled FEM Porous-Vessels CFD Model	32
3.3	Hepatic circulatory system modeling	33
3.4	Pipeline of the modeling method using the Fully Coupled LBM CFD Model.	34
3.5	Hepatic blood flow computation.	35
3.6	D3Q19 scheme for LBM.	36
3.7	(<i>Left</i>): Blood flow computation with the Decoupled FEM Porous-Vessels CFD Model; (<i>Right</i>): Blood flow computation with the Fully Coupled LBM CFD Model. In both cases, only the flow going through the portal inlet circled in white are shown. We can see differences between the results given by both computations. The area circled in black shows flow lines, which appear with the first computation and not the second one.	38
3.8	Cell state evolution over time.	41
3.9	D3Q7 scheme used in LBM.	42
3.10	Boundary conditions used in LBM.	44
3.11	Sphere modeling the RFA Probe.	45
3.12	Anatomical model of the liver.	46
3.13	Synthetic set-up for the verification against analytical solution.	47
3.14	Spatial and temporal convergence analysis	48
3.15	Computational speed-up time for LBM implementation	50
4.1	Pre and Post-operative images available for patient 04.	54
4.2	Results of registration on patient 10	56
4.3	Results of RFA simulation on patient 04	56
4.4	Box plot of the error on the 14 ablations.	58
4.5	Results of RFA simulation on patient 01	58
4.6	Box plot of the error on the 14 ablations without the advection.	59

4.7	Two hypotheses for the probe position on patient 06	60
4.8	Evaluation of the effect of the RFA probe.	60
4.9	Effect of the probe position on patient 06	61
4.10	Evaluation of the effect of the segmentation.	62
4.11	Two segmentation results of patient 03	62
4.12	Effect of the segmentation on RFA simulation for patient 03	62
4.13	Effect of the diffusivity parameter on the error for patient 02	63
4.14	Parameter tuning on patient 02	64
5.1	Parameter estimation pipeline.	71
5.2	Synthetic case for CFD solver validation.	72
5.3	Synthetic set-up for the verification of the optimization framework.	73
5.4	Convergence analysis of the optimization framework.	75
5.5	Results of the parameter estimation on the computed lesion.	76
5.6	Box plot of the error on the 7 ablations with personalized parameters.	77
5.7	Parameters personalized on three patients.	77
5.8	Comparison of personalized computation with post-operative registered ground truth for patient 1-1	78
5.9	Comparison of personalized computation with post-operative registered ground truth for patient 1-2	79
5.10	Comparison of personalized computation with post-operative registered ground truth for patient 1-3	80
6.1	Pre-operative CT and Post-operative MRI available for Fig 5.	82
6.2	Pipeline presenting the available pre-clinical data.	83
6.3	Surrogate tumor implantation.	84
6.4	RFA probe inside the liver and on a U.S image.	85
6.5	CFD Model.	87
6.6	Cell state evolution with the adapted cellular necrosis model.	88
6.7	Comparison of the 12 delivered power curves.	89
6.8	Delivered power over ablation : the RFA protocol is not exactly the same for all pigs.	90
6.9	Probe: mesh model used and interventional MRI data.	91
6.10	Segmentations.	92
6.11	Smooth vessel segmentation.	93
6.12	Velocity from Phase-Contrast MRI.	94
6.13	Porosity map used for Fig 1.	95
6.14	Evaluation of the Post-to-Pre registration.	96
7.1	Results of the Post-to-Pre registration on the liver of Fig 1 liver.	101
7.2	Post-to-Pre registration of arterial mesh of Fig 2.	102
7.3	Post-to-Pre registration of liver mesh of Fig 2.	103
7.4	Evaluation of the vessels smoothing of Fig 4	104
7.5	Synthetic case for the parameter estimation.	105

7.6	Computed delivered power and temperature on the synthetic case.	105
7.7	Convergence curves of 2 optimization algorithms.	106
7.8	Sensitivity analysis on the heat capacity of undamaged cells c_t^U	107
7.9	Sensitivity analysis on the heat capacity of vulnerable cells c_t^V	107
7.10	Sensitivity analysis on the heat capacity of necrotic cells c_t^N	108
7.11	Sensitivity analysis on the heat conductivity \bar{d}_t	108
7.12	Evaluation on synthetic data.	109
7.13	Computed delivered power and temperature for tumor 4-3.	110
7.14	Zoom on the cooling phase of the temperature for tumor 4-3.	111
7.15	Simulated necrosis versus ground truth for tumor 4-2.	111
7.16	Box plot of the error on the 12 ablations from 5 pigs.	112
7.17	Box plot of the similarity scores computed on the volume of the 12 ablations.	113
7.18	Box plot of the error in term of delivered power and temperature on the 12 ablations.	113
7.19	Simulated delivered power and temperature for tumor 3-1 with non- inial parameters and with personalized parameters.	114
7.20	Simulated delivered power and temperature for tumor 3-1 with and without the advection term.	115
7.21	Parameters personalized on two tumors from two different pigs.	117
7.22	Personalization of the delivered power and temperature curves for tumor 1.	118
7.23	Zoom on the cooling temperature curves for tumor 1.	118
7.24	Effect of the advection on the simulated necrosis versus ground truth for Fig 1.	119
7.25	Effect of the location of the ablation on the temperature dispersion.	120
7.26	Effect of the advection on the delivered power and the temperature curves for tumor 1.	121
7.27	Zoom on the cooling temperature curves for tumor 1 computed with the advection.	121
7.28	Effect of the registration on the simulated lesion on Fig 2.	123
7.29	Simulated lesion versus ground truth for tumor 3-2.	124
7.30	Cell death model versus thermal dose computation.	126
7.31	The simulated necrotic extent around tumor 3-3 is subcapsular.	128
7.32	MR thermometry acquired during the ablation of tumor 4-1.	130

List of Tables

2.1	Nominal Values of the model parameters.	19
3.1	RMS errors against the analytical solution for different spatial resolution.	49
4.1	Evaluation of the RFA modeling on patient data.	57
4.2	Evaluation of the RFA modeling on patient data without advection.	59
4.3	Evaluation of the effect of the RFA probe.	60
4.4	Evaluation of the effect of the segmentation.	61
4.5	Estimation of the heat conductivity.	64
5.1	Ranges of parameters values explored in the sensitivity analysis.	74
5.2	Total Sobol indices.	74
5.3	Evaluation of the parameter estimation framework on patient data.	76
5.4	Parameter estimation on patient 1-1.	78
5.5	Parameter estimation on patient 1-2 and patient 1-3.	79
6.1	Nominal parameter values.	87
7.1	Evaluation of the registration for Fig 2.	101
7.2	Evaluation of the smooth vessel trees.	103
7.3	Ranges of parameters values explored in the optimization framework.	106
7.4	Comparisons between personalized and nominal values of estimated parameters.	116
7.5	Modalities of the pre- and post-operative images used.	122
7.6	Evaluation of the effect of the registration on Fig 2-2.	123
7.7	Evaluation of the probe position on Fig 3-2.	125
7.8	Evaluation of the effect of the cell death model on Fig 2-2.	126
7.9	Evaluation of the effect of the cell death model on the shape of the simulated lesion for Fig 2-2.	127
7.10	Evaluation of the RFA modeling on Fig 1 and Fig 4.	131
7.11	Evaluation of the RFA modeling on Fig 2	132
7.12	Evaluation of the RFA modeling on Fig 3.	133
7.13	Evaluation of the RFA modeling on Fig 5.	134
7.14	Evaluation of the RFA modeling on the shape of the simulated lesion on Fig 1, Fig 2 and Fig 3.	135
7.15	Evaluation of the RFA modeling on the shape of the simulated lesion of Fig 4 and Fig 5.	136

Introduction

Contents

1.1 Motivation	1
1.1.1 Clinical Context	1
1.1.2 Role of Medical Imaging	2
1.1.3 Objectives	3
1.2 Main Contributions and Manuscript Organization	3
1.2.1 Main Contributions	3
1.2.2 Organization of the Thesis	4

1.1 Motivation

1.1.1 Clinical Context

The World Health Organization (WHO) estimated that 782,000 new liver cancer cases occurred worldwide in 2012, and that it is the second most common cause of death from cancer. The WHO estimated it to be responsible for nearly 745,000 deaths in 2012 that is to say 9.1% of the total cancer death worldwide [Ferlay 2015]. Hepatocellular carcinoma (HCC) for example is one of the most common malignancies encountered throughout the world (more than 1 million cases per year), with increasing frequency in Western countries due to the changing prevalence of hepatitis C [El-Serag 2003]. The prognosis for liver cancer is very poor (overall ratio of mortality to incidence of 0.95) [Ferlay 2015], since treatment of primary and metastatic malignancies in the liver remains a significant challenge in spite of recent advances in cancer therapy. Unfortunately, only less than 25% of patients with primary or secondary liver cancer are candidates for resection or transplantation, which are considered as the most effective treatments. These limitations are due to the patient's condition, the location, size and number of tumors. Proximity to vital organs like vascular and biliary structures is a problem. For example, many patients with cirrhosis or with multiple tumors have inadequate hepatic reserve to tolerate resection [Curley 1997]. Surgical resection are currently restricted to patient with solitary HCC, with liver function extremely well-preserved, without portal hypertension and without abnormal bilirubin. Whilst liver transplantation is limited by the number of donors [Bruix 2002]. Consequently, minimally invasive

ablative therapies such as radiofrequency ablation (RFA), high intensity focused ultrasound (HIFU), cryotherapy, microwave tumor coagulation and laser ablation have raised increasing interest for the treatment of liver tumors. Amongst them, RFA is the most widely used approach [Garrean 2008], but it is not yet a reliable alternative in clinical routine. There is a need for training tools for the less-experienced clinicians to improve outcome rates since the success rate increases quickly with experience [Hildebrand 2006]. For the same physician, survival rate of treated patients increased twofolds over a four years period. This learning curve is partly due to the difficult assessment of the cooling effect of the large vessels, porous circulation and blood coagulation, which results in suboptimal ablation and local recurrences in up to 60% of the cases [Kim 2006]. Nevertheless, RFA technology has increasingly been utilized for unresectable hepatic malignancies, is commonly used in the ablation of aberrant conduction pathways in the heart and has become a primary procedure in the treatment of cardiac arrhythmias [Bilchik 2001].

1.1.2 Role of Medical Imaging

Coagulation of living tissue by means of heat was mentioned around 3000 B.C. in the Edwin Smith papyrus, a copy of part of an ancient Egyptian textbook on trauma surgery [Rieder 2013]. It describes 8 cases of tumors or ulcers of the breast that were removed by cauterization with a tool called the fire drill. The writing says about the disease, "There is no treatment"¹. Nowadays, coagulation of living tissue by means of heat is used in clinical routine: radiofrequency ablation (RFA) is an established treatment for liver cancer when resection is not possible. Yet, incomplete treatment and an increased risk of recurrence are common. A tool that would enable the accurate planning of RFA is hence necessary.

To give additional guidance to radiologists or surgeons, several research groups are working on automatic tools to extract information from images and simulate the RFA behaviour in the liver parenchyma. Patient-specific anatomy is not enough since the optimal delivery of RFA is challenged by the presence of large blood vessels and the time-varying thermal conductivity of biological tissue. The effect of the blood flow, the temperature and the cell death on the tissue properties have to be accounted for, thus simulation is necessary. The idea is to combine anatomical data and the knowledge of biological mechanisms in order to build biophysical models that can reproduce the RFA in a patient-specific anatomy and therefore provide predictive tools. By simulating the outputs of the therapy *in silico*, these models might help clinicians to plan individual treatment for each patient since they would be able to study the effect of the probe position or the heat duration for example on the treatment output.

Theoretical models and computer simulations are powerful tools since they provide vital information on the thermal behavior of ablation rapidly and at low cost. They could provide efficient training tools for the less-experienced clinicians. They could also facilitate the feasibility of new probe geometries, and new RFA protocols.

¹www.cancer.org/the-history-of-cancer-pdf Retrieved 2015-07-15.

But, the two most important issues involved in the modeling of RFA include first the experimental validation of the proposed model and secondly the lack of an accurate characterization of the biological tissues [Berjano 2006].

1.1.3 Objectives

Within this scope, the objective of this work is to investigate new computerized predictive tools for patient-specific guidance and therapy planning of radiofrequency ablation (RFA) procedure. Patient-specific simulations may allow clinicians to better understand the intricacy of this procedure and could even allow to rehearse on a virtual patient-specific anatomy. These challenges require many improvements regarding the state of the art. The main questions, that we aim to answer in this thesis, are:

- How to simulate the radiofrequency ablation of hepatic tumors in computational time compatible with clinical settings while taking into account the main occurring biophysical phenomena ?
- Which parameters of this RFA model can be personalized and based on which data ?
- How to validate such RFA model on pre-clinical and clinical data ?

1.2 Main Contributions and Manuscript Organization

1.2.1 Main Contributions

During this PhD study, four main contributions to the modeling of RFA and its validation were proposed.

- A detailed computational model of the biophysical mechanisms (heat transfer, cellular necrosis, hepatic blood flow) involved in RFA of abdominal tumors based on patient images.
- A new implementation of the bio-heat equations coupled with a cellular necrosis model using the Lattice Boltzmann Method on Graphics Processing Units (GPU), which allows near real-time computation.
- The automatic estimation of the main parameters of the model. Two personalization strategies are tested and evaluated on clinical and pre-clinical data.
- The evaluation of the proposed model on clinical data as well as preclinical data from a comprehensive experimental set-up specially designed for RFA model validation.

1.2.2 Organization of the Thesis

The manuscript is organized following the published and submitted works. It follows rather well the studies performed during this thesis in chronological order: from the modeling of RFA to its main parameters personalization and finally to its evaluation on pre-clinical data. Figure 1.1 illustrates how the chapters are connected, and the available data used in each chapter.




Figure 1.1: A schematic diagram showing how the chapters are connected and the different available data used in each chapter.

This thesis focuses on hepatic tumors and the associated RF ablation procedures. **Chapter 2** introduces the background on hepatic anatomy and hepatic functions as well as details on radiofrequency ablation and other minimally invasive ablation techniques. Then, an introduction to the modeling of radiofrequency ablation establishes the state of the art, shows the different models and methods used in the literature, highlight their limits and those further used in our choice of RFA modeling.

Chapter 3 describes our first contribution, published in [Audigier 2013a] and further developed in [Audigier 2015a]. It consists of a new method to compute the extent of ablation based on the Lattice Boltzmann Method (LBM) and patient-specific, pre-operative images. To the best of our knowledge, this is the first time such a complete model, based on medical images has been proposed. A detailed anatomical model of the liver is obtained from volumetric images. Then a computational model of heat diffusion, cellular necrosis, and blood flow through the vessels and liver is employed to compute the extent of ablated tissue given the probe location, ablation duration and biological parameters. Furthermore, a second and improved patient-specific model of blood flow through liver and parenchyma is presented. We rely on the Lattice Boltzmann Method (LBM) to compute not only the heat transfer and cellular death mechanisms but also the blood and parenchyma flow in the liver tissue. This latter method is based on a Computational Fluid Dynamics (CFD) solver which incorporates a porous part to deal with the liver parenchyma. This second framework is particularly efficient for the modeling as it provides a fast solver and naturally accounts for the flow transition between veins and parenchyma. Finally, the model is verified against an analytical solution, showing good consistency.

In **Chapter 4**, based on [Audigier 2013a] and [Audigier 2015a], we evaluated the predictive power of the proposed framework on ten patients who underwent RFA, for whom pre- and post-operative images were available. Implemented on graphics processing units (GPU), our method simulates 1 minute of ablation in 1.14 minutes, allowing near real-time computation. This represents a 24% speed-up with respect to state of the art computation [Payne 2011]. By providing near real-time computation, our method may enable model-based planning of RFA in clinical settings. Comparisons between the computed ablation extent and ground truth, as observed in post-operative images, were promising (DICE index: 42%, sensitivity: 67%, positive predictive value: 38%). A mean point to mesh error between predicted and actual ablation extent of 10.17 mm is achieved. The clinical criterion is satisfied as the tumor is totally covered by the induced thermal lesion in all cases. The effect of liver perfusion while simulating electrical-heating ablation was also considering. In addition, the results of this chapter demonstrate the need for personalization of tissue properties, as illustrated in one patient. The heat conductivity was adjusted to reduce the point-to-mesh error with the ground-truth from 8.25 mm to 3.85mm for this patient.

Motivated by the results obtained in chapter 4, **Chapter 5** presents a framework for parameter estimation in order to model patient-specific radiofrequency ablation of multiple lesions in the case of metastatic diseases, published in [Audigier 2014a]. The proposed forward model is based upon the computational model of heat diffusion, cellular necrosis and blood flow through vessels and liver which relies on patient images presented in chapter 3. We identify the most sensitive material parameters, that need to be personalized from the available clinical imaging and

data. The selected parameters are the heat conductivity \bar{d}_t and the porosity ε . These parameters are then estimated using a gradient-free optimization method such that the point-to-mesh distance between the computed necrotic area and observed lesions is minimized. Based on the personalized parameters, the ablation of the remaining lesions are predicted. The framework is applied to a data set of seven lesions from three patients including pre- and post-operative CT images. In each case, the parameters were estimated on one tumor and RFA was simulated on the other tumor(s) using these personalized parameters, assuming the parameters to be spatially invariant within the same patient. Results showed significantly good correlation between predicted and actual ablation extent (average point-to-mesh errors of 4.03 mm).

In order to evaluate the RFA model in a more control environment, a pre-clinical study is proposed in **Chapter 6**. It is an experimental study especially designed for the validation of subject-specific multi-physics model of liver tumor radiofrequency ablation. The comprehensive experimental set-up is described first. It combines multi-modal, pre-, intra- and post-operative anatomical and functional images, as well as the interventional monitoring of the temperature and delivered power. The RFA model we aim to validate is then introduced. The RFA computation becomes subject-specific after several levels of personalization: anatomic, hemodynamics, heat transfer and an extended cellular necrosis model. Finally, a full processing pipeline required before using the image information into the model is presented. It copes with image noise, resolution and anisotropy. Methods of segmentation, registration and the estimation of smooth vessel tree tailored for computational fluid dynamics simulations are presented. This leads to an end-to-end pre-clinical validation framework that considers the most comprehensive data set for model evaluation.

In **Chapter 7**, results on the pre-clinical data are presented. The evaluation study encompasses twelve ablations from five healthy pigs: a mean point to mesh error between predicted and actual ablation extent of 5.3 mm is achieved, sufficient for clinical application as it is significantly lower than the difference of 1 cm in diameter configurations of the probe. We evaluate the error induced by each step of the pre-processing. After a sensitivity analysis of the model parameters, we also investigate the estimation of the cellular necrosis and heat transfer models parameters. By fitting the delivered power and the cooling temperatures measured by the RFA probe, key physiological parameters of the model can be correctly estimated as it leads to realistic predictions in terms of necrotic area, power and temperature. Such a framework would therefore not require any necrosis information, and would be better suited for clinical applications.

We conclude on the presented contributions and discuss perspectives in **Chapter 8**.

Finally, **Appendix A** gives the derivation of the Lattice Boltzmann equation

with the details of the required calculations. The derivations of diffusion equation, advection-diffusion equation and reaction-advection-diffusion equation are studied.

State of the Art on Radiofrequency Ablation of Hepatic Tumors

Contents

2.1 Structure and Function of the Liver	10
2.2 Radiofrequency Ablation	11
2.2.1 Radiofrequency Ablation Procedure	11
2.2.2 Monitoring of RFA based on MR Thermometry	13
2.3 Percutaneous Tumor Ablation Techniques	14
2.3.1 High Intensity Focused Ultrasound	15
2.3.2 Microwave Ablation	15
2.3.3 Cryotherapy	15
2.3.4 Laser Ablation	16
2.3.5 Electroporation	16
2.4 Technical Background	17
2.4.1 Models of Electrical Heating	17
2.4.2 Biophysical Models of Heat Transfer in Tissue	18
2.4.3 Models of Cellular Necrosis	20
2.5 State of the Art of Existing Radiofrequency Ablation Models 22	22
2.5.1 Computational Models of RFA	22
2.5.2 Model Simplification	23
2.5.3 Model Validation	24
2.5.4 Model Personalization	25
2.5.5 Surrogate Model for Optimization of the Probe Placement . .	26
2.6 MR Thermometry	27
2.6.1 Thermal Tissue Properties Acquisition during HIFU based on MR Thermometry	27
2.6.2 MR Thermometry Filtered by Bioheat Equation	27

2.1 Structure and Function of the Liver

Building an efficient and accurate computational model of RFA requires a high knowledge of the structure and function of the liver.




Figure 2.1: Anatomy of the human liver. *Image from the internet*¹.

The liver is a vital organ, which has a wide range of functions, including detoxification of various metabolites, protein synthesis, and the production of biochemicals necessary for digestion². A human liver has four lobes of unequal size and shape, whereas pig liver has five lobes. The liver is connected to two large blood vessels, the hepatic artery and the portal vein represented in Figure 2.1. The hepatic portal vein delivers approximately 75% of the liver's blood supply, and carries blood rich in digested nutrients from the entire gastrointestinal tract, its associated organs and also from the spleen and pancreas. The hepatic artery supply arterial blood to the liver, accounting for the remaining quarter of its blood flow. It carries oxygen-rich blood from the aorta.³ Oxygen is provided from both sources; approximately half of the liver's oxygen demand is met by the hepatic portal vein, and half is met by

¹<https://repairstemcell.wordpress.com/liver-failure-and-stem-cells/> Retrieved 2015-06-26.

²Anatomy and physiology of the liver - Canadian Cancer Society. Cancer.ca. Retrieved 2015-06-26.

³Medscape: Medscape Access. Emedicine.medscape.com. 2015-03-09. Retrieved 2015-06-26.

the hepatic arteries. Blood flows through the liver sinusoids and empties into each central vein. The central veins coalesce into hepatic veins, which leave the liver and drain into the inferior vena cava. The biliary tree, or biliary tract is derived from the branches of the bile ducts. It is the path by which bile is secreted by the liver then is transported to the first part of the small intestine, the duodenum.

2.2 Radiofrequency Ablation

2.2.1 Radiofrequency Ablation Procedure

During RFA procedure, the clinician places one or more probes using percutaneous, laparoscopic or open surgical (celiotomy) techniques within the target area in the liver parenchyma, within the malignant tissue. Although open surgical RFA procedure is more invasive, it allows repeated placement of the RFA probe at different locations to ablate several or large tumors, it can also be performed more easily [Minami 2011]. RF ablation relies on a complete electrical circuit to conduct RF current. This electrical circuit is created from the active probe through the body and is closed by surface electrodes (ground pads) as shown in Figure. 2.2.




Figure 2.2: A schematic diagram showing an RF needle deployed in a liver tumor. A complete electrical circuit is then created through the patient body allowing the alternating current of RF energy to go between the needle active electrode and the grounding pads. The white area is the targeted tumor, and the surrounding orange area, the necrosis lesion produced by RFA. *Image from [Curley 2003].*

From the active electrodes at the tip of the probe, a high frequency alternating

electric current flows through the surrounding region and heating is induced due to the electric resistance of the living tissue as illustrated in Figure. 2.3.




Figure 2.3: Drawing depicts how the heat expands around the needle electrodes into the surrounding tissue to form a thermal sphere during RFA intervention. In this case, the RFA probe just has 4 tips. *Image from Internet* ⁴.

Depending on the tumor size and the type of RFA probe used, the probe tips are deployed at a certain diameter in order to achieve a desired volume of ablation. This volume is established pre-operatively, ideally to destroy the visible tumor plus a 1 cm safety margin of ablation around it [Lencioni 2005]. Most of the time, in clinical settings, a temperature-controlled mode is used: the delivered RF energy increases up to a given maximum power until the given targeted electrode temperature is reached. Power is then scaled to maintain this temperature during a given ablation time. Following the ablation, the tips of the probe are retracted inside the main axis and the probe tract is cauterized as the RFA needle is withdrawn. This results in thermal coagulative necrosis at temperatures above 50°C due to irreversible protein denaturation of the cells.

Unlike extirpative therapies, the RFA treatment is difficult to monitor *in vivo* to prevent recurrence: success of the procedure depends on the complete coverage of the tumor by the thermal zone, which relies on optimal probe placements and the extent of conductive heat delivery controlled by the heating time. However, the latter is challenged by the hepatic blood vessels and the parenchyma perfusion that dissipate heat and make the size and shape of the ablation zone difficult to control thus potentially reducing RFA efficiency and increasing risks of recurrence [Kim 2006]. For example, Figure. 2.4 shows a liver extraction where the effect of the blood vessels on the necrotic area created on an *in vivo* RFA on a pig liver can be seen. Another great challenge for abdominal tumor treatment with RFA is the pre-interventional planning of the therapy, which has to be accurate in order to achieve a complete destruction of all the tumor cells.

In clinical practice, the planning is based on a 3D CT scan or MR image of the patient acquired pre-operatively. The clinician has to take into account the

⁴<http://www.radiologyinfo.org/en/info.cfm?pg=rfaliver>. Retrieved 2015-07-16.




Figure 2.4: Liver extraction showing RFA-induced lesion after *in vivo* ablation on a pig liver.

patient individual anatomy to successfully achieve thermal treatment of cancer cells. However the probe is inserted most of the time under ultrasound (US) guidance. The clinician mentally integrate the multi-modality data by visualizing the anatomy of the patient given by the pre-operative CT scan or the MR image and use the 2D slice visualization of the US alone in order to get the most appropriate position of the RFA probe, while, at the same time, manipulating the RFA probe. Once the RFA probe is implanted the probe location can be check in 3D by an anatomical MRI or a CT scan for example, like in Figure 2.5.

Besides the poor quality of the U.S guidance, the biophysical aspects have to be accounted for during the intervention. Even if a perfect guidance is achieved, the RFA outcome might still not be optimal. These are the reasons why guidance the RFA protocol for a specific patient is a challenging task and mathematical modeling has the potential to assist the radiofrequency ablation of liver tumors.

2.2.2 Monitoring of RFA based on MR Thermometry

Currently, the most widespread imaging technique for monitoring RFA of liver tumors is ultrasonography (US). However, real time MR temperature imaging covering the entire tumor and its surrounding area seems to be a more relevant strategy as it could be employed to interactively guide the RF probe to the targeted tumor and to monitor the effect of the therapy [Berjano 2006]. With temperature-based therapies such as RFA, a close relationship exists between cell death and the accumulated thermal dose (TD), which is dependent on both temperature increase and time of exposure [Sapareto 1984]. In their work [Lepetit-Coiffé 2010] show that it is possible to compute voxel-specific TD maps after the suppression of the electromagnetic interference by efficient filtering of the RF signal harmonics at the MR frequency




Figure 2.5: A CT slice showing radiofrequency ablation probe in the liver. *Image from Wikipedia*⁵

performed like in [Seror 2008], as this otherwise interferes with the measurements of the scanner. They demonstrated in two cases that RF ablation did not result in complete tumor ablation by showing insufficient coverage of the tumors. For ablation under MR guidance, the clinician could be notified in real-time that the end-point of the procedure is not reached whatever the source of the problem (heat-sink effect of flow, difficulties in tumor targeting) using MR thermometry. However, despite being the most sensitive imaging modality in detecting liver tumors, MR remains inefficient due to excessive positioning time that results from difficulty in accessing the patient and due to long procedure time [Guenette 2010]. Today, probe thermistors are used to assess the temperature at the tips of the probe. .

2.3 Percutaneous Tumor Ablation Techniques

Radiofrequency ablation (RFA) is the most common minimally-invasive technique. However other percutaneous techniques can be used to ablate tumors. We will present most of the currently used techniques in this section.

⁵Radiofrequency ablation. (2015, July 13). In Wikipedia, The Free Encyclopedia. Retrieved 16:44, August 14, 2015, from https://en.wikipedia.org/w/index.php?title=Radiofrequency_ablation&oldid=671315922

2.3.1 High Intensity Focused Ultrasound

High Intensity Focused Ultrasound (HIFU) is a medical procedure that applies high-intensity focused ultrasound energy to locally heat and destroy diseased or damaged tissue through ablation⁶. This technique creates short and local elevations of temperature at the focus of an ultrasonic therapy transducer and exposure time on the order of few seconds, inducing irreversible local coagulation necrosis. HIFU has advantages of no-invasiveness, good penetration, good selectiveness and easy power control [Zhang 2006]. Currently, further studies are needed to explore its treatment value for liver or prostate cancer for example.

2.3.2 Microwave Ablation

Microwave Ablation (MWA) is a form of thermal ablation used in interventional radiology to treat cancer. MWA uses electromagnetic waves in the microwave energy spectrum (300 MHz to 300 GHz) to produce tissue-heating effects. The oscillation of polar molecules produces frictional heating, ultimately generating tissue necrosis within solid tumors to damage and kill cancer cells or to make cancer cells more sensitive to the effects of radiation and certain anticancer drugs⁷. Theoretically, MWA has the potential advantages over RFA to generate larger zones of ablation in a shorter time and to generate higher temperatures. However, microwave power is inherently more difficult to generate and deliver safely and efficiently to the tissue when compared to RF power. Currently available microwave systems face technical limitations including large diameter probes, underpowered systems, shaft heating, long and relatively thin ablation zones and unpredictability of the size and shape of the zone of ablation. Further study and continued development of more robust clinical systems is still needed [Lubner 2010].

2.3.3 Cryotherapy

Cryotherapy (also called cryoablation or cryosurgery) is a clinical technique similar to RFA since the energy is delivered directly into the tumor by a probe, which is inserted through the skin. But rather than killing the tumor with heat, cryoablation uses an extremely cold gas (usually argon) to freeze it at a temperature of 233K to 248K. As the gas flows through the needle, a ball of ice crystals forms around the tip of the probe, thus immediately leading to cellular death of the surrounding tissue. This technique has been used for many years by surgeons in the operating

⁶High-intensity focused ultrasound. (2015, May 4). In Wikipedia, The Free Encyclopedia. Retrieved 14:08, August 5, 2015, from https://en.wikipedia.org/w/index.php?title=High-intensity_focused_ultrasound&oldid=660711436

⁷Microwave ablation. (2014, May 10). In Wikipedia, The Free Encyclopedia. Retrieved 14:10, August 5, 2015, from https://en.wikipedia.org/w/index.php?title=Microwave_ablation&oldid=607944914

room, but in the last few years, the needles have become small enough to be used by interventional radiologists through a small nick in the skin without the need for an operation. This technique has been applied to treat many kinds of tumors, such as breast cancer, primary or metastatic liver neoplasms, renal, lung, pancreas, and prostate cancer [Talbot 2014]. Cryotherapy may be less painful than RFA or MWA and ice ball can be visualized in real time (using CT, MRI or ultrasound) since there is no interference of the procedure with the imaging system. But the set-up is more complicated involving cryogens and tubing, the procedure duration is longer, the maneuverability of catheter is more difficult and it generates too small lesion size. Moreover, RFA allows the use of expandable electrodes and placement of multiple probes, which is not possible with cryoablation.

2.3.4 Laser Ablation

Percutaneous Laser Ablation (LA) is a technique for image-guided local tumor ablation within solid organs. Laser illumination of the optical fibers introduced into the target lesion delivers energy to the tumor [Pompili 2010], the targeted tissue is heated by the absorbed laser energy and evaporates or sublimates. It can be used on benign and malignant lesions in various organs. The main applications currently involve the destruction of primary and secondary malignant liver lesions. Laser ablation can make deeper lesions and does not predispose to blood clot formation. Yet, there still is a risk of damaging surrounding structures and a very limited experience with laser even during open-heart surgery.⁸

2.3.5 Electroporation

Electroporation is a molecular biology technique in which an electrical field is applied to cells to increase the permeability of the cell membrane, that allows chemicals, drugs, or DNA to be introduced into the cell [Neumann 1982]. Irreversible electroporation (IRE) applies short, high-voltage pulses to the tissue to permeabilize cell membranes and induce cellular apoptosis [Weiss 2013]. It can be relevant and efficient in organ like the pancreas, where the other technologies presented previously lack safety and efficacy owing to the risk of thermal damage to adjacent structures and significant thermal dissipation from large adjacent blood vessels. IRE is now used to treat pancreatic cancer previously thought to be unresectable.⁹ One potential benefit of IRE is that it generates less heat than RFA. Also, because the heat is generated during the pulses but dissipates while the electrode rests, there is not a constant cooking effect as with RFA, leading to potentially less damage to non targeted tissues. However, clinical experiences

⁸ <http://www.heartracing.com/physicians/cryo.vs.rf.asp>

⁹ "A Potential Boon for Pancreatic Cancer Patients". Johns Hopkins Surgery: News From the Johns Hopkins Department of Surgery. 2014-06-23.

are still limited before establishing the efficacy of IRE, whereas RFA have been time-tested for nearly three decades. Multiple needle placements are required within a prescribed distance, which can be challenging, and parallel placement of the probes may be hindered by issues such as intervening ribs. The actual timing of imaging follow-up and the best modality are still being determined. IRE also poses a unique challenge to the current imaging response criteria because unlike other ablation modalities, there is a marked decrease in the size of a successfully treated lesion [Narayanan 2013].

2.4 Technical Background

Computational models of RFA are divided into 3 different modeling parts: an electrical heating model, a heat transfer in living tissue model, and a cellular necrosis model. These models are coupled as illustrated on Figure. 2.6 and described in the following sections.




Figure 2.6: Sketch of the coupled modeling steps of the biophysical RFA model.

2.4.1 Models of Electrical Heating

Different types of radiofrequency (RF) probe can be used, either the cool-tip single probe or the RITA probe (StarBurst Radiofrequency Ablation, AngioDynamics, Latham, NY; www.angiodynamics.com) with three, four or six umbrella shaped prongs which can be deployed within the tumor (Figure. 2.7).

In the literature, both probe configurations are simulated [Tungjitkusolmun 2002, Panescu 1995]. The temperature and the voltage are usually set to be constant numbers on the surface of RF electrode probe needle and the potential field generated around the probe is solved numerically, using the Laplace equation (Eq. 2.1) for the electrical field in conductive media [Doss 1982].

$$\nabla \cdot \sigma \nabla V = 0 \quad (2.1)$$




Figure 2.7: RITA probe representation from [Payne 2011].

where V is the voltage and σ , the electrical conductivity. The Laplace equation (Eq. 2.1) allows to compute the current density \mathbf{J} and the electric field intensity \mathbf{E} , using the distributed heat source due to Joule loss q , given by Eq. 2.2:

$$q = \mathbf{J} \cdot \mathbf{E} \quad (2.2)$$

which can also be written as:

$$q = |\nabla V|^2 \quad (2.3)$$

Even if this approach is theoretically accurate, it requires a fine volume mesh on the surface of the very thin probe tips and then suffers from a high computational cost. Furthermore, the exact position of the whole electrode probe is not always available and the imperfect needle positioning has been found to severely affect the outcome of RFA procedure [Khlebnikov 2010]. The electric field strength is not high, most of the Joule effect heating is generated within the space 1 mm around the electrode surface [Panescu 1995] and most of the ablation zone at distances farther from the electrode is created by thermal conduction [Schramm 2006]. Moreover, the particular details of the heating point sources appear to have only a limited effect on the final lesion size [Payne 2011]. Thus, in our framework, the electrical heating is modeled with a Dirichlet boundary conditions fixing the delivered temperature either on a sphere, whose radius is defined pre-operatively by the protocol followed by the clinician if the position of the probe is unknown, or on the probe model itself if we know the probe position.

2.4.2 Biophysical Models of Heat Transfer in Tissue

Computing heat diffusion in biological tissues amounts to solving the coupled bio-heat equations derived from the theory of porous media (Eq. 2.4), where each elementary volume is assumed to comprise both tissue and blood with a certain fraction [Nakayama 2008].

$$(1 - \varepsilon)\rho c_{ti} \frac{\partial T_{ti}}{\partial t} = \underbrace{(1 - \varepsilon)Q}_{source} + (1 - \varepsilon) \underbrace{\nabla \cdot (d\nabla T_{ti})}_{diffusion} + H(T_{bl} - T_{ti}) \quad (2.4a)$$

$$\varepsilon\rho c_{bl} \left(\frac{\partial T_{bl}}{\partial t} + \underbrace{\mathbf{v} \cdot \nabla T_{bl}}_{advection} \right) = \varepsilon Q + \varepsilon \nabla \cdot (d\nabla T_{bl}) - H(T_{bl} - T_{ti}) \quad (2.4b)$$

In these equations, T , Q , \mathbf{v} , ε stand for temperature, source term, blood velocity, blood volume fraction (fraction of blood volume over total volume) and subscripts 'ti' and 'bl' refer to tissue and blood phase respectively. The other parameters are listed in Table 2.1. Different models similar to (Eq. 2.4) have been proposed in the litera-

Table 2.1: Values from literature [Payne 2011] of the parameters used in the computation of RFA.

parameter	description	value
ρ	blood and tissue densities	$1.06 \times 10^3 \text{ kg m}^{-3}$
c_{bl}	blood heat capacity	$4.18 \times 10^3 \text{ J}(\text{kg K})^{-1}$
c_{ti}^U, c_{ti}^V	tissue heat capacity	$3.6 \times 10^3 \text{ J}(\text{kg K})^{-1}$
c_{ti}^N	tissue heat capacity in necrotic cells	$0.67 \times 10^3 \text{ J}(\text{kg K})^{-1}$
d	blood and tissue heat conductivities	$0.512 \times (1 + 0.00161 \times (T - 310)) \text{ W}(\text{m K})^{-1}$
H	convective transfer coefficient	$24.4 \times 10^5 \text{ W}(\text{m}^3 \text{ K})^{-1}$
ε	blood volume fraction	0.1
κ	permeability	$4.0 \times 10^{-11} \text{ m}^2$
μ	dynamic viscosity of the blood	0.0035 Pa s
$\varphi_{vc_{in}}$	vena cava inflow	2.0 L min^{-1}
φ_i	flow through the inlets of the hepatic veins	1.6 L min^{-1}
p_0	vena cava outlet pressure	3 mmHg
\bar{k}_f	forward rate constant	$3.33 \times 10^{-3} \text{ s}^{-1}$
k_b	backward rate constant	$7.77 \times 10^{-3} \text{ s}^{-1}$
T_k	parameter of cell state model	40.5°C

ture [Roetzel 1998, Khaled 2003, Nakayama 2001, Shrivastava 2009]. Some models include a metabolic heat generation term [Mahjoob 2009], which can be neglected since it has been shown to be insignificant in thermal ablation [Labonté 1994]. The advection term in the blood bio-heat equation (Eq. 1.1b) is sometimes simplified by considering an averaged perfused tissue volume [Shrivastava 2009] but we preserve it in our model as it is essential when modeling the heat transfer between the tissue and the small vessels with low blood velocity magnitude. To account for the heat transfer associated with the transcapillary fluid exchange, some models include an additional perfusion heat transfer term [Nakayama 2008]. However, it has been proven that this term can be neglected as the blood in the capillary network usually reaches equilibrium with tissue temperature [Chato 1980]. Some bioheat models examined counter-current heat transfer in arterial-venous vessels (the venous flow is warmed through heating from the nearby arteries) [Mitchell 1968, Roetzel 1998, Chato 1980].

One common simplification of the coupled bio-heat equations is the *Pennes model* [Pennes 1948] where the blood temperature is assumed constant, which is valid within and close to large vessels, where the blood velocity magnitude is high. In this case, the coupled bio-heat equations (Eq. 2.4) reduce to one unique equation with only one temperature distribution T describing the temperature inside the liver:

$$\rho c_{ti} \frac{\partial T}{\partial t} = Q + \nabla \cdot (d \nabla T) + c_{bl} \omega (T_{b0} - T) \quad (2.5)$$

where ω represent the blood volumetric perfusion rate.

This simplification has been used widely in the literature to model the electrical-thermal heating process happening in RFA. However, this model assumes that the blood is a volumetric heat sink and that it is uniformly distributed throughout the tissue [Peng 2011]. Hence the term accounting for heat-transfer convection between tissue and blood in the Pennes equation is oversimplified. This simplification may be suitable for low vascularized organs like the heart or the kidney and so it can be used more accurately to model cardiac ablation or kidney cryotherapy [Linte 2013b, Talbot 2014] but it is not appropriate for the liver which is highly vascularized. Therefore, we have to account for the perfusion in the parenchyma in another way. Studies [Peng 2011] demonstrate that microvasculature perfusion does not act as a spatially homogeneous heat sink, which invalidates the fundamental assumption of the Pennes model when applied to the liver in which different types of vessels are present. For these reasons another simplification has been proposed, which results in the *Wulff-Klinger (WK) model* [Klinger 1974, Wulff 1974]. It assumes equilibrium between tissue and blood temperature which is accurate for highly perfused organ with small vessels where the blood velocity magnitude is low and accounts for the directional effect of the blood flow, with a blood volume fraction $\varepsilon \ll 1$:

$$(1 - \varepsilon)\rho_{cti} \frac{\partial T}{\partial t} = (1 - \varepsilon)Q + (1 - \varepsilon)\nabla \cdot (d\nabla T) - \varepsilon\rho_{cbt}\mathbf{v} \cdot \nabla T \quad (2.6)$$

This equation holds in a porous medium where tissue is dominating. The main difference between these two models lies in their cooling terms (last term of the right-hand side). The former is a reaction term and acts as a volumetric homogeneous heat sink whilst the latter is an advection term accounting for the directional effect of blood flow on the tissue temperature. It can be shown [Peng 2011] analytically and computationally but in 1D that the coupled bio-heat equations (Eq. 2.4) can be simplified into the Pennes model in the case of large vessels, and to the WK model in the case of small vessels.

2.4.3 Models of Cellular Necrosis

Thermal treatments aim at transporting heat energy within the cancerous tumor then creating a zone of dead cells surrounded by tissue which could eventually recover after the ablation procedure. For a cell to go to an apoptosis state, a critical temperature should be exceeded during a sufficient time duration, due to the thermal tolerance of cells. Various cellular necrosis models [Berjano 2006, Diller 2008, O'Neill 2011] have been studied.

The simplest one is to use a single temperature threshold above which cells stop instantaneously functioning, and below which cells remain fully functional. Although different values have been used for this threshold, the *in vivo* lesion volume (i.e. necrotic tissue after RF ablation) can be defined by the volume enclosed by the 50 °C isothermal surface [Panescu 1995], which has been widely used for RFA

computational simulations. This isothermal model does not take into account the duration for which cells are at high temperatures although this factor has an influence on the cellular response since tissue damage is a function of both temperature and time [ONeill 2011].

Arrhenius-based models have been proposed [Henriques Jr 1947, Moritz 1947], which extend the Arrhenius law proposed for chemical reaction rates. The rate of cell damage is then proportional to $\exp(-Ea/RT)$, where Ea is an activation energy, R is the universal gas constant, and T is the temperature. In this case, the tissue damage increases linearly with time and hyperbolically with temperature.

Thermal dose models have also been proposed. The earliest work was based on the concept of "equivalent minutes" [Sapareto 1984], where they characterize the thermal dose in an equivalent time at a temperature fixed at 43°C. The thermal dose in equivalent minutes at 43°C can be computed at each position according to this empirical model:

$$t_{43} = \sum_{t=0}^{t_c} C^{(43-\bar{T})} \Delta t \quad (2.7)$$

where t_{43} is the thermal dose in equivalent minutes at 43°C, \bar{T} is the average temperature during the time interval Δt , and C is a constant resulting from the Arrhenius model:

$$\begin{aligned} C &= 0.5 & \text{when } \bar{T} \geq 43^\circ\text{C} \\ C &= 0.25 & \text{when } \bar{T} < 43^\circ\text{C} \end{aligned}$$

A thermal dose of 240 (in equivalent minutes at 43°C) is considered lethal. This model is based on the concept of a thermal dose, whereby cumulative damage leads to cell death. However, the possibility for a cell to recover is not allowed with this model as the thermal dose is monotonically increasing.

In [Linte 2013b], the authors define cumulative exposure as a measure of induced tissue injury. Cumulative exposure is the area under the temperature-time curve calculated on a voxel-basis over the duration of the ablation, which takes into account the voxel exposure prior to reaching the cell-death temperature (i.e., reversible damage and lesion penumbra) and also the voxel exposure beyond cell-death temperature for 5 s or longer (i.e., irreversible damage and core lesion). Their proposed exposure criterion yields consistent results with Arrhenius and the 50°C isotherm tissue injury criteria but at a lower computational expense.

Cell-death models based on the Arrhenius law are widely used [Pop 2010, Chen 2009a, Chang 2004, Schwarzmaier 1998], but are limited since the interpretation of model parameters characterizing the cell damage formulation is not easy to find. These models compute the degree of tissue damage at a tissue location r , using [Pearce 1995]:

$$\Omega(r, t) = \ln \left\{ \frac{[C(r, 0)]}{[C(r, t)]} \right\} = \int_0^t A e^{\frac{E_a}{RT(r, \tau)}} d\tau \quad (2.8)$$

where $\Omega(r, t)$ is the thermal damage index (dimensionless), τ denotes the exposure time (s), A is the frequency constant (s^{-1}), E_a is the activation energy ($J mol^{-1}$) and R is the universal gas constant ($8.314 J mol^{-1} K^{-1}$). The thermal damage is defined as the logarithm of the ratio of the original concentration of native tissue at time zero: $C(r, 0)$, to the remaining native state tissue at time t : $C(r, t)$. A value of 1 (respectively 4.6) for Ω represents 63% (respectively 99%) of the cells in a damaged state. In some cases, this model is coupled to the heat transfer model: the perfusion term is multiplied by $\alpha = \frac{1}{\Omega}$, since once tissue coagulation occurs, tissue perfusion ceases [Chang 2004].

Those models are sensitive to small changes in parameters, and they are not able to predict cellular injury over a wide hyperthermic temperature range and throughout the entire heating process [Feng 2011]. To tackle these drawbacks, two-compartment models have been proposed containing either fully alive or fully dead cells [Feng 2008, Feng 2011]. In order to characterize the biological state changes, transitions between the two compartments are usually modeled with first-order rate processes.

Finally a three-compartment cell death model [ONeill 2011] can be used to simulate both fast and slow cell death over a temperature range extending to $100^\circ C$. This last model uses a single continuous function where a backward, recovery process is accounted for by adding a vulnerable state of the cell from which cells can either die or heal and return to the alive compartment. In this model, thermal-based necrosis depends on both the heating temperature and the heating time. This model accounts for the recovery process of the cells, allows to define three different states for the cell properties, however, it required to estimate several parameters.

2.5 State of the Art of Existing Radiofrequency Ablation Models

2.5.1 Computational Models of RFA

Several research groups are working on methods to extract information from the images, and on models to simulate the RFA behaviour. More specifically, such models could potentially calculate the size, the shape and the location of the necrotic area, given the position and settings of the ablation probe in the abdomen. Thus they could give additional guidance to clinicians in deciding where to place the heating probe and for how long heating must be applied in order to fully ablate the tumor and limit the number of recurrences. For instance, the placement of the probe has been optimized in [Schumann 2015] by using segmentation masks and by the exploration of the set of pareto-efficient solutions, in order to improve the established planning of RFA.

Several approaches have been developed to describe and simulate RFA of liver tumors. They differ in their choice of the biophysical phenomena considered and the type of experimental data used to design and validate them.

All simulations are based on the bio-heat equation considering a cooling effect that is either expanded [Chen 2009b] or localized at neighboring vessels [Altrogge 2012, Jiang 2010, Payne 2011, Audigier 2015a]. Those studies couple radiofrequency electrical fields to thermal transport. They have been developed to compute heat diffusion in the liver, predict the temperature distribution during the procedure, study the biophysical impact of the temperature and eventually evaluate the optimal placement of the RFA probes [Kröger 2010].

Early studies use simplified 3D geometries to determine tissue temperature distribution during RF hepatic ablation [Tungjitkusolmun 2002]. Other studies [Chen 2009b, Jiang 2010] simulate the heat transfer on generic human anatomies but they do not rely on patient-specific data. Few authors [Altrogge 2012, Audigier 2015a, Payne 2011] have proposed to simulate RFA on realistic subject-specific geometries extracted from images and only [Audigier 2013a] has personalized biophysical parameters on patient data in order to minimize the discrepancy between simulated and measured necrotic (ablated) regions, which is paramount since parameters in the literature often come from experiments of tissue from different species ex-vivo [Hall 2015].

Some authors simulated the bio-heat equation based on the most common discretization method: the Finite Element Method (FEM) [Chen 2009b, Altrogge 2012, Jiang 2010], Instead we introduced the Lattice-Boltzmann Method (LBM) [Audigier 2015a, Audigier 2013a]. The traditional method to discretize RF ablation PDE is indeed the Finite-Element Method (FEM). However, the use of classic FEM is often more computationally demanding (execution time is usually in the range of hours, which is not suitable for clinical purposes nor therapy guidance) than using recent numerical methods such as the Lattice Boltzmann Method (LBM) [Rapaka 2012]. Other authors proposed simplified model using a weighted distance-based method [Rieder 2011, Schumann 2015] to approximate the ablation zone in real-time utilizing the GPU, but the analysis of the effect of using an approximation instead of the complex numerical simulation has not been done yet. Furthermore, a preprocessed thermal equilibrium representation of the liver parenchyma and blood vessels is needed to incorporate its cooling effect. In [Rieder 2011], they only compare the ablation zone given by their method with the coagulative necrosis mask from the numerical simulation in complex vascular situations, indicating a good correlation.

2.5.2 Model Simplification

In most of the developed models, numerous simplifications are assumed, such as the homogeneity and the isotropy of the tissue, a blood perfusion rate unaffected by the heating process, no boiling of tissue during heating [Berjano 2006]. Also the tissue itself is neither isotropic nor homogeneous, particularly in diseased states, which are cases where ablative treatment are commonly used [Payne 2010]. In this work, the properties of the tissue (heat capacity, etc) vary with respect to its state and different parameters are related to different location inside the liver. The reac-

tion coefficient, H and the porosity, ε are related to large vessels, and parenchyma respectively. Nevertheless, some models include the water evaporation during the RFA procedure [Yang 2007, Pätz 2010], porous circulation [Crezee 1990], or blood coagulation [Chang 2004]. An important issue is the relationship between tissue characteristics and temperature. RF ablation procedures involves heating at temperature around 105°C, the parameters values found in the literature may not be correct in this range of temperature. Some models ignore the thermal dependence of the heat diffusivity of tissues [Chen 2009b, Jiang 2010], but it is still not clear if this dependence of the tissue properties have an important impact on the RFA output. For this reason, we keep in this work the definition used in [Payne 2011], where the heat diffusivity is temperature-dependent (Table 2.1).

The liver is a very complex thermal organ due to its inhomogeneity. It is composed of three different types of blood vessels (hepatic arteries, portal veins, hepatic veins), which have different diameters and different flow velocities [Tungjitkusolmun 2002]. The large blood vessels have a significant cooling effect on RF lesions as it has been shown on ex-vivo porcine liver [Lehmann 2009]. The heterogeneous cooling effect of large neighboring vessels have been studied [Patterson 1998], but the effect of the perfusion is still not totally defined. For these reasons, the vascular system of the liver has to be considered, and we model it in this work but it is often neglected or simplified in most of the studies. Moreover, blood flow circulation is often not computed based on patient-specific clinical information, whereas it is the case in this work. To give an example, the blood temperature in the large vessels is unaffected by the thermal field in the surrounding tissue in [Tungjitkusolmun 2002]. Other models [Pennes 1948, Panescu 1995, Labonté 1994] used convective boundary conditions at the tissue - vessels interfaces to simulate the heat transfer due to blood flow, but in order to achieve accurate prediction of RFA output, blood flow needs to be included in the model [Jain 2000].

It is also particularly important to take into account the effect of perfusion, which is often neglected in these models, it is done in this work, as in [Audigier 2015a, Payne 2011] where the cooling effect due to venous flow in the liver parenchyma is considered.

2.5.3 Model Validation

Subject-specific modeling of liver tumor radiofrequency ablation (RFA) needs further validation before impacting clinical use.

Up to now, the comparison between simulated and measured necrotic regions has been used by several authors [Altrogge 2012, Payne 2011, Audigier 2015a] as the main criteria of success in calculating the effect of RFA on abdominal tumors, for either model validation or personalization. In [Payne 2011], the authors compare the simulated volume with the measured one on two pigs and they have an average error of 40% on the volume. Whereas, in our work, point-to-mesh errors of 5 mm on average are computed between the simulated and the actual lesion.

Computational models of RFA depend on a large number of different parameters, which depend mostly on the patient, the temperature or the current state of the tissue. They also have a high computational cost. RFA model validation is then challenging with clinical data but is more feasible on pre-clinical data in a more controlled environment. This extensive validation step is important prior to any translation into clinical settings. It is paramount because the size and the location of the RFA lesion has to be predictable and controllable for clinical applications. A comprehensive understanding of detailed multi-physics model is required first to simplify only the aspects that are not needed, which could then enable model personalization or simplification and eventually computational models of RFA can be helpful for interventional guidance or therapy planning. Several experimental validation of RFA modeling have been proposed, but they rely on *ex vivo* or *in vitro* data [Linte 2013a, Jain 2000]. In [Linte 2013a], an *ex vivo* study on two beef muscle samples is performed to validate an endocardial tissue RFA model by comparing the predicted versus the experimentally measured temperature at 2 locations. Validation on *in vivo* data has been performed as well [Payne 2011], but the validation is done by comparing the volumes of the simulated and post-operative lesions on two pigs only, with nominal biophysical parameters coming from the literature.

2.5.4 Model Personalization

The accuracy of the RFA computation is challenged by the inter-subject variability in anatomies and tissue characteristics [Hall 2015]. Tissue properties depend upon both the tissue state and the temperature. There is little information available on tumor or disease tissue, which are of interest in our case [Payne 2010]. Because tissue properties are patient-specific and can depend on the current state of the tissue, a proper estimation of those parameters is needed but has been often overlooked in the modeling of RFA.

In cardiac modeling, model personalization, i.e. estimation of patient-specific model parameters has been studied. The tissue conductivity as well as the biophysical model parameters for the electrophysiology are estimated in [Relan 2011], whereas in [Marchesseau 2013], the mechanical parameters of an electromechanical model of the heart are personalized.

In RFA modeling, studies [Altrogge 2012, Chen 2009b, Jiang 2010] have investigated the Finite Element Method (FEM) to simulate the heat transfer on generic human anatomies. Simulations with animal-specific [Payne 2011] or patient-specific anatomies were also recently considered [Audigier 2013a] with the inclusion of cooling effects computed from simulated hepatic venous flow and hepatic parenchymal flow. However, nominal tissue parameters were employed in these studies with values often based on *ex vivo* experiments on animal tissue sometimes with a large varying range between published studies [Hall 2015]. In [Altrogge 2012], the uncertainty associated with biophysical tissue properties (electrical conductivity and thermal diffusivity) is taking into account. The RFA output is evaluated for certain realizations of tissue parameters at an extremely high computational cost. However,

they do not tackle the problem of patient-specific parameters by trying to obtain more accurate material parameters, as the uncertainty is considered as an intrinsic attribute of the modeling process.

In this work, we first attempt to estimate patient-specific parameters by minimizing the error between the simulated necrosis and the post-operative registered ground-truth, in the case of multiple lesions ablated per patient [Audigier 2014a]. However, the necrosis of tissue is the resultant of several combined physical phenomena, mainly the heat transfer and cell death mechanisms, meaning that a given ablated region may be explained by several combinations of parameters. In addition to this identifiability issue, the size of the tumor extent can only be known reliably from post-operative imaging and its shape may be highly asymmetric [Hall 2015], which makes it difficult to eventually update the ablation plan during the procedure, which is the clinical end-point. A method that relies also on pre-operative or interventional data for personalization is therefore required for RFA models to be clinically useful. Those observations are complementary information to the necrotic extent for model evaluation and personalization. A patient-specific tool showing the extent of ablation given the probe position, the heat duration, personalized parameters and patient images will potentially be beneficial in providing a personalized treatment planning and guidance, as it could improve the current clinical outcome.

2.5.5 Surrogate Model for Optimization of the Probe Placement

Inserting the RFA probe in the liver parenchyma requires considerable experience and constitutes a significant mental task for the clinician. During the procedure, multiple criteria for all possible trajectories have to be taken into account like the effect of the large blood vessels, the distance to the lungs, the ribs, etc. For these reasons, several methods to determine the optimal access path have been proposed [Chen 2009a, Altrogge 2012, Schumann 2015].

Some automatic and semi-automatic approaches concentrate on calculation and simulation of the ablation zone and the resulting optimal probe placement with respect to the coverage of the tumor. In [Chen 2009a], the probe placement is optimized on a simplified geometry such that the resulting necrotic lesion totally covers the targeted area. The probe placement can be optimized by maximizing the minimum temperature inside the tumor area such that the volume of destroyed tumor tissue is maximized [Altrogge 2012]. Whereas in [Schumann 2015], the optimization of the probe placement is based on numerical optimization and on image processing: segmentations of the area of interest like vessels, lungs, ribs, tumors, liver.

Other methods propose to optimize the planning of the trajectory for percutaneous needle insertions by taking into account the critical structures but also the instrument shape and penetration angle. Methods taking these constraints into account in order to automatically compute insertion trajectories [Villard 2004, Villard 2005, Baegert 2007, Seitel 2011] have been proposed. A simplified radiofrequency ablation simulator can be used [Villard 2003] to get an automatic planning of needle positions considering the tumor environment. The necrosis zone is

represented by spheroid deformed by the surrounding vessels to take into account the heat-sink effect. In [Seitel 2011], they use pareto optimality to have a weight-independent approach whereas in [Baegert 2007] the planning of the automatic trajectory in the abdomen is done by a weighting of the different constraints.

2.6 MR Thermometry

MRI provides excellent soft-tissue contrast anatomical images as well as a continuous temperature monitoring, thanks to the combination with proton resonance frequency shift (PRFS) based MR thermometry [Ishihara 1995]. Hence, MR thermometry could estimate the resulting lesion sizes of any thermal ablation and thus provide an adequate therapy end point [Quesson 2000]. However, precise temperature measurements using PRF require to elaborate correction techniques, if motion and magnetic field variations are present during the treatment [De Senneville 2007]. More recently, thermal tissue properties have been estimated using MR thermometry during HIFU [Cornelis 2011].

2.6.1 Thermal Tissue Properties Acquisition during HIFU based on MR Thermometry

The development of MR hybrid hyperthermia systems with thermal imaging capability has opened the door to acquisition of thermal tissue properties during heating. In [Cornelis 2011], the authors evaluate quantitatively *in vivo* the tissue thermal properties during high-intensity focused ultrasound (HIFU) heating of pig kidneys monitored in real time by volumetric MR thermometry. The analysis of the volumetric temperature distribution during the HIFU intervention allows the determination of the thermal parameters, and may therefore improve the quality of the planning of noninvasive therapy with MR-guided HIFU.

2.6.2 MR Thermometry Filtered by Bioheat Equation

Extending MR temperature imaging to include the entire targeted area and its vicinity with a high accuracy is a difficult problem as the trade-off among spatial resolution, temporal resolution, and volume coverage is inherent in MRI. In addition, the precision of real time MR-thermometry for therapy guidance is generally limited by the available signal-to-noise ratio (SNR) and the influence of physiological noise. In [De Senneville 2013], they generate 3D temperature maps by combining 2D slices measured by MR thermometry and a priori knowledge of 3-D data derived from predictions based on a physical model (the bio-heat transfer equation model) using an extended Kalman filter (EKF). This is done to monitor HIFU intervention but could be applicable to RFA.

Modeling of Radiofrequency Ablation

Contents

3.1	Introduction	29
3.2	Method	30
3.2.1	Model of the Patient Hepatic Blood Circulation System	30
3.2.2	Fully Coupled Model of Heat Transfer in Liver Tissue and Cellular Necrosis	38
3.2.3	Estimation of Anatomical Model from Patient Images	45
3.3	Heating Power and Cooling Temperature Computation	46
3.3.1	Heating Stage	46
3.3.2	Cooling Stage	47
3.4	Model Verification	47
3.4.1	Quantitative Verification of the Heat Transfer Model	47
3.4.2	Computational Efficiency	49
3.5	Discussion	50
3.5.1	Summary	50
3.5.2	Model Limitations	50
3.6	Conclusion	51

Based On: the conference paper [Audigier 2013a] and the extended journal version [Audigier 2015a] for the description of the model.

3.1 Introduction

Turning computational models of RFA into clinical practice remains challenging. The lack of an integrated, efficient, patient-specific framework for RFA modeling based on patient data is the major difficulty encountered. Studies based on *in vivo* animal images have been reported, but time-consuming FEM computations are still required [Payne 2011]. Because of the expensive computation time needed by FEM, current approaches cannot enable a personalization of the model-based planning of RFA in a clinical setting. Therefore a fast and efficient model would help to optimize the treatment protocol pre-operatively.

As a first step towards efficient patient-specific planning of RFA, we propose an integrated multi-physics approach that combines, the computation of the hepatic blood flow with a biophysical model of the heat transfer and cellular necrosis to simulate RFA therapies in patients and predict the extent of ablation. The computational model of heat transfer and cellular death during RFA is only based on volumetric image of the anatomies estimated from CT (Sec. 3.2.3).

This chapter presents the Lattice Boltzmann Method (LBM) for heat in biological tissue. This method is a kinetic-based approach for fluid flow computations which has been successfully used as an alternative numerical method for solving Navier-Stokes type equations [Chen 1998]. In a first step, the computation of the hepatic blood flow is done using a Computational Fluid Dynamics (CFD) method in the large vessels and Darcy's law in the parenchyma. We will refer to this method throughout the manuscript as the Decoupled FEM Porous - Vessels CFD Model. To go further, an improved hepatic blood flow computation is presented, in which we rely on LBM to compute not only heat diffusion, cellular necrosis as previously proposed but also blood and parenchyma flow in the liver tissue. This latter method is based on a CFD solver which incorporates a porous part to deal with the liver parenchyma. We will refer to this method throughout the manuscript as the Fully Coupled LBM CFD Model. This second framework is particularly efficient for the personalization as it provides a fast solver and naturally accounts for the flow transition between veins and parenchyma. This framework is adapted to situation where no temperature map is available.

A detailed convergence analysis against an analytical solution of the heat transfer equation in Sec. 3.4.1 and a computational efficiency study in Sec. 3.4.2 are presented.

3.2 Method

As illustrated in Fig. 3.1, and detailed in the following section, we first estimate a comprehensive level set representation of the liver, including parenchyma, blood vessels and tumors from medical CT images. Then a computational model of heat propagation, cellular necrosis and blood flow through the vessels and liver is solved to estimate the extent of the ablated tissue.

3.2.1 Model of the Patient Hepatic Blood Circulation System

The heat diffusion in liver tissue, involved during the RFA procedure depends on the velocity field of the hepatic blood flow. In this section, we first propose a Decoupled FEM Porous-Vessels CFD Model to estimate the blood flow in the liver parenchyma. The 3D blood flow is first computed on the large vessels and then the porous flow in the liver parenchyma is computed using the Finite Element Method (FEM) on a tetrahedral mesh. As this method relies on FEM, meshing techniques have to be employed, which increases the complexity of the model and the computational errors. For this reason a second Fully Coupled LBM CFD Model is proposed,




Figure 3.1: Steps of the proposed method (green: input, blue: processes, red: output).

where the blood flow in the large vessels as well as in the liver parenchyma are computed simultaneously on a Cartesian grid using an Lattice Boltzmann (LBM) implementation.

3.2.1.1 Decoupled FEM Porous-Vessels CFD Model

In this section, we present our first approach: the Decoupled FEM Porous-Vessels CFD Model. The modeling pipeline is illustrated on Figure 3.2, and Algo. 1 presents the RFA computation method.

Algorithm 1 Computational model of RFA

- 1: Estimate patient-specific model of liver anatomy
 - 2: Compute 3D blood flow in hepatic veins, vena cava and portal vein
 - 3: Compute blood velocity field inside parenchyma
 - 4: **while** $t < t_{end}$ **do**
 - 5: Update temperature T using Pennes model in the large vessels, Wulff-Klinger model elsewhere
 - 6: Update cell-state
-

Heat transfer in liver tissue is highly dependent on the blood flow circulation. To solve the WK model (Eq. 2.6), we need the blood velocity field \mathbf{v} everywhere in the parenchyma. The blood inside the liver is modeled as a Newtonian fluid with pre-specified density ρ and viscosity μ . \mathbf{v} is calculated according to Darcy's




Figure 3.2: Steps of the proposed method using the Decoupled FEM Porous-Vessels CFD Model (green: input, blue: processes, red: output).

law [Brinkman 1949]:

$$\mathbf{v} = -\kappa/(\mu\varepsilon^{2/3})\nabla p \quad (3.1)$$

where p is the pressure. Because of mass conservation, this amounts to solving the Laplace equation:

$$\nabla \cdot (-\kappa/(\mu\varepsilon^{2/3})\nabla p) = 0 \quad (3.2)$$

At the border of the liver, Neumann boundary conditions are employed (no flow is leaking). Dirichlet boundary conditions are applied at the tip of the portal and hepatic veins, to define the pressure drop between them. As we cannot estimate these pressures *in vivo*, we rely on a CFD model of the hepatic venous circulation system to estimate them (Fig. 3.3).

We used a full 3D CFD solver (unsteady incompressible Navier-Stokes equations with viscous terms, expressed in an Eulerian framework which embeds the domain boundary using a level set representation of the segmented vessels [Ralovich 2012]). From the segmentation, a tetrahedral multi-domain mesh is generated based on the resulting multi-label mask image using CGAL [Alliez 2011] (www.cgal.org) to compute the porous flow, which is then calculated using FEM on the linear multi-domain tetrahedral mesh. The resulting flow is tri-linearly rasterized on the Cartesian grid after computation. In this work, the effect of heat on the viscosity of the flow is neglected to decouple flow-related from the heat diffusion calculation for computational efficiency. CFD and porous flow are calculated only once at the beginning of the algorithm.

Let $\varphi_{vc_{in}}$ be the vena cava inflow, φ_p the portal vein inflow and $\varphi_{vc} = \varphi_{vc_{in}} + \varphi_p$ the vena cava outflow (conservation of mass, the hepatic artery is neglected in this




Figure 3.3: Model of the hepatic circulatory system. Arrows denote blood flow. Circles and squares denote portal and hepatic vessel tips. *See text for details.*

study but could be added without modification into the framework). We also set the vena cava outlet pressure $p_0 = 3 \text{ mmHg}$ in the range of physiological values of healthy patients. The values are listed in Table 2.1. First, we compute the 3D blood flow and pressure distribution within the vena cava and hepatic veins as follows: a plug profile velocity field is applied at the inlets (squares in Fig. 3.3), computed from the outflow φ_p and the cross-sectional area of each inlet. The CFD calculation gives the downstream pressures p_i^- and the 3D blood flow φ_i for each inlet of the hepatic vein. Then we estimate the upstream pressure p^+ , assumed constant, of the portal vein outlets (circles in Fig. 3.3). We solve Darcy's law and optimize over p^+ such that the computed perfused flow through the hepatic vein inlets φ_i matches the one computed at the first step using the 3D CFD model. Then once p^+ is estimated, we compute the blood flow using the 3D CFD solver. Finally, we compute the blood velocity field inside the parenchyma using Eq. 3.1.

3.2.1.2 Fully Coupled LBM CFD Model

In order to go further and reduce the rasterization errors introduced by the FEM part of the former model, an improved hepatic blood flow computation is presented: a Fully Coupled LBM CFD solver. This method is based on a computational fluid Dynamics (CFD) solver which incorporates a porous part to deal with the liver parenchyma, and is implemented using the Lattice Boltzmann Method (LBM) to compute simultaneously the blood flow in the large vessels and in the liver parenchyma. The modeling pipeline is modified and illustrated in Fig. 3.4.

As detailed in the following section, we still rely on an estimation of a compre-




Figure 3.4: Steps of the proposed improved method (green: input, blue: processes, red: output). The blood flow is computed in the large vessels and in the parenchyma simultaneously.

hensive level set representation of the liver, including parenchyma, blood vessels and tumor from medical CT images. The computational model of heat propagation and cellular necrosis remains unchanged. However the blood flow computation through the vessels and liver parenchyma is done simultaneously. This method enables to estimate the extent of the ablated tissue as well. Algo. 3 presents the modified RFA computation method, with the differences with the previous model highlighted in red.

Algorithm 3 Computational model of RFA

- 1: Estimate patient-specific model of liver anatomy
 - 2: Compute 3D blood flow in hepatic veins, vena cava, portal vein, and inside parenchyma simultaneously
 - 3: **while** $t < t_{end}$ **do**
 - 4: Update temperature T using Pennes model in the large vessels, Wulff-Klinger model elsewhere
 - 5: Update cell-state
-

Description

The blood in the main vessels and in the parenchyma are combined in the generalized 3D incompressible Navier-Stokes equation for fluid flow in porous media, thus considerably easing the definition of boundary conditions between vessels and parenchyma, improving on the previous method [Audigier 2013a, Audigier 2015a]. More precisely, writing \mathbf{v} as the blood velocity and p the pressure inside the liver,

we solve:

$$\frac{\partial \mathbf{v}}{\partial t} + \mathbf{v} \cdot \nabla \mathbf{v} = -\frac{1}{\rho} \nabla p + \frac{\mu}{\rho} \nabla^2 \cdot \mathbf{v} + \mathbf{F} \quad \mathbf{F} = -\frac{\mu(1-\varepsilon)^2}{\rho \alpha^2 \varepsilon^2} \mathbf{v} \quad (3.3)$$

The added force \mathbf{F} represents the total body force due to the presence of a porous medium [Guo 2002]. \mathbf{F} depends on the porosity coefficient ε (fraction of blood volume over the total volume) whose default values are 1 in the CT-visible vessels, 0.1 in the porous parenchyma [Payne 2011], and 0.04 in the vessel walls to model an impermeable medium. Experiments have been performed to obtain a sufficiently small porosity (0.04) to avoid flowing through the vessel wall. In Eq. 3.3, ρ stands for the blood density, μ , the dynamic viscosity of the blood and α^2 an effective parameter. At the border of the liver, no flux boundary conditions are used whereas Dirichlet boundary conditions are applied at the inlets of portal vein and vena cava and at the outlet of the vena cava: the portal vein and vena cava inflow, φ_p and φ_{vcin} are fixed as well as the vena cava outlet pressure p_0 (see Figure 3.5 for more details). This method makes the boundary conditions simple to treat: no boundary conditions are fixed on the extremities of the vessels inside the parenchyma thanks to the use of the porosity map. This framework mainly avoids the occurrence of shear stress on the vessel walls due to their much lower value of porosity. Figure 3.5 illustrates flows calculated in a patient-specific geometry.




Figure 3.5: Set-up and results of the hepatic blood flow computation with a zoom inside the vena cava on the right and at the extremities of the hepatic veins on the left.

Link between the Decoupled FEM Porous-Vessels CFD and the Fully Coupled LBM CFD Models

In the Decoupled FEM Porous - Vessels CFD Model(Section 3.2.1.1), the blood flow is computed according to Darcy's law: Eq. 3.1 [Brinkman 1949]. In the Fully

Coupled LBM CFD Model (Section 3.2.1.2), the blood flow is computed according to the generalized 3D incompressible Navier-Stokes equation for fluid flow in porous media: Eq. 3.3. At steady state, Eq. 3.3 gives the blood velocity as a function of the gradient of the pressure:

$$\mathbf{v} = -\frac{\alpha^2 \varepsilon^2}{\mu(1-\varepsilon)^2} \nabla p \quad (3.4)$$

$\alpha^2 = d_p^2/150$, where d_p is the diameter of the solid particle [Guo 2002]. From (Eq. 3.1) and (Eq. 3.4), the same law should link the velocity and the gradient of the pressure. To this end, the value of d_p is set to $d_p = 1.5mm$ after computing both coefficients with biophysical values of parameters from Table. 2.1.

Numerical Resolution using LBM

Eq. 3.3 is solved using the Lattice Boltzmann Method (LBM) for fast computation on general purpose graphics processing units (GPU). LBM has been developed for CFD and is now a well-established discretization method [Chen 1998]. In RFA, it has been validated in [Audigier 2013a] through a comparison with an analytical solution, for a similar accuracy as FEM. In this case, LBM is used to compute the porous and blood flow circulation in the liver. To this end, an isotropic Cartesian grid with 19-connectivity topology (D3Q19 scheme defined in Figure 3.6) and Neumann boundary conditions is employed.




Figure 3.6: D3Q19 scheme used in the LBM computation of the blood flow.

A Multiple-Relaxation-Time (MRT) model is used for increased stability [Pan 2006]. At position \mathbf{x} for the edge \mathbf{e}_i , the governing equation is:

$$\mathbf{f}(\mathbf{x} + \mathbf{e}_i \Delta x, t + \Delta t) = \mathbf{f}(\mathbf{x}, t) + \mathbf{A}[\mathbf{f}^{eq}(\mathbf{x}, t) - \mathbf{f}(\mathbf{x}, t)] + \Delta t \mathbf{g}(\mathbf{x}, t). \quad (3.5)$$

In this equation, $\mathbf{f}(\mathbf{x}) = \{f_i(\mathbf{x})\}_{i=1..19}$ is the vector of distribution function with $f_i(\mathbf{x})$ being the probability of finding a particle travelling along the edge \mathbf{e}_i of the

node \mathbf{x} at a given time; $c = \Delta x / \Delta t$; $c_s^2 = 1/3$; Δx is the spacing;

$$f_i^{eq}(\mathbf{x}, t) = \omega_i \rho \left[1 + \frac{\mathbf{e}_i \cdot \mathbf{V}}{c c_s^2} \right] \quad g_i(\mathbf{x}, t) = \omega_i \rho \frac{\mathbf{e}_i \cdot \mathbf{F}}{c_s^2} \quad (3.6)$$

$\boldsymbol{\omega} = \{\omega_i\}_{i=1..19}$ is the vector of weighting factors and \mathbf{A} the MRT matrix. The fluid mass density and velocity are computed from the LBM distributions as:

$$\rho = \sum_{i=1}^{19} f_i(\mathbf{x}, t) \quad \rho \mathbf{v} = \sum_{i=1}^{19} \mathbf{e}_i f_i(\mathbf{x}, t) + \frac{\Delta t}{2} \rho \mathbf{F} \quad (3.7)$$

The fluid mass density and the velocity are updated at every node of the grid for every timestep Δt .

3.2.1.3 Differences between the Decoupled FEM Porous-Vessels CFD Model and the Fully Coupled LBM CFD Model

We first proposed the Decoupled FEM Porous-Vessels CFD Model to compute the blood flow in the hepatic vessels and in the liver parenchyma. This model requires two steps: first the computation of the blood flow in the large vessels using a Computational Fluid Dynamics (CFD) solver and then the computation of the blood flow in the parenchyma using Laplace equation and Darcy's law. This method requires mesh generation and a significant computational time. For these reasons, we proposed a second method where both flows (in large vessels and parenchyma) are computed simultaneously on the Cartesian grid using the Lattice Boltzmann Method (LBM).

There are differences in practice between those two methods. The Decoupled FEM Porous-Vessels CFD Model requires two decoupled computations. Those two computations are computationally demanding and we have to fit the values found by the second computation at the tips of the hepatic and portal veins to those found by the first computation. Moreover, the computation of the porous flow is done using the Finite Element Method (FEM). The computed flow is tri-linearly rasterized on the Cartesian grid to be used in the heat transfer model, which introduced interpolation errors. On the other hand, the Fully Coupled LBM CFD Model considerably eases the definition of boundary conditions between vessels and parenchyma: no boundary conditions are fixed on the extremities of the vessels inside the parenchyma thanks to the use of a porosity map, which is generated from the vessel mask. Secondly, no advanced meshing techniques are required, reducing the rasterization errors. Fig 3.7 shows the results of the two different computations on a patient-specific geometry.




Figure 3.7: (*Left*): Blood flow computation with the Decoupled FEM Porous-Vessels CFD Model; (*Right*): Blood flow computation with the Fully Coupled LBM CFD Model. In both cases, only the flow going through the portal inlet circled in white are shown. We can see differences between the results given by both computations. The area circled in black shows flow lines, which appear with the first computation and not the second one.

3.2.2 Fully Coupled Model of Heat Transfer in Liver Tissue and Cellular Necrosis

3.2.2.1 Model of Heat Transfer in Liver Tissue

This model describes how the heat flows from the probe through the liver while accounting for the cooling effect of the main vessels and parenchyma.

As current imaging techniques do not support an accurate measurement of the ratio between blood and liver tissue (large vessels can be clearly identified, but small capillaries are difficult to image), we use two simplifications of the coupled bioheat equation (Eq. 2.4) in the parenchyma and in visible blood vessels. Both equations can be easily implemented in a modular way to cope with tissue inhomogeneity. Continuity between the two models has to be ensured at the extremities of the large vessels. To that end, reaction coefficient (H or R) are chosen large enough in order to reduce the effect of the high temperatures relative to normal body temperature, i.e. 37°C . In our framework, a two-compartment model is used according to the spatial location within the anatomy: either the Pennes model or the WK model is used. Assuming that large blood vessels and surrounding tissue are isolated from each other, we compute the temperature by solving the diffusion equation:

$$\rho c_{ti} \frac{\partial T}{\partial t} = Q + \nabla \cdot (d \nabla T) \quad (3.8)$$

everywhere in the domain, to which we add the cooling term:

$H(T_{b0} - T)/(1 - \varepsilon)$ when a point belongs to a large vessel, where blood velocity is high, (Pennes model) or $-\varepsilon \rho c_{bl} \mathbf{v} \cdot \nabla T / (1 - \varepsilon)$ when it belongs to the parenchyma, where tissue is dominating (WK model). T , Q , \mathbf{v} and T_{b0} stand for temperature,

source term, blood velocity and the mean temperature (assumed constant) of the blood in large vessels.

In the case of a Fully Coupled LBM CFD Model, as the porosity ε vary spatially inside the liver, the temperature T is computed by solving the following equations: a reaction-diffusion equation in the large vessels, and an advection-diffusion equation in the liver parenchyma:

$$\rho c_{ti} \frac{\partial T}{\partial t} = Q + \nabla \cdot (d \nabla T) \begin{cases} +R(T_{b0} - T) \\ -\alpha_v \rho_b c_b \mathbf{v} \cdot \nabla T \end{cases} \quad (3.9)$$

where ρ , c , d are the density, heat capacity, conductivity; subscripts ti and b stand for tissue and blood respectively. Q is the source term, α_v , the advection coefficient, R , the reaction coefficient and T_{b0} the blood temperature (assumed constant) in large vessels. In order to be consistent with the previous model R is set to $R = 27.1 \times 10^4 W(m^3 K)^{-1}$ and $\alpha_v = 0.11$.

It can be shown [Peng 2011] analytically and computationally but in 1D that the coupled bio-heat equations (Eq. 2.4) can be simplified into the Pennes model in the case of large vessels, and to the WK model in the case of small vessels.

Our model includes the heat sink of all hepatic vessels (veins and arteries) as well as the effect of the blood flow within the parenchyma considered as a porous medium. A weakly coupling model is considered: the blood flow has an influence on the temperature distribution through the WK model and through the reaction term in the Pennes model but the temperature does not affect the blood flow (coagulation is not considered here), which allows us to speed up the calculations since the blood flow distribution is computed only once, at the beginning of the simulation. The CFD solver is run until a steady state is reached.

3.2.2.2 Cellular Necrosis Model

General Formulation

Tissue necrosis is calculated based on the computed temperatures using a three-state model [O'Neill 2011, Breen 2002], coupled with the bio-heat equation. Each cell has a probability to be either undamaged (U), vulnerable (V) or necrotic (N). Those probabilities vary with the temperature spatially and temporally according to the following transition diagram (Eq. 3.10):



where $\beta(T) = \bar{\beta} e^{T/T_k}$, $\delta(T) = \alpha \beta(T)$ are the rates of cell damage and γ is the rate of recovery. Unlike in [O'Neill 2011], a constant α is introduced. This diagram results in three coupled ODEs (Eq. 3.11), solved with a first order explicit scheme on the same grid and with the same time step as the bio-heat equation:

$$\begin{cases} \frac{dU}{dt} = -\beta(T)U + \gamma V \\ \frac{dV}{dt} = \beta(T)U - (\gamma + \delta(T))V \\ \frac{dN}{dt} = \delta(T)V \end{cases} \quad (3.11)$$

The following property is also imposed at each point:

$$U + V + N = 1 \quad (3.12)$$

The initial conditions are chosen as in [ONeill 2011]: $U = 0.99$, $V = 0.01$ and $N = 0$. We solve the resulting three coupled ODEs with a first order explicit scheme at each vertex of a Cartesian grid, yielding a spatially-varying cell state field used in the bio-heat solver. The computation is done on the same grid and with the same time step as the bio-heat equation. For both heat transfer and cell death models, parameters are initially set to values from the literature [Pennes 1948] reported in Table 2.1. The cell death model is strongly coupled to the bio-heat equation as the heat capacity depends on the state of the cell (c_t^U , c_t^V and c_t^N correspond respectively to the heat capacity of undamaged, vulnerable and necrotic tissue) and the conductivity d_t depends on the temperature through $d_t = \bar{d}_t * (1 + 1.61 * (T - 310).10^{-3})$ as in [Payne 2011]. All the remaining parameters are constant.

Nominal Formulation from [ONeill 2011]

To get the same equation as in [ONeill 2011], we set $\alpha = 1$, and the two damage rate coefficient are equal. In this case, we rename the parameters for the sake of clarity and Eq. 3.10 becomes Eq. 3.13:



$k_f(T) = \bar{k}_f e^{T/T_k} (1-U)$ and k_b are the rates of cell damage and recovery respectively. \bar{k}_f is a scaling constant and T_k is a parameter that sets the rate of the exponential increase with respect to the temperature.

This diagram results in three coupled ODEs (Eq. 3.14):

$$\begin{cases} \frac{dU}{dt} = -k_f(T)U + k_b V \\ \frac{dV}{dt} = k_f(T)U - (k_b + k_f(T))V \\ \frac{dN}{dt} = k_f(T)N \end{cases} \quad (3.14)$$

In this case, a single damage process incorporates all physiological damage mechanisms, thus the transitions $U \rightarrow V$ and $V \rightarrow N$ are the same. The vulnerable state is an arbitrary position representing the "point of no return", it is not a change in




Figure 3.8: Cell state evolution over time when tissue is heated at 105°C during 10 minutes.

the mechanism of thermal damage [ONeill 2011]. To avoid stagnation in its initial condition due to the formula of $k_f(T)$, an initial fraction of the cells has to be in the vulnerable compartment: the initial conditions are still chosen as in [ONeill 2011]: $U = 0.99$, $V = 0.01$ and $N = 0$. One percent seems reasonable since very small RMS error is achieved in [ONeill 2011] indicating a high quality of model fit to experimental data.

Figure 3.8 represents the solution of (Eq. 3.14) at one vertex over time if a constant temperature of 105°C is applied.

3.2.2.3 Numerical Resolution using LBM

We rely on the Lattice Boltzmann Method (LBM) where a statistical description of the system is used to compute heat diffusion and cellular necrosis in the liver tissue. LBM is a new computational method, which discretizes the velocity space on a grid, with mass, momentum and energy conservation conditions. To meet these three conservation conditions, two separate distribution functions are usually used [Mai 2010]. In the Decoupled FEM Porous-Vessels CFD approach, FEM solvers (CFD solver in the large vessels and porous solver in the parenchyma) give the blood flow distribution in the liver and enforce the mass and momentum conservation. Energy conservation is modeled using a distribution function for the thermal energy. Whereas, in the case of Fully Coupled LBM CFD solver, to meet the three conservation conditions of mass, momentum and energy, two separate distribution functions are used as in [Mai 2010]. In this approach, two LBM solvers (one CFD solver in the large vessels and in the parenchyma and one heat transfer solver) are

used, the first one gives the blood flow distribution in the liver and the second one the temperature and cell state distributions.

LBM has emerged as a powerful technique for efficient computation of second order elliptic partial-differential equations [Yoshida 2010]. It is used in Computational Fluid Dynamics to solve Navier-Stokes equations [Chen 1998], in heat transfer problem [Dawson 1993], or in cardiac electrophysiology [Rapaka 2012] to solve reaction-diffusion equations. Contrary to FEM, this discretization method uses a grid, allowing an easy parallelisation. We used an isotropic Cartesian grid and, LBM is performed with a D3Q7 scheme (DnQm denotes m discrete velocities in n dimensions): 6 directions are considered + the current point [Rapaka 2012] as described in Figure 3.9.




Figure 3.9: The D3Q7 scheme used in the heat diffusion solver from [Yoshida 2010].

Numerically, LBM uses fictitious particles, which perform consecutive collision and streaming processes over the discrete lattice mesh. When there is a non-zero probability that particle distributions move to the same node from different directions, then the probability of having particle distributions at that node with a given velocity direction is changed due to the application of a collision operator. The governing equation at position $\mathbf{x} = (x, y, z)$ for the direction defined by the vector \mathbf{e}_i is given by the two equations (Eq. 3.15, Eq. 3.16).

$\mathbf{f}(\mathbf{x}) = \{f_i(\mathbf{x})\}_{i=1..7}$ is the vector of the temperature energy distribution function with $f_i(\mathbf{x})$, the probability of finding a particle traveling along the vector \mathbf{e}_i of the node \mathbf{x} at a given time.

$\hat{\mathbf{f}}(\mathbf{x})$ describes the post-collision distribution.

$\boldsymbol{\omega} = \{\omega_i\}_{i=1..7} = (1/4, 1/8, 1/8, 1/8, 1/8, 1/8, 1/8)$ is the vector of weighting factors [Rapaka 2012] that depends on the lattice connectivity.

First, the collision step is:

$$\hat{\mathbf{f}}(\mathbf{x}, t + \Delta t) = \mathbf{f}(\mathbf{x}, t) + \mathbf{M}^{-1} \hat{\mathbf{S}}[\mathbf{M}\mathbf{f}^{eq}(\mathbf{x}, t) - \mathbf{M}\mathbf{f}(\mathbf{x}, t)] + \omega \Delta t H(T_{b0} - T(\mathbf{x}, t)) \quad (3.15)$$

and then, the streaming step is:

$$f_i(\mathbf{x} + \mathbf{e}_i \Delta x, t + \Delta t) = \hat{f}_i(\mathbf{x}, t + \Delta t) \quad \forall i \in 1 \dots 7 \quad (3.16)$$

with:

$$f_i^{eq}(\mathbf{x}, t) = \omega_i T(\mathbf{x}, t) \left[1 + \frac{\mathbf{e}_i \cdot \mathbf{v}}{cc_s^2} \right]$$

We denote $c = \Delta x / \Delta t$, $c_s^2 = 1/4$ and Δx as the spacing. The set of vectors \mathbf{e}_i is defined as:

$$[\mathbf{e}_1, \mathbf{e}_2, \mathbf{e}_3, \mathbf{e}_4, \mathbf{e}_5, \mathbf{e}_6, \mathbf{e}_7]^T = \begin{pmatrix} 0 & 1 & -1 & 0 & 0 & 0 & 0 \\ 0 & 0 & 0 & 1 & -1 & 0 & 0 \\ 0 & 0 & 0 & 0 & 0 & 1 & -1 \end{pmatrix}$$

At each time step, the entire 3D lattice domain is traversed and for each cell, new distribution function values are computed from its six neighbors (plus itself). The temperature corresponding to the 0th moment is computed as:

$$T(\mathbf{x}, t) = \sum_{i=1}^7 f_i(\mathbf{x}, t) \quad (3.17)$$

Only the data of the cell and its six neighbors are used, so the lattice can be traversed in any order since values from the neighbors are computed from the previous time step. The LBM offers high parallel scalability, second order accuracy in space and the simplicity of an implementation on a uniform Cartesian grid [Rapaka 2012]. The relation between LBM equations (Eq. 3.15, Eq. 3.16) and the continuous model can be derived by employing the Chapman-Enskog expansion, (a formal multi-scale expansion) [Chen 1998]. The derivation are given in Appendix A. For stability reason, we use a Multiple-Relaxation-Time model since the numerical stability can be improved by separating the relaxation rates of the conserved and non-conserved moments [d'Humières 2002]. First, \mathbf{f} is brought to a new basis in which each component corresponds to a certain moment of the vector $\mathbf{M}\mathbf{f}$ (0th order is the temperature T). Then each component relaxes to the equilibrium $\mathbf{M}\mathbf{f}^{eq}$ with a different relaxation coefficient. Finally, the vector is projected back onto the original seven-dimension space [Yoshida 2010]. In (Eq. 3.15), instead of writing $\mathbf{A} = \mathbf{M}^{-1} \hat{\mathbf{S}} \mathbf{M} = 1/\tau \mathbf{I}$, the relaxation towards equilibrium is performed in the moment space, where

$$\mathbf{M} = \begin{pmatrix} 1 & 1 & 1 & 1 & 1 & 1 & 1 \\ 0 & 1 & -1 & 0 & 0 & 0 & 0 \\ 0 & 0 & 0 & 1 & -1 & 0 & 0 \\ 0 & 0 & 0 & 0 & 0 & 1 & -1 \\ 6 & -1 & -1 & -1 & -1 & -1 & -1 \\ 0 & 2 & 2 & -1 & -1 & -1 & -1 \\ 0 & 0 & 0 & 1 & 1 & -1 & -1 \end{pmatrix}$$

and $\hat{\mathbf{S}} = \text{diag}(1, 1/\tau, 1/\tau, 1/\tau, 1/\tau_1, 1/\tau_1, 1/\tau_1)$ is the collision matrix in moment space. The relaxation time τ is directly related to the diffusion coefficient D through $\tau = 1/2 + 4D\Delta t/\Delta x^2$ [Yoshida 2010]. After a stability analysis, we chose $\tau_1 = 1.33$ [Rapaka 2012] to get a stable and well-behaved solution. Based on the Chapman-Enskog expansion, this value does not change the accuracy of the solution but it affects its stability.

No-flux boundary conditions are applied at the border of the liver. The boundaries are treated according to the level set representation of the liver computed from the segmentation using linear interpolation without requiring advanced meshing techniques. We use a second-order accuracy model for curved walls presented by [Bouzidi 2001] who proposed a simple boundary condition based on interpolation and the bounce back scheme. This method needs to treat the boundary conditions separately for two cases: $\Delta \leq \frac{1}{2}$ or $\Delta > \frac{1}{2}$, i.e either the boundary is closer to the lattice point which is inside the domain or it is closer to the lattice point which is not in the domain. Δ represents the fraction of an intersected link in the domain of interest (Figure 3.10 illustrated the description of Δ) and is computed based on the level-set representation of the liver.




Figure 3.10: Layout of the isotropic Cartesian grid for the implementation of LBM boundary condition with a curved wall boundary.

If the node from which the post collision values travels ($\mathbf{x} - \mathbf{e}_\alpha \Delta x$) is outside the domain, then the value of $f_\alpha(\mathbf{x}, t + \Delta t)$ is:

$$\text{for } \Delta > \frac{1}{2} : f_\alpha(\mathbf{x}, t + \Delta t) = \frac{1}{2\Delta} \hat{f}_{\bar{\alpha}}(\mathbf{x}, t + \Delta t) + \frac{2\Delta - 1}{2\Delta} \hat{f}_\alpha(\mathbf{x}, t + \Delta t)$$

$$\text{for } \Delta \leq \frac{1}{2} : f_\alpha(\mathbf{x}, t + \Delta t) = 2\Delta \hat{f}_{\bar{\alpha}}(\mathbf{x}, t + \Delta t) + (1 - 2\Delta) \hat{f}_{\bar{\alpha}}(\mathbf{x} - \mathbf{e}_{\bar{\alpha}} \Delta x, t + \Delta t)$$

where $\mathbf{e}_{\bar{\alpha}} = -\mathbf{e}_\alpha$.

Finally, we model the heat source term through a Dirichlet boundary condition at the location of the probe. We emulated the RFA protocol by computing a sphere centered at the center of the tumor and with the radius defined pre-operatively by the clinician given the size and the location of the tumor, as illustrated on Figure 3.11.

The temperature of the points inside the sphere is enforced at 105°C for a duration defined also pre-operatively.




Figure 3.11: Scheme showing the modeling of the RFA probe with a sphere centered at the center of the tumor.

3.2.3 Estimation of Anatomical Model from Patient Images

The pre- and post-operative images are semi-automatically segmented, yielding a detailed anatomical model of the patients liver. The user provides different strokes for the background and the foreground of the image, and the strokes can be modified interactively [Criminisi 2008]. Finally the resulting segmentation is manually refined using itk-snap [Yushkevich 2006]. We generate volumetric binary images of the parenchyma, tumors, hepatic veins, vena cava, portal vein, hepatic artery whenever visible (Figure 3.12). The duration of the full segmentation process depends on the extent of the visible vessels. For example, typically for one patient, it took 20 min to segment the liver mask, 10 min for the tumor, 35 min for the vena cava and 15 min for the portal vein, i.e 1 hour 20 minutes in total. A multi-label mask image is created to identify the different structures of this detailed anatomical model of patient's liver and circulation tree as well. To define the computational domain, a level set representation of the liver, without tumor and vessels is computed.

The Decoupled FEM-LBM CFD solver rely on FEM. In this case, a smooth polygonal surface mesh is created for each region. A multi-label mask image is also created to identify the structures of interest for the computation.

Contrarily, the Fully Coupled LBM CFD solver does not rely on polygonal meshes at all, but the porosity needs to be defined everywhere in the computational domain. In this case, we do not generate a surface mesh but instead a porosity map is necessary. From the vessel masks, a porosity map is created to identify the porous parenchyma and the vessels. Because non-visible, the walls of the vessels are extrapolated using 26 connectivities dilatation of the vessels masks. The vessel extremities, which do not have walls are manually identified.




Figure 3.12: A detailed anatomical model of the liver is estimated from a standard clinical CT image. *See text for details.*

3.3 Heating Power and Cooling Temperature Computation

During the heating period, a Dirichlet boundary conditions is used to fix the delivered temperature either on a sphere, whose radius is defined pre-operatively by the protocol followed by the clinician if the position of the probe is unknown, or on the probe model itself if we know the probe position. In this case, RFA is simulated by imposing as input the temperatures measured at the five thermistors and the four remaining tips temperature are linearly interpolated from these values. The heating stops at time $t = t_a$. During this period ($t < t_a$) the heating power can be computed, whereas the cooling temperature can be computed when ($t > t_a$).

3.3.1 Heating Stage

We assume that the measured power is strongly correlated (proportional) to the heat power $P(t)$ delivered through radio-frequency to heat the liver tissue. Proportionality is assumed to account for power dissipation due to electrical resistance, and the unknown surface ratio of the probe being heated. The heat power $P(t)$ delivered to the tissue can be computed at each time step of the simulation from the bio-heat equation according to Fourier's law:

$$P(t) = \alpha \int_S d_t \frac{\partial T(t)}{\partial \mathbf{n}} dS \quad (3.18)$$

where S is the probe surface, \mathbf{n} is the outer normal at that surface and α , the proportionality coefficient ($\alpha = 0.4469$), found by matching the peak value of the measured power for pig 1 with the peak value of the simulated power with personalized parameters. This value is then used in all the computations.

3.3.2 Cooling Stage

In the absence of any delivered heat power, the nine tips of the probe cool down at a speed which depends on the local conductivity d_t and the heat capacity c_t . Thus, during the cooling period $t_c - t_a$ (cooling stops at time $t = t_c$), the cooling temperature can be simulated.

3.4 Model Verification

The computer hardware used in all the experiments was a Windows 7 desktop machine (Intel Xeon, 2.80 GHz, 45GB RAM, 24 CPUs) with a Nvidia Quadro 6000 1.7 GB with 448 CUDA cores.

3.4.1 Quantitative Verification of the Heat Transfer Model




Figure 3.13: Synthetic set-up for the verification against analytical solution: (*Left*): Synthetic set-up used in the quantitative verification of the framework at time $t = 0$ s. (*Right*): Initial temperature distribution along the x-axis of the cuboid domain. See text for details.

3.4.1.1 Experimental protocol

To evaluate the LBM heat transfer solver, its behavior is compared on a regular cuboid domain with an analytical solution. For a source released at \mathbf{x}_0 at time t_0 , the 3D analytical solution of the advection-diffusion equation: $\frac{\partial T}{\partial t} + \mathbf{v} \cdot \nabla T = \nabla \cdot (D \nabla T)$ is:

$$T(\mathbf{x}, t) = \frac{M}{[4\pi(t - t_0)D]^{3/2}} \exp\left(-\frac{\|\mathbf{x} - \mathbf{x}_0 - (t - t_0)\mathbf{v}\|^2}{4D(t - t_0)}\right)$$

To have the same conditions as the RFA computation on a patient's liver, parameters were chosen to get the heat diffusion in a physiological range. We set a Gaussian-shape source of 70°C at the center of the cuboid at time $t = 0$ (Figure 3.13) with

this set of parameters: $D = 0.1 \text{ mm}^2/\text{s}$, $\mathbf{v} = (2, 0, 0) \text{ mm/s}$, $M = 35000 \text{ }^\circ\text{C}\cdot\text{mm}^3$, $t_0 = -50 \text{ s}$. We initialized the temperature values at each point of the domain with the analytical solution at time $t = 0$, then the advection-diffusion equation is solved using our LBM solver. The temperature at a typical point of the domain is reported and compared with the analytical solution values. Neumann boundary conditions were used at the border of the domain, which was chosen to be large enough to get rid of the boundary effect at the probed points.

3.4.1.2 LBM Convergence Analysis

We performed a spatial and a temporal convergence analysis of the solution computed by the LBM solver. The solutions are compared to the analytical solution at one typical point of the domain (Figure 3.14).




Figure 3.14: (*Left*): Spatial convergence analysis for a fixed time step of 0.01 s for one point in the domain described in Figure 3.13. As one can see, the proposed framework quickly converges to the analytical solution of the advection-diffusion equation. (*Right*): The computed solution for a resolution of 1 mm and a time-step of 0.075 s compared to the analytical solution for the same point. The RMS error is reduced to 0.26°C .

Spatial Convergence Analysis First, the time-step is fixed to a constant value and different solutions with different resolutions are computed. As shown in (Figure 3.14, Left), the smaller the spatial resolution, the closer the computed solution is to the analytical one. Quantitatively, the Root-Mean-Square (RMS) errors between the computed solution and the analytical one decreased with the resolution (Table. 3.1).

This analysis confirms qualitatively and quantitatively the accuracy of the implementation of the heat transfer model with LBM.

Temporal Convergence Analysis For a given resolution, an upper and lower bound for the time step were provided by the simulated physics from the

Table 3.1: The Root-Mean-Square (RMS) errors between the computed solution and the analytical with respect to the resolution for a fixed time step of $0.01 s$.

Resolution	RMS
$5 mm$	$11.86^{\circ}C$
$2 mm$	$2.42^{\circ}C$
$1 mm$	$0.55^{\circ}C$
$0.5 mm$	$0.11^{\circ}C$

LBM implementation and the Courant-Friedrichs-Lewy conditions: $|\mathbf{v}\Delta t/\Delta x| \leq 1$ [Courant 1928]. A time-step of $75 ms$ and a resolution of $1 - 2 mm$ appeared to be a good compromise between accuracy and computational cost (Figure 3.14, Right).

3.4.2 Computational Efficiency

A GPU-based version of our model has been developed, relying on the CUDA toolkit (dedicated software for NVIDIA’s GPUs) since LBM is easily parallelizable. In our model, the values of interest are the temperature related to the distribution function and the state (alive, vulnerable, or dead) of the cell computed at each vertex separately. This implementation uses classic parallelization methods, similarly to a CPU parallel approach: each thread is dedicated to one vertex and computes the contribution of the temperature and state of the cell for this vertex ensuring a tiled access in memory. Two distribution functions are actually needed (at time t and time $t - 1$: \mathbf{f}_{past} and $\mathbf{f}_{\text{present}}$), to avoid concurrency reading and writing when one thread writes a vertex value while another thread attempts to read it. The implementation is described in Algorithm 4.

Algorithm 4 Implementation of LBM RFA

- 1: Initialization of each vertex temperature T .
 - 2: Computation of each vertex distribution functions \mathbf{f} .
 - 3: **while** $t < t_{RFA}$ **do**
 - 4: Enforce temperature at the probe points.
 - 5: Compute the corresponding distribution functions \mathbf{f} .
 - 6: Compute $\hat{\mathbf{f}}$ and then \mathbf{f}
 - 7: Compute the cell state and the temperature T .
 - 8: Update τ , c_{ti} and d .
 - 9: Do $\mathbf{f}_{\text{past}} = \mathbf{f}_{\text{present}}$.
-

In order to show the benefit of our GPU-based approach, it is compared against a CPU implementation with multithreading. The same synthetic set-up as described in Figure 3.13 is used. As reported in Figure 3.15, experiments showed a maximum speed-up of 11 with multithreading and 22 threads (OpenMP) and 45 with graphical processing units (GPU) implemented in CUDA with respect to a single-core implementation of LBM. Moreover, after a quantitative verification of the FEM

computation against the analytical solution, experiments showed that a $60\times$ speed-up was obtained with respect to an FEM implementation on CPU for a similar accuracy.




Figure 3.15: (*Left*): Computation time for simple CPU implementation without any parallelization, for multithreading implementation with OpenMP and 22 threads and GPU implementation with CUDA. (*Right*): Computation speed-up with respect to the number of threads used (parallel optimization with OpenMP).

3.5 Discussion

3.5.1 Summary

We have presented two multi-physics models for efficient patient-specific planning of RFA based on medical CT images. In a first approach, we rely on LBM to solve the bioheat equations and the level set representations of the structures are directly computed from images, but this framework still requires advanced meshing techniques to compute the flow in porous media, adding one step in the processing pipeline with the associated computation time, errors and parameters. To solve this issue, we propose a second model, where we rely on LBM not only to solve the bioheat equations and the cellular necrosis model, but also to compute the blood flow simultaneously in the large visible vessels and in the parenchyma. The coupled computation of the porous and hepatic flow eliminates the difficulties in the setting of boundary conditions in the parenchyma, and the occurrence of shear stress on the vessels wall is avoided as the Pennes Model is used in the big vessels where the flow is not accounted for.

3.5.2 Model Limitations

In this study, a two-compartment model was used to describe the heat propagation, coming from two simplifications of the coupled bio-heat equations which are accurate in the cases of small and large vessels respectively, whereas previous

studies suggested that heat dissipation may arise precisely in the medium vessels [El-Serag 2003]. This simplification was motivated by the fact that the liver is highly vascularized with different types of vessels of different diameters. For example, the large hepatic veins have a diameter of 6 mm on average and the capillaries, a diameter of $8 \cdot 10^{-3}$ mm [Peng 2011]. With the resolution of the images used in clinical settings, we cannot segment such small vessels and thus we cannot include them in the current modeling pipeline. Modeling the impact of all vessels is not practical in clinical practice. However, it will be interesting to study the effect of the vasculature on the necrosis, using a synthetic case with a deeply detailed vessel tree for example. Those experiments are necessary to understand the impact of every hepatic blood vessels. It will be interesting to know until which resolution it is necessary to go in order to get an accurate prediction of the ablation extent. Only then computational models could be simplified and used in clinical settings.

3.6 Conclusion

Turning computational models of RFA into clinical practice is a challenging task, which can be beneficial for therapy understanding. Through the use of the Lattice Boltzmann Method, our radiofrequency ablation (RFA) model allows near real time and state of the art computations of heat transfer that are suitable for model-based therapy planning or guidance in the future, even if the target is to go beyond real-time, as we need to stay under 1-2 minutes of computation time for clinical use. Based on our approach, the clinician could test different therapeutical strategies *in-silico*, assess their outcome before the intervention and finally choose the most appropriate therapy for a specific patient. The model has been verified against an analytical solution, evaluation on clinical data is presented in the next chapter.

Evaluation on Clinical Data

Contents

4.1	Introduction	53
4.2	Available Clinical Data	54
4.3	Patient-Specific RFA Computation	55
4.3.1	Experimental Protocol	55
4.3.2	Computation Time	55
4.4	Evaluation on Patient Data	56
4.4.1	Systematic Study	56
4.4.2	Effect of Advection	58
4.4.3	Effect of the Probe Position	60
4.4.4	Effect of the Segmentation	61
4.4.5	Effect of Biophysical Parameters	63
4.5	Discussion	65
4.5.1	Effect of Large Vessels	65
4.5.2	Effect of Advection	65
4.5.3	Effect of the Probe Position	66
4.5.4	Effect of the Segmentation	66
4.5.5	Towards Personalization	66
4.6	Conclusion	67

Based On: the conference paper [Audigier 2013a] and on the extended journal version [Audigier 2015a] for the clinical evaluation of the model on ten patients who underwent radiofrequency ablation (RFA) of hepatic tumors.

4.1 Introduction

In this chapter, we describe the application of the framework described in the previous chapter, on fourteen tumors from ten patients. The Decoupled FEM Porous-Vessels CFD Model is used to compute the blood flow (the work presented here has been realized at the beginning of this work and at that time, the Fully Coupled LBM CFD Model was not introduced yet), the nominal formulation from [ONeill 2011] of the cell death model (with $\alpha = 1$) is used, and different hypotheses are tested in Section 4.4. The comparison with the real outcome extracted from post-operative

data in terms of necrotic area suggests a promising correlation between the predicted and actual ablation extent, but also the importance of considering the blood perfusion inside the parenchyma. The generalization of the biophysical model with respect to tissue properties (i.e. we simulated a RFA procedure by using patient-specific geometry and boundary conditions but generic tissue parameters) shows that reasonable results can be achieved (average values of point-to-mesh distance: 10.17 ± 8.52 mm, DICE: 41.8%, sensitivity: 66.9%, positive predictive value: 38.3%). This study presents the generalization of the biophysical model with respect to tissue properties tested on ten patients data, and an analysis of the perfusion effect on the necrotic area. A first experiment of tissue parameters fitting has been performed on one patient. It highlighted a more accurate prediction power but also the need for personalization.

4.2 Available Clinical Data

Our database is made of CT scans from ten patients who underwent RFA of one or several hepatic tumors. Figure 6.1 illustrates the case of patient 04. For each patient, pre- and post-operative late venous phase CT are available. We generate volumetric binary images of the parenchyma, tumors, hepatic veins, vena cava, portal vein, without the hepatic artery since only single-phase CT images are available. The clinical data were processed off-line. The results were not available to the interventionist during the RFA procedure. This set of patient data come from the Johns Hopkins Hospital in Baltimore, MD, USA.




Figure 4.1: (*Left*) Pre-operative image of patient 04. The tumor is visible. (*Right*) Post-operative image for patient 04, the necrotic area is visible around hepatic vein.

4.3 Patient-Specific RFA Computation

4.3.1 Experimental Protocol

The computation using the RFA model has been described in the previous chapter and is now evaluated on clinical, retrospective data from ten patients, with fourteen ablations (some patient had several tumors ablated) for whom pre- and post-operative CT images were available. In this chapter, the nominal formulation from [O'Neill 2011] of the cell death model (with $\alpha = 1$) is used to compute necrotic extent and the Decoupled FEM Porous-Vessels CFD Model is used to compute the blood flow. For all patients, nominal tissue parameters were employed (Table 2.1). Clinical RFA protocol requires that the probe is deployed within the tumor with a probe diameter defined pre-operatively according to the size of the tumor, and then maintained for 7 minutes after the target temperature of 105°C was reached, as measured by the probe thermistors. For large tumors, the process was iterated with sequentially increasing diameters. In all cases, a single probe placement was utilized (no separate overlapping ablations). After anatomical model extraction, we emulated the RFA protocol by placing the virtual probe at the center of the tumor. Cells around the probe tip within the probe diameter sphere were heated at 105°C during 7 minutes or twice 7 minutes. In all cases, the computation continued for 3 additional minutes without the probe so that each cell reaches a steady state. In order to evaluate the results of the computation, we compared the computed necrotic region with the patient-specific ground truth. For each patient, the lesion is manually segmented by an expert on the post-operative image and then non-rigidly registered to the pre-operative image. The elastic registration from the Advanced Normalization tools (ANTs) is used [Avants 2009] with the vessels and parenchyma binary images employed as landmarks. Indeed, the thermal induced lesion on the post-operative image creates registration issues, especially in the area of interest (Figure 4.2). Finally the accuracy of the registration is visually checked by the expert.

4.3.2 Computation Time

One minute of ablation is computed in almost one minute. In comparison, using FEM on CPUs, the identical process takes around one hour. To the best of our knowledge, this is the first time that near real-time physiological computations of RF ablation has been achieved. A single probe ablation of 7 minutes is computed in around 8 minutes depending on the liver size, while an increased probe diameter ablation of 14 minutes is computed in around 15 minutes.




Figure 4.2: The registration is performed using masks with vessels and parenchyma as landmarks. The post-operative mask is registered (*Right*) to the pre-operative mask (*Left*), with an elastic registration algorithm using ANTS [Avants 2009]. Presented here are the results for patient 10.

4.4 Evaluation on Patient Data

4.4.1 Systematic Study

Qualitatively, computed ablation followed closely the boundaries of the vessels, due to the heat sink effects of the blood (Figure 4.3).




Figure 4.3: Results of the computation for patient 04, the streamlines represent the parenchyma flow and are color-coded with respect to the velocity magnitude. As one can see the lesion computed with advection follows the blood flowing from the portal vein to the sushepatic veins. *See text for details.*

The shape of the ablated area is also dependent on the heat advection through the liver parenchyma (Figure 4.3 and Figure 4.5). Cell death area computed using the model compared qualitatively well with the observed post-operative necrosis zone as we can see on Figure 4.3 and Figure 4.5. Quantitatively, in most of the cases, average point-to-mesh errors were within clinical acceptance as they were significantly lower than the different diameter configurations of the probes (3 to 5 cm), as shown on Figure 4.4. The values are reported in Table 4.1.

Given V_m and V_s , the volume of measured (respectively simulated) necrotic area, then the DICE score ($DICE$), the sensitivity (S) and the Positive Predictive Value (PPV) are defined as:

$$DICE = \frac{2|V_m \cap V_s|}{|V_m| + |V_s|} \quad S = \frac{|V_m \cap V_s|}{|V_m|} \quad PPV = \frac{|V_m \cap V_s|}{|V_s|}$$

Table 4.1: Evaluation of the RFA modeling on patient data (PPV: Positive Predictive Value).

Patient	Probe diameter (cm)	Point-to-mesh error (mm)	DICE (%)	Sensitivity (%)	PPV (%)
01	3	8.24 ± 8.48	54.3	46.0	66.2
02	4 then 5	8.25 ± 5.92	61.7	45.6	95.6
03	4	21.65 ± 16.96	17.7	22.7	14.5
04	4 then 5	10.91 ± 8.67	31.4	89.5	19.0
05-1	4	7.28 ± 7.03	45.0	80.8	31.2
05-2	3.5	6.80 ± 6.25	54.2	44.8	68.5
06	4 then 5	10.77 ± 5.01	37.6	90.7	23.7
07-1	4 then 5	8.34 ± 6.79	40.8	66.4	29.4
07-2	4 then 5	12.92 ± 12.32	45.2	63.8	35.0
07-3	4 then 5	11.61 ± 11.92	40.3	93.6	25.7
08-1	3	7.97 ± 5.41	30.6	32.1	29.1
08-2	3	5.32 ± 4.76	61.6	65.5	58.2
09	3	11.96 ± 8.50	21.3	98.9	11.9
10	4 then 5	10.33 ± 11.20	44.1	96.8	28.5
Mean	-	10.17 ± 8.52	41.8	66.9	38.3

More importantly, in all but one case (patient 01) the computation predicted that the selected protocols completely covered the entire tumor, which is the clinical criterion for ablation planning. The sensitivity is also reasonable (67% on average). The average Positive Predictive Value (PPV) of 38% is low as the values are drastically different for each patient (from 95.6% for patient 02 to 11.9% for patient 09). Some cases presented a computed lesion far from the registered post-operative one. For example, as one can see on Figure 4.12, the computed lesion in red is different and far from the registered post-operative lesion in white for patient 03 (PPV of 14.5%). This may be due to the uncertainty of the actual probe position used in the clinical protocol or to registration errors as discussed in the following section (Section 4.5). Some cases presented a larger necrosis area compared to the ground truth (Figure 4.3) or a smaller one (Figure 4.14, Left). The diffusion coefficient used from the literature was either too high or too low to get a perfect match, as exposed in the following experiment. It may be due to the general state of the liver, which can be cirrhotic or hyperperfused [Van Beers 2001].




Figure 4.4: Box plot of the error on the 14 ablations. (*Left*): Point-to-Mesh error in mm computed on the surface of the lesion. (*Right*): Similarity scores in % computed on the volume of the lesion.




Figure 4.5: Results of the computation for patient 01, the streamlines represent the parenchyma flow and are color-coded with respect to the velocity magnitude. As one can see the lesion computed with advection is closer to the ground truth as it follows the blood flow path from one inlet of the portal vein to the vena cava. *See text for details.*

4.4.2 Effect of Advection

In order to better understand the effect of advection, we perform the computations on the same data and we remove the blood flow in the parenchyma. The evaluation results are shown on Figure 4.6 and the values are reported in Table 4.2. Qualitatively, we can observe that the advection has an impact on the shape, the extent and the size of the lesion (Figure 4.3 and Figure 4.5). As the blood flow in the parenchyma tends to go from the extremities of the portal vein to the hepatic veins (the extremities of the vena cava), the temperature follows the same path. Notably, if the tumor is close to the portal vein, the advection will tend to enlarge the extent of the necrotic region as the blood flow will evacuate the temperature inside the liver parenchyma, whereas if the tumor is closer to the vena cava, the advection will tend to reduce the size of the necrotic core as the temperature will be dissipated in the general blood flow vasculature. Therefore, we can claim that the effect of

the advection is roughly proportional to the signed geodesic distance between the vena cava and the portal vein. Thus the impact of parenchyma perfusion is highly dependent on the tumor location.

Table 4.2: Evaluation of the RFA modeling on patient data without advection.

Patient	Point-to-mesh error (mm)	DICE (%)	Sensitivity (%)	PPV (%)
01	9.03 ± 9.74	52.4	43.7	65.3
02	8.00 ± 5.01	64.6	48.9	95.2
03	24.44 ± 20.91	18.3	22.4	15.5
04	10.57 ± 5.59	28.8	96.6	16.9
05-1	5.56 ± 5.20	51.7	83.8	37.4
05-2	6.19 ± 4.56	54.5	59.8	50.1
06	8.25 ± 5.92	37.9	82.6	24.6
07-1	6.06 ± 3.99	55.3	79.7	42.3
07-2	7.98 ± 5.05	56.8	71.2	47.2
07-3	4.73 ± 3.30	67.8	87.9	55.2
08-1	6.70 ± 4.43	38.8	51.6	31.0
08-2	5.07 ± 4.45	60.5	61.3	59.7
09	6.91 ± 4.65	33.0	97.9	19.9
10	6.76 ± 4.08	51.5	98.8	34.8
Mean	8.30 ± 6.21	48.0	70.4	42.5




Figure 4.6: Box plot of the error on the 14 ablations computed without the advection term. (Left): Point-to-Mesh error in mm computed on the surface of the lesion. (Right): Similarity scores in % computed on the volume of the lesion.

4.4.3 Effect of the Probe Position

The position of the probe is a major unknown in our experiments. We assumed that the clinician put the probe at the center of the tumor (Figure 4.7, Left) but it may not always be the case. To check the sensitivity of the computations to the probe position, a new computation is performed for patient 06. The same configuration was used except that the probe center is at the barycenter of the registered post-op lesion (Figure 4.7, Right), not anymore at the center of the tumor. With this technique, the results are significantly improved (Table 4.3 and illustrated on Figure 4.8). The probe position affects the error measured between the computed necrotic area and the ground truth (Figure 4.9).




Figure 4.7: The 2 different hypotheses used for the probe placement (patient 06). (*Hypothesis 1*): The probe is placed at the center of the tumor = regular hypothesis. (*Hypothesis 2*): The probe is placed at the center of the post-op registered lesion.

Table 4.3: Evaluation of the effect of the probe placement with the two hypotheses described in Figure 4.7.

Patient 06	Point-to-mesh error	DICE	Sensitivity	PPV
Hypothesis 1	10.77 ± 5.01 mm	37.6 %	90.7 %	23.7 %
Hypothesis 2	8.01 ± 6.50 mm	50.0 %	96.3 %	33.8%




Figure 4.8: Evaluation of the effect of the probe placement with the two hypotheses described in Figure 4.7.




Figure 4.9: The different results in red with the different hypotheses are presented. (*Left*): Post-operative lesion registered to the pre-operative image. (*Middle*): Computed lesion with the first hypothesis. (*Right*): Computed lesion with the second hypothesis. The boundary of second lesion is closer the one of the ground truth (patient 06).

4.4.4 Effect of the Segmentation

The segmentation also has an influence on the computed lesion (Table. 4.4 and illustrated on Figure 4.10). To notice the effect of user variability on the segmentation, the expert segmented twice the data of patient 03. This second segmentation was done independently and weeks apart from the first one. Figure 4.11 shows the superposition of the two segmentations. Based on these two segmented datasets, we computed twice the RFA ablation and compared the two computed lesions (Figure 4.12). It can be noted that the main difference between the two computations lies in the CFD results which is really sensitive to the segmentation, especially the segmentation of the vessels inlets.

Table 4.4: Evaluation of the effect of the segmentation on the simulated lesion for patient 03.

Patient 03	Point-to-mesh error	DICE	Sensitivity	PPV
Segmentation 1	21.65 ± 16.96 mm	17.7 %	22.7 %	14.5 %
Segmentation 2	19.07 ± 17.11 mm	22.8 %	28.4 %	19.0%




Figure 4.10: Evaluation of the effect of the segmentation on the simulated lesion for patient 03.




Figure 4.11: Two segmentations of the same patient (patient 03) performed by the same expert several weeks apart. The parenchyma boundaries are similar, but the vessel segmentation differ greatly from one segmentation to the other.




Figure 4.12: The two computed lesions for patient 03, the streamlines represent the parenchyma flow and are color-coded with respect to the velocity magnitude. The total blood flow in the liver is the same (around 25 mL/s) but the parenchyma blood does not flow similarly in both cases due to the differences in the extremities segmentations.

4.4.5 Effect of Biophysical Parameters

The accuracy of the computation depends on the material properties since they are patient-specific, temperature- and space-dependent and they cannot be easily measured *in vivo*. Nominal tissue parameters were employed, with values often based on *ex vivo* experiments on animal tissue sometimes with a large varying range between published studies [Tiesler 2011]. A proper estimation of those parameters is needed but has often been overlooked due to its difficulty. A first step towards the personalization of those parameters is to perform a sensitivity study of the computation to the main parameters. From the results based on nominal parameters for patient 02, we can see that the heat conductivity is too low. For a heat conductivity 8 times greater than the nominal value ($4.096 W(mK)^{-1}$), we manage to have a better outcome (Figure 4.14). Qualitatively, Figure 4.13 shows the variation in term of point-to-mesh error and similarity scores with respect to the diffusivity value. Quantitatively, Table 4.5 shows improvements in the point-to-mesh error (from 8.25 ± 5.92 mm to 3.85 ± 3.21 mm) and in the Dice index (from 61.7 % to 82.3 %).




Figure 4.13: Effect of the diffusivity parameter on the mean and standard deviation of the point to mesh error, the DICE, Sensitivity and PPV scores (patient 02). D x 8 gives the larger DICE score and the smaller point-to-mesh error.

Table 4.5: Estimation of the heat conductivity on patient 02 (PPV: Positive Predictive Value). D x 8 gives the larger DICE score and the smaller point-to-mesh error.

Patient 02 Probe diameter: 4 cm then 5 cm	Point-to-mesh error (mm)	DICE (%)	Sensitivity (%)	PPV (%)
D	8.25 ± 5.92	61.7	45.6	95.6
D x 2	6.97 ± 5.26	68.2	54.1	92.2
D x 4	5.17 ± 4.07	76.3	67.3	88.0
D x 6	4.17 ± 3.38	80.8	77.5	84.4
D x 8	3.85 ± 3.21	82.3	84.8	80.1
D x 10	4.15 ± 3.38	81.3	89.3	74.6
D x 12	4.88 ± 3.87	78.6	92.0	68.6
D x 14	5.83 ± 4.53	75.0	93.6	62.6




Figure 4.14: (Left): computed lesion in red is smaller than the post-op lesion in blue. (Right): A better match is achieved between the computed lesion and the ground truth with adjusted heat conductivity for the same patient (patient 02).

4.5 Discussion

We have evaluated the first multi-physics model proposed for efficient patient-specific planning of RFA based on medical CT images. Despite possible biases in establishing correspondences between the post- to the pre-operative images due to registration or segmentation errors, and the use of nominal biological parameters, which are not patient-specific, our model provided promising results, opening new opportunities for RFA planning and guidance.

4.5.1 Effect of Large Vessels

The computed lesions show that large vessels act as heat sinks as the lesion follows the large vessel's walls. A generic vena cava pressure, inflow and flow through the inlets of the hepatic veins were used for all computations. Since these values were not available, we assumed the pressure and the flow to be a constant in the physiological range of healthy patients. Yet, the pressure and the flow should be personalized as it can vary from patient to patient but also spatially and temporally inside the liver. And the blood flow should instead be modeled as a pulsatile flux. These parameters could be adjusted based on catheter or imaging information such as Phase-Contrast MRI, which can measure the flow in the visible vessels.

This study was performed with a weak coupling of heat propagation and CFD. Change in blood viscosity was neglected as we compute the blood flow only once, at the beginning and then we focused on the heat propagation and cellular necrosis only. A fully coupled computation where the blood viscosity and coagulation are considered could improve the outcome of the RFA model, i.e the extent of ablation and may provide insights on the entire physical mechanisms involved in RFA and tissue properties for long-term therapy prognosis. This aspect of RFA modeling will be investigated in future work.

4.5.2 Effect of Advection

The results of the fourteen ablations computed with the complete model and without the advection term demonstrate that the perfusion in the parenchyma does have an effect on the shape and the extent of the thermal induced lesion, depending on the location of the tumor. Therefore the perfusion in the liver may be an important factor to quantify the extent of ablation and predict therapies. Recent experimental studies on pig livers show an increase in the thermal conductivity with decreased distance to large blood vessels and with the perfusion rate and significant directional differences in thermal conductivity [Podhajsky 2009]. In this study, nominal values were used for the advection coefficient, which control the extent of the heat transport. Yet, this parameter could be estimated and animal data could help to better understand this phenomenon [Cheng 2002] using MRI-thermometry acquired intra-operatively for example.

4.5.3 Effect of the Probe Position

We showed that the probe position is an important unknown and the RFA computation could be improved by adding this information. As the probe insertion during the RF ablation is done under ultrasound guidance, 3D ultrasound images could be used to know the exact position of the probe. Additionally, the simulation of the probe using Dirichlet boundary condition is simplistic, and a geometrical model describing the exact shape of the probe would probably be beneficial for this framework.

4.5.4 Effect of the Segmentation

We saw that the advection and large vessels may have an important effect on the size, shape, and extent of the lesion. More extensive sensitivity analysis of this effect with respect to the segmentation of the vessels should be performed. Modification of the reconstructed anatomy, in particular in the circulation system tree may have an important impact on the parenchyma blood flow and therefore on the advection influence, as shown for patient 03 on Figure 4.12.

In this study, as only late venous phase CT images were available, we could not segment the hepatic arteries, thus the effects of the arterial flow were not considered to predict the extent of ablation. Yet, it is worth noting that the veins account for more than 70 % of the blood inflow of the liver parenchyma [Schenk Jr 1962]. A complete study would benefit from including the hepatic artery inflow, accounting for about one fifth of the hepatic blood inflow, but which comes at a higher pressure and pulsation characteristics compared to the portal vein inflow. If the three phases (arterial, portal and venous) CT images were available, it would be straightforward to include hepatic arteries as well as more venous vessels without any modifications for improved accuracy, our framework being modular.

4.5.5 Towards Personalization

Finally, tissue properties in the liver are spatially-varying and may also vary with the disease-state of the liver (e.g cirrhosis), but they are difficult to estimate from clinical data since they cannot be quantified from CT imaging for instance. In this study, all patients had normal liver parenchyma without cirrhosis or steatosis and heat propagation was modeled with standard tissue parameters reported in the literature for all patients. However our method makes it possible to estimate global patient-specific tissue parameters, which may increase the accuracy of the computation as it is the case in adjusting the heat conductivity for one patient in this study. Moreover, the perfusion of the tumor was not taken into account which can impact its thermal conductivity. Yet, the border zone beyond the margin of the tumor is of greater interest and is considered as regular hepatic tissue. Further investigations and experiments on patient-specific tissue parameters will be necessary to include these new findings into our model.

Preliminary experiments suggest that systematic sensitivity analysis and personalized computations with patient-specific and not nominal parameters are necessary to get a better match with the actual outcome [Audigier 2014a].

4.6 Conclusion

Before deploying model-based therapy planning or guidance in clinical settings, clinical validation is a necessary step. The computational model of RFA needs to be validated on extensive data from a large population dataset. The proposed framework make it possible due to its level of integration. Our system may thus constitute a first step towards clinical application of a RFA computational model.

From a research point of view, the proposed framework paves the way to the quantitative and systematic evaluation of computational models of RFA. There is growing evidence that patient-specific anatomical models are necessary to fully comprehend RFA treatments, in particular in the light of recent *in vivo* studies [El-Serag 2003]. By providing a complete system for patient-specific modeling, our system may be used as input to more comprehensive inverse problem studies, and constitute a useful surrogate tool for RFA planning, potentially improving the outcomes for the patients.

Parameter Estimation For Personalization of Liver Tumor Radiofrequency Ablation

Contents

5.1	Introduction	69
5.2	Available Clinical Data	70
5.3	Mathematical Model of RFA Simulation	70
5.3.1	Forward Model	70
5.3.2	Parameter Estimation	71
5.4	Experiments and Results	72
5.4.1	Evaluation of the CFD Solver on a Synthetic Case	72
5.4.2	Sensitivity Analysis	73
5.4.3	Verification of the Optimization Framework	74
5.4.4	Evaluation on Patient Data	74
5.5	Discussion	79

Based On: the conference paper [Audigier 2014a] for the parameter estimation.

5.1 Introduction

Motivated by the promising results obtained by fitting the diffusivity in the last chapter (section 4.4.5), we present in this chapter, a framework for the personalization of RFA which enables patient-specific parameter estimation in the case of multiple hepatic lesions. The forward model relies on the computational model of heat transfer and cellular death during RFA, presented in Chapter 3 and it is based on anatomies estimated from CT (Section. 3.2.3). The computation model of the blood flow is the Fully Coupled LBM CFD model so the Lattice Boltzmann Method (LBM) is used to compute not only the heat diffusion and cellular necrosis as in [Audigier 2013a] but also blood and parenchyma flow in the liver tissue. This framework is particularly efficient for the personalization as it provides a fast solver and naturally accounts for the flow transition between veins and parenchyma. The

effect of the small vasculature inside the liver which are not visible on the CT data is taking into account.

The model is then personalized based on a first ablation. Our framework is adapted to situation where no temperature map is available. This information is used to plan subsequent ablation(s) of the same or additional lesions to treat for a more refined, targeted ablation that minimizes probabilities of recurrence. This can be validated in case of the ablation of multiple tumors inside the liver, assuming that the parameters are spatially invariant within the same patient.

In Section. 5.4, heat conductivity and porosity were selected as the most sensitive parameters for predicting the necrosis extent. After their estimation on patient-specific data, we demonstrate improved prediction accuracy. We discuss the contributions of this chapter and perspectives in Section. 5.5.

5.2 Available Clinical Data

In this chapter, the personalization method used constrains us to consider only patients who had several tumors ablated. Our database is a subset of the database presented in Section. 4.2, which comes from the Johns Hopkins Hospital in Baltimore, MD, USA. Our database is made of CT scans from three patients who underwent RFA of several hepatic tumors. For each patient, pre- and post-operative late venous phase CT are available. We generate volumetric binary images of the parenchyma, tumors, hepatic veins, vena cava, portal vein, without the hepatic artery since only single-phase CT images are available.

5.3 Mathematical Model of RFA Simulation

5.3.1 Forward Model

The forward model used in this chapter relies on the computational model of RFA, presented in Chapter 3. The simulation of heat transfer and cellular death inside the liver depends on the patient-specific anatomy (Section. 3.2.3). The nominal formulation from [ONeill 2011] of the cell death model (with $\alpha = 1$) is used to compute necrotic extent. The computation model of the blood flow is the Fully Coupled LBM CFD model so the Lattice Boltzmann Method (LBM) is used to compute not only the heat diffusion and cellular necrosis but also the blood flow inside the main vessels and the parenchyma considered as a porous medium (Section 3.2.2). This latter method is based on a computational fluid Dynamics (CFD) solver which incorporates a porous part to deal with the liver parenchyma. The different steps of our method are illustrated in Figure 5.1. The main parameters of the model of heat transfer and cellular necrosis in liver tissue are optimized to match the observed extent of the necrosis.




Figure 5.1: Estimation of the personalized parameters and the forward model (blue: input, green: processes, purple: output).

5.3.2 Parameter Estimation

As parameters in the heat transfer and cellular necrosis equation are customarily taken from the literature, we aim to personalize them given the observed extent of the necrotic region measured post-operatively as temperature maps are not readily available. Most of the parameters are defined as constant whereas the heat capacity c_t and conductivity $dt = \bar{dt} * (1 + 0.00161) * (T - 310)W(mK)^{-1}$ are temperature dependent and therefore spatially distributed [Audigier 2013a]. Using DAKOTA ¹, we first perform a sensitivity analysis to know which parameters mostly influence the volume and the point-to-mesh error [Zheng 2007] of the computed necrosis area. Then, we optimize the most sensitive ones: the heat conductivity and the porosity, as to minimize the average point-to-mesh error between computed and observed necrotic region. To this end, we use a gradient-free optimization method, the Constrained Optimization BY Linear Approximations (COBYLA), which required only a few numbers of forward simulations.

5.3.2.1 Global Sensitivity Analysis: Variance-Based Decomposition

In order to know which parameters of the model mostly influence the output of the RFA simulation, we first perform a global sensitivity analysis using a variance-based decomposition.

Variance-based Decomposition (VBD) is a global sensitivity method that summarizes how the uncertainty in the model output can be apportioned to uncertainty in individual input variables [Adams 2014]. VBD uses two primary measures called the Sobol indices: the main effect sensitivity indice S_i and the total effect indice T_i .

¹<http://dakota.sandia.gov> - multilevel framework for sensitivity analysis

Let us call Y , the given output of the computation (in our case, it is either the volume of the necrosis or the point-to-mesh error between the simulated necrosis and the ground truth). Y is a function of m input variables: $Y = f(\mathbf{x})$. We also consider: $\mathbf{x}_{-i} = (x_1, \dots, x_{i-1}; x_{i+1}, \dots, x_m)$. The Sobol indices can be computed as follow:

$$S_i = \frac{Var_{x_i}[E(Y|x_i)]}{Var(Y)} \quad (5.1)$$

and:

$$T_i = \frac{E(Var(Y|\mathbf{x}_{-i}))}{Var(Y)} = \frac{Var(Y) - Var(E[Y|\mathbf{x}_{-i}])}{Var(Y)} \quad (5.2)$$

The main effect sensitivity index, S_i corresponds to the fraction of the uncertainty in the output, Y , that can be attributed to input x_i alone. The main effect sensitivity index compares the variance of the conditional expectation $Var_{x_i}[E(Y|x_i)]$ against the total variance $Var(Y)$. Whilst, the total effect index, T_i corresponds to the fraction of the uncertainty in the output, Y , that can be attributed to input x_i and its interactions with other variables. It should be noted that: $\sum_{i=1}^m T_i \geq 1$, due to the fact that the interaction effect between two different input variables is counted twice [Adams 2014].

5.4 Experiments and Results

All experiments were executed on a Windows 7 desktop machine (Intel Xeon, 2.80 GHz, 45GB RAM, 24 CPUs) with a Nvidia Quadro 6000 1.7 GB (448 CUDA cores).

5.4.1 Evaluation of the CFD Solver on a Synthetic Case




Figure 5.2: Synthetic case used to validate the CFD solver. (Top Left): A cylindrical mesh with spherical inlet and outlet is used. The boundary conditions are the inflow and the outlet pressure. (Top Right): The velocity field given by the CFD solver in the second case. (Down Left): The porosity field used in the first case. (Down Right): The porosity field used in the second case.

A synthetic case, presented on Figure 5.2, is used to validate the CFD solver. A cylindrical mesh with spherical inlet and outlet is used. Blood flow is set at the inlet and pressure at the outlet. Two cases are considered. The first case presented on (Figure 5.2, Left), simulate a single vessel, with porosity set to 1 inside the vessel, 0.04 in the endothelium and 0.1 outside. In the second case, presented on (Figure 5.2, Right), the porosity map is used to emulate two veins and the liver parenchyma. The mass conservation is verified by comparing the flow through the surface of the "vessel" at $z = 20mm$: Φ_{in} and at $z = 80mm$: Φ_{out} . The computation is run in both cases with a spatial resolution of 0.5 mm and until a steady state is reached (tolerance: 10^{-4}). In the first case, the difference between these two fluxes is 3%, whereas in the second case, 2.8% difference between these two flux is achieved.

5.4.2 Sensitivity Analysis

We want to know the sensitivity of four uncertain parameters of our model: $\bar{d}t$, H , \bar{k}_f , ε on the volume of the computed necrosis but also on the point-to-mesh error between the computed necrosis area and the one computed with the nominal parameters from the literature. To this end, a synthetic case has been setup to speed-up the process (Figure 5.3).




Figure 5.3: Set-up of the synthetic case for the verification of the optimization framework. (*Top*): The cylinder with the porosity field used. The boundary conditions are the output pressure and the input flow. (*Down*): The heat distribution initially applied and the velocity distribution used.

The range of parameters values used [Altrogge 2012] are reported in Table 5.1.

These parameters of interest were modeled with a uniformly distributed uncertainty, and the sensitivity analysis was performed using variance based decomposition to compute the global sensitivity indices (so-called Sobol indices). $\bar{d}t$ has the largest total effect (0.58) as compared with \bar{k}_f , H and ε (0.16, 0.15, 0.43) on the volume of the lesion, whereas ε and $\bar{d}t$ have the same larger total effect (0.37) as compared with \bar{k}_f and H (0.16, 0.35) on the point-to-mesh error with respect to

Notation	Parameter Name	Min - Max
H	convective transfer coefficient	24.4×10^4 48.8×10^4 W (m ³ K) ⁻¹
\bar{d}_t	heat conductivity	0.25 1.24 W(m K) ⁻¹
\bar{k}_f	damage rate coefficient	3.2×10^{-3} 3.4×10^{-3} s ⁻¹
ε	porosity	0.1 0.9

Table 5.1: Ranges of parameters values explored in the sensitivity analysis.

lesion obtained with nominal values. Thoses values are reported in Table 5.2. As it is not reasonable to try to estimate all of these four parameters at once, we decided to estimate only two of them: we chose \bar{d}_t for its effect on the volume, as the nominal value of \bar{k}_f [ONeill 2011] seemed accurate and ε . H was choosen large enough to maintain a constant temperature of 37°C in the CT-visible vessels. We fixed the other parameters to nominal values for the personalization on patient data.

Parameters	Total Sobol indices	
	volume	point-to-mesh error
\bar{d}_t	0.58	0.37
H	0.15	0.35
\bar{k}_f	0.16	0.16
ε	0.43	0.37

Table 5.2: Total Sobol indices on the volume of the lesion and the average point-to-mesh error with respect to lesion obtained with nominal parameters.

5.4.3 Verification of the Optimization Framework

In order to confirm the accuracy of the optimization framework, we considered a synthetic case on a regular cuboid domain (Figure 5.3) where all the phenomena occuring during RFA were present (diffusion, reaction and advection) and with nominal parameters of tissue properties. First we simulated a necrotic area with generic parameters by emulating the clinical RFA protocol: during 7 minutes, we heated at 105°C, the simulation continued for 3 more minutes without heating so that each cell reached a steady state. Then, the main parameters: \bar{d}_t and ε are estimated by minimizing the mean of the point-to-mesh error between the computed necrotic region and the one created at the first step. We managed to obtain the estimated parameters with 6.1% of error on \bar{d}_t and 2.1% on ε in 32 minutes after 36 iterations with a mean of the point-to-mesh error of 10^{-3} mm. The convergence of the COBYLA algorithm is illustrated on Figure 5.4.

5.4.4 Evaluation on Patient Data

We evaluated our model on 3 patients, with 7 ablations (several tumors ablated for each patient) for whom pre- and post-operative CT images were available. Clinical




Figure 5.4: Convergence analysis of the optimisation framework. We use DAKOTA and manage to estimate two parameters (\bar{d}_t and ε) in 32 minutes after 36 iterations with a mean of the point-to-mesh error of 10^{-3} mm.

RFA protocol was simulated: the probe was deployed within the tumor and cells in a diameter defined pre-operatively around the center of the tumor probe tip were heated at 105°C during 7 minutes or 2 times 7 minutes. The diameter and heat duration were iteratively increased according to the size of the tumor. The simulation continued for 3 more minutes without any heat source so that each cell reached a steady state. The parameters were considered spatially invariant within the same patient. To the best of our knowledge, all the proposed computational models of RFA assume numerous simplifications, such as the homogeneity and isotropy of the tissue. All the characteristics (electrical, thermal, etc) are normally considered to be isotropic. Nevertheless, the cell death model locally changes the properties of the tissue, and different parameters are related to different location inside the liver. For instance the heat capacity of the tissue (c_t) is different from the heat capacity of the blood (c_b). Moreover, the way the heat is propagating depends on the location inside the liver. There is a reaction term (H , convective transfer coefficient) in the large vessels, and an advection term in the parenchyma, which depends on the porosity ε , but we consider that they have a constant value.

The parameters are estimated on one tumor by reducing the error with the ground-truth. Then, we computed the cell death area of the other tumor(s) of this

patient with the personalized parameters. The computed cell death area compared qualitatively well with the observed post-operative necrosis zone for tumor located at different place inside the liver, close to large vessels, or on the border. The predicted lesion was manually segmented by an expert and registered to pre-operative image. Figure 5.5 illustrates the results of the personalization framework for each patient on one tumor.




Figure 5.5: Computed necrosis compared qualitatively well with the predicted lesion after personalization on the first tumor of each patient.

Table 5.3: Evaluation of the parameter estimation framework on patient data: 7 tumors from 3 patients are considered.

patient	tumor size	probe diameter	point-to-mesh error (mm)	DICE	Estimated Parameters	
					\bar{d}_t $\text{W}(\text{m K})^{-1}$	ε
1-1	5 cm	4 cm then 5 cm	4.06 ± 2.56	72.0%	0.250	0.0997
1-2	3.5 cm	4 cm then 5 cm	4.65 ± 3.52	74.9%		
1-3	4.2 cm	4 cm then 5 cm	4.62 ± 3.32	69.0%		
2-1	1.5 cm	4 cm	2.57 ± 1.89	77.2%	0.275	0.1028
2-2	1 cm	3.5 cm	5.66 ± 4.25	60.9%		
3-1	1 cm	3 cm	3.01 ± 2.05	74.0%	0.489	0.1
3-2	1 cm	3 cm	3.64 ± 2.89	74.1%		
Mean	-	-	4.03 ± 2.92	72.4%	-	-

Quantitatively, average point-to-mesh errors (Table 5.3) were within tolerance in clinical routine for the four tumors estimated with the personalization of the main biological parameters, as the probes can be deployed in steps of 1 cm. Figure 5.6 shows the error in term of point-to-mesh errors and Dice scores for the 7 ablations simulated with personalized parameters. The estimated heat conductivities were lower than the nominal value ($0.512 \text{ W}(\text{m K})^{-1}$), whereas the porosity was very close (0.1), as illustrated on Figure 5.7.




Figure 5.6: Box plot of the error on the 7 ablations with personalized parameters. (*Left*): Point-to-Mesh error in mm computed on the surface of the lesion. (*Right*): DICE score in % computed on the volume of the lesion.




Figure 5.7: Comparison of the parameters (the porosity ϵ , and the heat conductivity \bar{d}_t) personalized on three patients against the nominal values from the literature.

Patient 2 presents a quite large Dice difference between the two cases. It might be due to segmentation or registration issues, and potentially to model limitations (assumptions made, etc.), which will be further investigated in a pre-clinical setup.

The main parameters are first considered with nominal values taken from the literature and then optimized to match the observed extent of the necrosis. Current errors can be explained by segmentation and registration processes but also by the limited number (2) of personalized parameters.

Other experiments on patient 1 showed also a significant improvement of the correlation between predicted and actual ablation extent compared to the prediction using only nominal parameters (average point-to-mesh errors of 4.44 mm vs 4.98 mm, average Dice score of 72.0% vs 68.5%). Those results are illustrated on Figure 5.8, Figure 5.9 and Figure 5.10. The comparative results with simulations with nominal parameters for patient 1 are reported on Table 5.4 and Table 5.5.




Figure 5.8: Parameter Estimation on the first tumor of the first patient (patient 1-1). (*Left*): The post-operative lesion appears around the tumor. (*Middle-Left*): Computed necrosis with personalised parameters. (*Middle-Right*): Overlay of the computed lesion with personalised parameters and ground truth. (*Right*): Overlay of the computed lesions with nominal parameters, with personalised parameters and ground truth. In each case, the blood flow in the area of interest is represented by streamlines in white.

patient	tumor size	probe diameter	point-to-mesh error (mm)	DICE (%)	\bar{d}_t ($W(m K)^{-1}$)	ε
1-1	5 cm	4 cm then 5 cm	4.8 ± 3 4.1 ± 3	67 72	0.512 0.250	0.1 0.0997

Table 5.4: Parameter estimation on the first tumor of the first patient (patient 1-1). The values in black come from computation with nominal parameters, and the values in red from computation with personalized parameters. The nominal values of \bar{d}_t and ε are reminded in black, the estimated values for this specific patient are in red.

patient	tumor size	probe diameter	point-to-mesh error		DICE	
1-2	3.5 cm	4 cm then 5 cm	4.9 ± 3 mm	4.7 ± 4 mm	73%	75%
1-3	4.2 cm	4 cm then 5 cm	5.3 ± 4 mm	4.6 ± 3 mm	66%	69%

Table 5.5: Evaluation of the parameter estimation of patient 1. The parameters estimated on the first tumor (patient 1-1) reduce the point-to-mesh error and increase the DICE score (values in red) between the computed lesion and the ground truth, compared to the lesion computed with nominal parameters (values in black) on the two other tumors for the same patient (patient 1-2 and patient 1-3).




Figure 5.9: Evaluation of the parameter estimation on the second tumor of the first patient (patient 1-2). (*Left*): The post-operative lesion appears around the tumor. (*Middle*): Computed necrosis with personalized parameters. (*Right*): Overlay of the computed lesion with personalized parameters and ground truth. In each case, the blood flow in the area of interest is represented by streamlines in white.

5.5 Discussion

The personalization of the sensitive tissue parameters allows to have a better estimation of the necrosis and to predict the outcome of RFA in case of multiple tumors inside the liver. As we are using the Fully Coupled LBM CFD solver, our framework totally rely on LBM, no advanced meshing techniques are required. All the computations are directly done from patient images: heat propagation and cell death modeling as well as the heat sink effect of blood vessels and porous circulation in the liver. The current method needs several tumors for validation and is worth using only when no temperature maps are available, but it could be easily translated into clinical settings. Adaptation for RFA under image-guidance is considered: RFA procedure is usually done in several steps (increase in probe diameters for example). An intra-operative image is acquired at the end of the first step and used to personalize key parameters, providing a powerful guidance tool. No post-operative images are required. A necessary step before deploying this method in clinical settings is a pre-clinical validation with extensive data on larger populations to evaluate the




Figure 5.10: Evaluation of the parameter estimation on the third tumor of the first patient (patient 1-3). (*Left*): The post-operative lesion appears around the tumor. (*Middle*): Computed necrosis with personalized parameters. (*Right*): Overlay of the computed lesion with personalized parameters and ground truth. In each case, the blood flow in the area of interest is represented by streamlines in white. One area of the post-operative lesion, circled in white, is not covered by the computed lesion.

computational model of RFA and to consider potential safety issue of the proposed application. Even if promising results are achieved with the use of patient-specific parameters, the impact of possible biases in the post- to the pre-operative image registration like the impact of the average simulation of the probe need to be investigated as well as the sensitivity of the results with respect to segmentation and registration.

Pre-Clinical Study: Protocol and Data Analysis

Contents

6.1	Introduction	81
6.2	Available Pre-clinical Data	82
6.3	Pre-clinical Experiment	82
6.3.1	A Pre-clinical Study for RFA Model Validation	82
6.3.2	Experimental Set-up	84
6.4	Computational Experiment	85
6.4.1	Overview	85
6.4.2	RFA Modeling	86
6.4.3	Numerical Computation using LBM	88
6.4.4	Parameter Estimation from Probe Measurements	88
6.5	Data Pre-processing	90
6.5.1	3D Modeling of the RFA Probe	91
6.5.2	Segmentation	92
6.5.3	CFD Computational Domain	92
6.5.4	Post-To-Pre Registration	95
6.6	Conclusion	96

Based On: the conference paper [Audigier 2015b] and on an extended journal version under preparation for the pre-processing of the pre-clinical data.

6.1 Introduction

In this chapter, we introduce a pre-clinical study for the validation of RFA model, based on pre-, intra- and post-operative data.

The sophisticated experimental protocol is described first (Section 6.3), followed by the RFA model we aim at validating. The computational model relies on subject-specific anatomies estimated from Computed Tomography (CT) (Section 6.4). Heat transfer is computed according to the bio-heat equation (Eq. 2.4) and the Fully Coupled LBM CFD Model is used: a Computational Fluid Dynamics (CFD) solver incorporates a porous part to deal with large vessels and the liver parenchyma

hepatic blood flow simultaneously as described in Section 3.2.2. Moreover, this is coupled with a novel cellular necrosis model.

Finally, we present the data analysis required before using the information from the pre-clinical images into the model (Section 6.5).

Section 6.6 concludes this chapter.

6.2 Available Pre-clinical Data

Our database is made of CT scans, MR images, invasive and device-based measurements from five pigs who underwent RFA of one or several hepatic surrogate tumors. Figure 6.1 illustrates the case of Pig 5. For each pig, pre-, intra-, and post-operative images are available. We generate volumetric binary images of the parenchyma, tumors, hepatic veins, vena cava, portal vein and the hepatic artery since three phases CT images are available. The pre-clinical data were processed off-line. The results were not available to the interventionist during the RFA procedure. This set of swine data has been undertaken at the IHU of Strasbourg in France.




Figure 6.1: (*Left*) Venous phase of the pre-operative CT of Pig 5. The surrogate tumor is visible, as well as the gall bladder. (*Right*) Post-operative MRI T2 for Pig 5, the necrotic area is visible around the surrogate tumor, as well as the gall bladder.

6.3 Pre-clinical Experiment

6.3.1 A Pre-clinical Study for RFA Model Validation

A comprehensive animal experiment has been realized for model validation and parameter identification including several modalities at pre-, intra- and post-operative stages. Figure 6.2 illustrates the available pre-clinical data (in blue), the different pre-processing steps needed (in red, green and white) before performing the computation (in grey).




Figure 6.2: Pipeline presenting the available clinical data in blue, the pre-processing needed to use those data (segmentation in green, registration in red), the different computations done in grey and finally a possible personalization framework.

We have developed this pre-clinical study specifically for the validation of the model developed in chapter 3. We have seen that this model is based on patient images. For this reason a pre-operative CT scan of the liver with the implanted surrogate tumors is acquired in order to get an accurate representation of the anatomy. We showed in chapter 4 that the large vessels play an important role in the shape of the lesion, so we acquired pre-operative CT images with the three available phases including portal, venous and arterial phases. One limitation of the proposed model is the probe representation and its location, as described in section 4.5.3. To tackle this issue, a CT scan of the probe alone was first acquired to get an accurate model of the RFA probe, and then an anatomical MR image was acquired after the insertion of the probe to know exactly its position. The boundary conditions of the CFD solver were fixed with nominal flow and pressure values coming from the literature, as described in section 4.5.1. To get a more personalized computation, Phase-Contrast MR images were acquired to get a subject-specific blood flow and pressure measurements from catheter were performed to get invasively the subject-specific

pressure. The clinical protocol employed for each specific patient was unknown, so in this study, we decided to register for each ablation procedure the delivered power and the temperature distribution at the thermistors provided by five tips of the probe. Finally, as the evaluation of the model relies on the comparison with the actual necrotic lesion segmented on the post-operative image, we decided to acquire post-operative images of the lesion two days after the intervention.

6.3.2 Experimental Set-up

The present experimental study (No. 38.2014.01.063) received approval from the local Ethical Committee on Animal Experimentation. All animals used in the experimental laboratory were managed according to French laws for animal use and care and according to the directives of the European Community Council (2010/63/EU).

Pigs are considered as a relevant animal model as their hepatic system is similar to the human one. Several surrogate tumors (diameters $< 3\text{cm}$) are inserted on five swines at various locations of the liver (close to vessels or the Glisson capsule) under ultrasound (US) guidance (SIEMENS ACUSON S3000), followed by the acquisition of pre-operative CT images including portal, venous and arterial phases (CT scanner: SIEMENS Somatom Definition A5). The surrogate tumors are made of a specific gel (a mix of biocompatible gelatin, alginate and nanoparticles) which exhibits an hyper-intense signal in CT and MRI as illustrated in Figure 6.3.




Figure 6.3: Two CT images of Pig 1 before and after surrogate tumors implantation. (*Left*): Before tumor implantation; (*Right*): After tumor implantation.

The surrogate tumors are mainly made of gelatin, whose thermal properties can be found in the literature. In [Culjat 2010], the gelatin density is reported to be 1050 kg.m^{-3} , which is close to the value used in this work (1060 kg.m^{-3} for blood and tissue density). In [Dunmire 2013], the gelatin heat capacity is reported to be $3.6 \text{ kJ}(\text{kg K})^{-1}$, which is exactly the value used for the heat capacity of healthy tissue in this work. However, the reported value of conductivity is approximately $0.303 + 0.0012 (T - 310) \text{ W}(\text{mK})^{-1}$ if T is in Kelvin, which differs from the value used: $0.512 (1 + 0.00161 (T - 310)) \text{ W}(\text{mK})^{-1}$. In order to know how the surrogate nature of the

”tumors” impacts the heat transfer and subsequent ”necrosis”, it could be possible to run the experiments by setting this specific value at the points belonging to the surrogate tumor. However as we used the conductivity value of the tissue for the tumor cells in the clinical study (Chapter 4), similarly we kept this assumption in the pre-clinical study.

Invasive catheter is introduced through the jugular vein to get the free and the wedge pressure in a subhepatic vein and in the vena cava. An MR-compatible RFA probe, the radiofrequency interstitial tumor ablation (RITA) probe (StarBurst RFA, AngioDynamics; www.angiodynamics.com), is deployed at 2 cm of diameter (the diameter of the area defined by the tips of the probe is 2cm) under US guidance (Figure 6.4) next to the targeted surrogate tumor. An MR image is then acquired to get the position of the probe in the liver (SIEMENS Magnetom Aera 1.5T) and flow data. The temperature and delivered power are monitored and recorded intra-operatively during and after the ablation. Finally, a post-operative CT, T2 or T1 + gadolinium MRI is acquired two days after the ablation to assess the extent of the necrotic areas. Overall, pre-, intra- and post-operative images are available, along with interventional device measurements (Figure 6.2). To the best of our knowledge, no such validation set-up has been reported previously in the literature.




Figure 6.4: (*Left*): Photo of the probe inside the pig liver; (*Right*): U.S image of the probe next to the surrogate tumor. The probe diameter is shown in red, the tumor diameter in green.

6.4 Computational Experiment

6.4.1 Overview

Model validation can be performed, as it is done in chapter 4 through a comparison between the computed necrotic extent and the actual lesion observed on the post-operative image. We will proceed as in chapter 4, where all the computations

are performed on the pre-operative image, and the accuracy of the computation is evaluated by a comparison with the post-operative image. We rely on the RFA computation described in chapter 3 to compute the heat transfer from the bio-heat equations (Eq. 2.4). To compute the hepatic blood flow, the Fully Coupled LBM CFD Model described in Section 3.2.2 is used as we want to obtain simultaneously the blood flow in the large vessels, in the arteries and in the parenchyma. Chapter 5 also showed the importance of patient-specific parameters. In this chapter, we rely on intra-operative data (the delivered power and the probe temperature) to show that subject-specific parameters can be estimated.

6.4.2 RFA Modeling

6.4.2.1 Blood Flow Computation

The Fully Coupled LBM CFD Model described in Section 3.2.2 is used to compute simultaneously the flow in the main vessels, in the arteries and in the parenchyma. This model depends on the porosity coefficient ε (fraction of blood volume over the total volume) defined in this chapter through a porosity map (Figure 6.13) as described later in the next section (section 6.5.3.4).

6.4.2.2 Boundary Conditions

This method makes the boundary conditions simple to treat: no boundary conditions are fixed on the extremities of the vessels inside the parenchyma thanks to the use of the porosity map, contrary to [Audigier 2015a]. At the border of the liver, no flux boundary conditions are used (Neumann) whereas Dirichlet boundary conditions are applied at the inlets of portal vein, vena cava and hepatic artery and at the outlet of the vena cava and hepatic artery. Pre-operative Phase-Contrast MR images are used to set the portal vein, vena cava and arterial inflows, φ_p , $\varphi_{vc_{in}}$ and φ_a , whereas invasive pressure measurements through catheter are used to set the vena cava and arterial outlet pressure p_0 and p_a . Figure 6.5 illustrates boundary conditions on a subject-specific geometry.

6.4.2.3 Heat Transfer Model

In this chapter, the heat transfer model described in chapter 3 is used. It depends on the patient-specific anatomy (the segmentation process is presented in Section 6.5.2) and on the blood flow inside the main vessels and the parenchyma considered as a porous medium (Section 6.5.3.4). The Pennes Model is solved in the large visible vessels and the WK model in the liver parenchyma. Nominal values of parameters reported on Table 6.1, are used.

6.4.2.4 Adapted Formulation of Cellular Necrosis Model

The three-state model described in section 3.2.2.2 is used to compute tissue necrosis. Unlike in [ONeill 2011], the constant α is different than 1 ($\alpha = \bar{\delta}/\bar{\beta}(1 + 10N)$) to




Figure 6.5: Model of the hepatic circulatory system. Black arrows (resp. grey arrows) denote blood flow (resp. outlet pressures) which are fixed by the boundary conditions. *See text for details.*

Table 6.1: Nominal parameter values.

Notation	Parameter Name	Nominal
T_k ($^{\circ}\text{C}$)	parameter of cell death model	40.5
$\bar{\beta}$ (s^{-1})	damage rate coefficient	3.3×10^{-3}
γ (s^{-1})	recovery rate coefficient	7.7×10^{-3}
$\bar{\delta}$ (s^{-1})	vulnerable rate coefficient	3.3×10^{-3}
c_t^V ($\text{J}(\text{kg K})^{-1}$)	heat capacity of vulnerable tissue	3.6×10^3
c_t^N ($\text{J}(\text{kg K})^{-1}$)	heat capacity of necrotic tissue	0.67×10^3
c_t^U ($\text{J}(\text{kg K})^{-1}$)	heat capacity of undamaged tissue	3.6×10^3
\bar{d}_t ($\text{W}(\text{m K})^{-1}$)	heat conductivity	0.512
R ($\text{W}(\text{m}^3 \text{K})^{-1}$)	reaction coefficient	27.1×10^4
α_v	advection coefficient	0.11

decouple the damage rate coefficient from the vulnerable rate coefficient, so that three distinct transition rates are considered to allow cells to reach the vulnerable state. Figure 6.6 represents the solution of Eq. 3.11 at one vertex over time if a constant temperature of 105°C is applied.




Figure 6.6: Cell state evolution over time when tissue is heated at 105°C during 10 minutes with the adapted cellular necrosis model proposed in this chapter.

6.4.3 Numerical Computation using LBM

For a time step of $\Delta t = 0.5$ s and a spatial resolution of $\Delta x = 0.5$ mm, better than real-time computation can be achieved on a desktop machine (Windows 7, Intel Xeon, 3.30GHz, 16GB RAM, 12 CPUs, Nvidia Quadro K5000 4.0 GB). For example, it took 577 seconds to compute 1000 s of ablation of tumor 4-1.

6.4.4 Parameter Estimation from Probe Measurements

During the intervention, the delivered power and the temperature distribution are measured by the ablation probe itself. We explore how this can be used to estimate the parameters of the model. During the heating phase, the simulated heat power P_s can be compared to the measured one P_m . During the cooling phase, the simulated temperatures \mathbf{T}_s can be compared with the measurements \mathbf{T}_m , read from five tips of the probe (four tips do not have any thermistors).

In this study, the parameters in the heat transfer and cellular necrosis equation come from the literature. After a sensitivity analysis reported in the next section 7.4, we choose to estimate the heat capacity c_t^U and the constant part of the conductivity \bar{d}_t as they mainly influence the delivered power, the temperature distribution and the size of ablated regions.

As temperature maps are not readily available, these two parameters are personalized from probe measurements (temperature and power) by minimizing the

following cost function:

$$\arg \min_{c_t^U, d_t} \left(\frac{1}{t_a} \sum_{t=0}^{t_a-dt_m} \frac{(P_m(t) - P_s(t))^2}{\sigma_{P_m}^2} + \frac{1}{t_c - t_a} \sum_{t=t_a}^{t_a+t_c} \frac{(m_{T_m}(t) - m_{T_s}(t))^2}{\sigma_{T_m}^2} \right) \quad (6.1)$$

where dt_m is the time step of the measurements, σ_{P_m} and σ_{T_m} are the standard deviations associated with the heat power and the temperature, both of them evaluated from the variability in the available observations (equal to 13.3 W and 5.1 °C in our experiments). σ_{P_m} is the standard deviation of the power of all the twelve ablations at one time point during the heating phase ($t = 50$ s), which acts as a normalization factor in Eq. 6.1. We chose $t = 50$ s to define σ_{P_m} since all the power curves are coherent at that time. Moreover, as illustrated in Figure 6.7, this time is a good compromise since it is not too early, avoiding the boundary conditions effect, and not too late, before the cell death could affect the delivered power.




Figure 6.7: Plot of the power delivered for the 12 ablations for the first 200 seconds. The mean and standard deviation are computed at time $t = 50$ s.

Figure 6.7 represents the power delivered for the 12 ablations for the first 200 seconds. However, the applied RFA protocol is not exactly the same for all ablations. Eight ablations are performed through several short cooling and heating periods, in this case, the delivered power is reduced during the different cooling stages. On the other hand, the other four ablations included only one long final cooling stage after a continuous heating period (red, blue, cyan and green curves in Figure. 6.7). Figure. 6.8 shows the four cases whose the delivered power over ablation is a long heating phase followed by a cooling stage, and one typical case whose ablation is performed through several short cooling and heating periods (fig 5 - tumor 1).




Figure 6.8: (*Left*): Delivered power for the four ablations included only one long final cooling stage after a continuous heating period; (*Right*): Delivered power for pig5 - tumor 1, for which ablation was performed through several short cooling and heating periods. In both case, the grey area represents the domain considered in Figure. 6.7

σ_{P_m} is defined as detailed here:

$$m_P = \frac{1}{12} \sum_{i=1}^{12} P_i(t = 50s)$$

$$\sigma_{P_m}^2 = \frac{1}{12} \sum_{i=1}^{12} (P_i(t = 50s) - m_P)^2$$

σ_{T_m} is the standard deviation of the temperature computed from one ablation during the cooling phase (we did not use all the twelve ablations as their cooling stage did not start at the same time), where all the temperature curves at the 5 tips of the probe are coherent. It is defined as follow:

$$m_T(t) = \frac{1}{5} \sum_{i=1}^5 T_i(t)$$

$$\sigma_{T_m}^2 = \frac{1}{t_c} \sum_{t=t_a}^{t=t_a+t_c} \left(\frac{1}{5} \sum_{i=1}^5 (T_i(t) - m_T(t))^2 \right)$$

To cope with the uncertainty in the rotation of the probe along its axis, the mean tip temperature $m_T(t)$ is used for the personalization instead of directly mapping the tip temperatures.

6.5 Data Pre-processing

The model is based on a representation of the anatomy obtained from pre-operative images. It takes as inputs the geometry and the location of the probe and as

boundary conditions blood flows and pressures. Finally the validation of the model outcome is based on a comparison with the actual lesion observed post-operatively. In order to get this information, data pre-processing is required. Segmentations of the RFA probe and of subject-specific anatomy coming from the pre- and post-operative images are needed. The position of the RFA probe is known from an intra-operative anatomical MR image, but its registration to the pre-operative image is not straightforward, given the artifacts induced by the RFA probe during the acquisition. Moreover, the segmentations of the visible vessels come from anisotropic images. To get stable solution of the CFD solver, smooth vessel trees have to be generated. Since the acquisition process of the Phase-Contrast MR images was complex, pre-processing of the images was necessary to get coherent settings of the boundary conditions. Finally the registration of the post-operative image to the pre-operative image was extremely challenging. It required the use of a robust and advanced registration algorithm. We present here one registration method which gave the most significant results out of several methods which have been tried.

6.5.1 3D Modeling of the RFA Probe

The position and the geometry of the probe are inputs of the RFA model, thus we need to extract these information from images. The geometry of the probe deployed at 2 cm is acquired from a CT image (resolution: 0.2x0.2x0.9 mm) of the probe alone. A 3D mesh is then reconstructed by thresholding the image intensity (Figure 6.9, left panel), and manually registered to the pre-operative CT using the main axis of the probe and intra-operative MR data.




Figure 6.9: (*Left*): Mesh model of the probe with the 9 tips derived from a CT image of the probe only; (*Right*): Anatomical MRI with the MRI-compatible probe implanted in the liver. Four tips of the probe are visible.

The manual rigid registration is done using Paraview ¹ and is visually checked by an expert up to a rotation along the probe axis while the main axis is clearly visible as well as three or four tips (depending on the ablation considered) but the MR resolution does not allow to distinguish between the nine tips of the probe individually (Figure 6.9, right panel).

6.5.2 Segmentation

As each structure of interest is needed as input of the RFA computational model, the segmentation of patient images is a determinant task. From the pre-operative CT data, the following anatomical and pathological regions are segmented semi-automatically and meshed by the tools of Visible Patient, Strasbourg, France (Figure 6.10): parenchyma, hepatic veins, vena cava, portal vein, hepatic arteries and all tumors. These regions are then used to define the computational domain. The meshes are rasterized and a multi-label mask image is created to identify the different structures. To define the computational domain, a level set representation of the liver without tumor and vessels is computed. From post-operative CT or MR data, necrotic areas are segmented and meshed as well.




Figure 6.10: Segmentation of the liver in magenta, arterial vessels in red, portal and hepatic networks in light blue and dark blue respectively, surrogate tumors in dark green and gall bladder in light green. (*Left*): Overlay of the segmented areas on pre-operative CT image. (*Middle*): Creation of the 3D model. (*Right*): Vessels and liver parenchyma.

6.5.3 CFD Computational Domain

6.5.3.1 Vessel Trees

The pre-operative images from which the vessels are segmented have anisotropic resolution. To avoid unstable solution of the CFD solver, smooth vessel trees have to be generated from the semi-automatic segmentations. To this end, centerlines are extracted from each vessel segmentation using VMTK ² (Figure 6.11, left panel). Along each branch of the centerline, the mean radius is computed and the smooth

¹<http://www.paraview.org/>

²The Vascular Modeling Toolkit website, www.vmtk.org.

vessel tree is generated by overlaying cylinders having this mean radius in a piecewise fashion (Figure 6.11, right panel). All those cylinders are rasterized on a single image.




Figure 6.11: (*Left*): Centerlines extraction from the vessel segmentation. The segmentation gives non-smooth vessels due to the anisotropy and the low resolution of the input image. (*Right*): Smooth vessels generated from the centerlines.

6.5.3.2 Blood flow from Phase-Contrast MRI

The computational model of RFA requires as inputs the blood flow entering the vena cava, the portal vein and the hepatic artery. Instead of fixing nominal values from the literature, several Phase-Contrast MR images have been acquired in order to impose subject-specific values as boundary conditions.

2D+t Phase-Contrast MRI have been acquired pre-operatively before the probe implantation. For each pig, several 2D+time sequences are acquired at different time points, for the reproducibility of data, but also at different locations: at the inlet of the hepatic artery, at the inlet of the portal vein and at the inlet of the vena cava. Given one 2D+t Phase-Contrast MRI sequence acquired at one time point, the user places a single seed in the vessel of interest in the first image and then an automatic method is used to segment this vessel on each 2D slice (Figure 6.12, right panel) [Gulsun 2006, Guetter 2011]. Using the vessel area defined by the segmentation, the mean blood flow can be computed at each time. If N 2D+t sequences are acquired at the same location but at N different time points, this is done N times and the RMS of the mean of those N curves is used. The acquisition process of flow is complex to handle and requires a significant learning curve. Due to the non-reproducible and non-coherent measures on the first four pigs, the measures of Pig 5 are used for the five pigs in all the computations (Figure 6.12). We assume that flow would not vary so much from pig to pig. It is a reasonable assumption, since it was demonstrated that liver weight and hepatic blood flow in all mammalian species could be readily related to body weight by a simple equation [Boxenbaum 1980]. The 5 pigs used in the experiments were all healthy, of similar age (3-4 months) and

weight (33, 29.6, 26.7, 25, and 28.4 kg), that is why we assumed that the blood flow would not vary so much from pig to pig.

For Pig 5 (vena cava), the acquisition has been done at 3 different time points ($N = 3$). The RMS of the mean of those three curves is used (Figure 6.12, left panel) since numerical results have shown that pulsatile velocity profile in large blood vessels has little difference in effect on the thermal lesion region of tissue compared with uniform or parabolic velocity profile [Horng 2007].




Figure 6.12: (Left): Velocity in the vena cava of Pig 5 from Phase-Contrast MRI with respect to trigger time (all measures are registered on the cardiac pulse). In cyan, the mean of different measurements at different times (red, blue, green curves). The RMS of this curve (black curve) is then used as boundary conditions in the computation as φ_{vcin} . (Right): Phase-Contrast MRI of Pig 5: the vena cava is segmented in green, the portal vein in blue and the aorta in yellow.

6.5.3.3 Pressure from invasive measurements

The computational model of RFA requires also vena cava and hepatic artery blood pressures as inputs. In order to impose subject-specific values as boundary conditions, these blood pressures have been measured to avoid the use of nominal values from the literature.

The pressures at the outlet of the vena cava and the hepatic artery have been measured invasively by catheter introduced through the jugular vein. The same values (from Pig 5) are used as boundary conditions in the five pigs, as it is the case for the blood flow measurements.

6.5.3.4 Porosity Map

The porosity has to be defined everywhere in the computational domain as it is an input of the CFD solver. The vessel walls are defined as follow. The smooth vessel trees are rasterized on a single image. On this image, we perform a 26-connectivity dilation on the voxels of the vessels to model the endothelium and avoid the flow

to go through the vessel wall. The extremities are detected using the centerlines previously used and the porosity at the extremities is set to the parenchyma porosity value. Figure 6.13 shows an example of porosity map. The porosity value is 1 in the CT-visible vessels, 0.1 in the porous parenchyma [Payne 2011], modeled then as a porous medium. Experiments are performed to obtain a sufficiently small porosity (0.04) in the vessel walls (impermeable medium) to avoid the occurrence of shear stress on the vessel walls (leakage).



Figure 6.13: (*Left*): The porosity map of Pig 1 created from the vessels segmentation. The porosity is 1 in the vessels (in white), 0.04 in the endothelium (in black) and 0.1 in the parenchyma. (*Right*): Zoom on the porosity map. The extremities of the vessels are detected using the centerlines so that the blood flow can go through the vessel extremities.

6.5.4 Post-To-Pre Registration

Due to ethical reasons, CT with contrast agent could not be acquired two days after intervention, making accurate post- to pre-operative registration challenging. In order to compare the results of the computation with the ground truth given by the post-operative images, registration of the post-operative image to the pre-operative image is performed. For each pig, the pre-operative image (CT from different phases) does not necessarily belong to the same modality as the post-operative image (CT or MRI (T2 or T1 + gad)) due to individual experimental changes. Moreover, the pigs are neither in a similar position nor a similar condition, as they had an empty stomach the day of the intervention, which was not true two days after. Breathing has also an important impact on the shape of the liver. For those reasons, we choose to register the post-operative meshes to the pre-operative image instead of pure image-to-image registration. From the semi-automatic segmentation of the

post-operative image, a tetrahedral mesh is generated using CGAL³. From the segmentation of the pre-operative image, a binary mask of the liver including vessels and arteries is generated. Using the SOFA framework⁴, the post-operative mesh is deformed elastically to match the intensity profile in the pre-operative image using the Finite Element Method (FEM) and a co-rotational model [Faure 2012].




Figure 6.14: (*Left*): Superposition of the post-operative mesh model in red on the pre-operative binary image; (*Right*): Comparison between the registered post-operative vessel mesh and the pre-operative one.

First-order implicit Euler time integration is employed and the system of equation is solved with the conjugate gradient algorithm. Figure 6.14 shows the pre-operative binary image with the surface of the post-operative mesh before the registration (Figure 6.14, Up) and after the registration (Figure 6.14, Down).

The computed deformation field is then applied to the necrotic surface mesh using a barycentric mapping between the coordinates of the surface mesh and the tetrahedral mesh. The registration is fast to perform. As an example, it took 110 seconds for Fig 5 to register the post-operative mesh to the pre-operative image.

6.6 Conclusion

In this chapter, we presented a sophisticated pre-clinical set-up to validate a complete and complex multi-physics model of radiofrequency ablation (RFA). This comprehensive validation is based on pre-, intra-, post-operative images and device-

³www.cgal.org.

⁴www.sofa-framework.org.

based measurements. To our knowledge, it is the most complete validation set-up proposed in the literature. The RFA computational model considered takes into account the main occurring biophysical phenomena (heat transfer, cellular necrosis, hepatic blood flow). The advection effect of the porous circulation in the parenchyma [Audigier 2013a] is taken into account by including a CFD model, robust to image noise and anisotropy to compute the venous and arterial blood flow in the liver parenchyma. By adding such a component, the proposed model is further improved.

The pre-clinical validation approach is difficult to establish. The experimental set-up requires different acquisitions of different modalities (CT, MRI, US), at different times (pre-, intra-, post-operative images). The different images at Day 1 (day of intervention) should be acquired quickly, as the anesthesia of the pig cannot be too long since the pig will be kept alive for 2 additional days. Due to the high complexity of the experimental set-up, pre-processing is necessary. The segmentation was performed semi-automatically, but the vessel meshes had to be smoothed in order to avoid unstable solution of the CFD solver. The probe was segmented from a CT image with a good resolution, the probe geometry considered in the model is convincing. However the artifacts induced by the probe and the bad resolution of the interventional MR images, did not allow to accurately register the probe position. The setting of subject-specific boundary conditions for the blood flow and the pressure was not straightforward, as the acquisition process of Phase-Contrast MR images was complex to handle and required a significant learning curve. Due to the non-reproducible and non-coherent measures on the first four pigs, the measures of blood flows and pressures of Pig 5 only were used for the five pigs in all the computations. Finally, after several trials of registration algorithms, we were able to reliably estimate the deformation between pre- and post-operative imaging for a precise validation of the necrosis extent prediction. The validation of the RFA model on the pre-clinical data presented in this chapter is performed in the following chapter.

Pre-Clinical Study: Results and Discussion

Contents

7.1	Introduction	100
7.2	Evaluation of the Pre-processing	101
7.2.1	Evaluation of the Registration	101
7.2.2	Evaluation of the Smooth Vessels Trees	103
7.3	Verification of the Parameter Estimation Framework	104
7.3.1	Synthetic Data Generation	104
7.3.2	Evaluation	105
7.4	Sensitivity Analysis of the Model Parameters	106
7.5	Evaluation on Synthetic Data	108
7.6	Evaluation on Swine Data	110
7.6.1	Systematic Study	110
7.6.2	Towards Personalization	116
7.7	Sensitivity Analysis on Swine Data	119
7.7.1	Effect of the Advection	119
7.7.2	Effect of the Registration	122
7.7.3	Effect of the Probe Position	123
7.7.4	Cell Death Model Evaluation	125
7.8	Discussion	127
7.8.1	Model Limitations	127
7.8.2	Effect of the Registration	128
7.8.3	Towards Personalization	128
7.9	Conclusion	129

Based On: the conference paper [Audigier 2015b] and on an extended journal version under preparation for the evaluation of the model on pre-clinical data.

7.1 Introduction

In this chapter, the pre-clinical study presented in chapter 6 is used for the evaluation of RFA model. Since our approach relies on pre-clinical data of healthy pigs in a controlled environment, the evaluation leads to an increased confidence in the computed information.

The computational model evaluated here was presented in the last chapter (Section 6.4). An evaluation of each pre-processing steps is performed first (Section 7.2). In Section 7.6, the model is evaluated on twelve ablations from five healthy pigs, to which surrogate tumors have been implanted. We demonstrate accurate predictions on pre-clinical data both for temperature, power evolution and necrotic region extension: the mean error between measured and simulated temperature (respectively power) is 5.1 °C (respectively 25.9 W), which is good compared to the targeted ablation temperature of 105 °C and the maximal power of 150 W. The mean point-to-mesh error between predicted and actual ablation extent is 5.3 ± 3.6 mm, which is smaller than the diameter of the targeted surrogate lesions (around 2 to 3 cm). As we have a LBM implementation on graphics processing units (GPU), a single simulation is faster than real time, which allows us to adjust model parameters for a better evaluation in Section 7.6.2. Key biophysical parameters (heat conductivity and heat capacity) are estimated from this pipeline to show that parameter personalization could be achieved. We use inverse modeling to minimize the error between the computed power and cooling temperature and the observed values (the delivered power directly measured by the RFA device itself and the observed temperature drop during cooling measured at the 5 tips of the probe). Section 7.9 concludes this chapter.

7.2 Evaluation of the Pre-processing

7.2.1 Evaluation of the Registration

First, in each case, the accuracy of the registration was visually checked by an expert as it is illustrated on Figure 7.1.




Figure 7.1: Results of the Post-to-Pre registration on the liver of Fig 1. (*Up*): Superposition of the post-operative liver in red with the pre-operative liver before the registration; (*Down*): Superposition of the post-operative liver in red with the pre-operative liver after the registration.

Table 7.1: Evaluation of the registration: Point-to-mesh errors from the pre-operative meshes to the post-operative meshes in millimeter before and after the registration for Fig 2 are computed. Before the non-rigid registration, all the meshes are centered.

Mesh	Before Registration	After Registration
hepatic veins	4.53 ± 2.13	2.30 ± 0.97
portal vein	6.13 ± 3.81	4.88 ± 4.38
arteries	21.34 ± 23.25	17.60 ± 19.71
tumor 1	2.25 ± 1.38	2.80 ± 1.49
tumor 2	8.80 ± 6.51	5.28 ± 4.50
liver	6.91 ± 3.57	3.53 ± 1.60

Whenever available, the non-rigid transformation between pre- and post-operative images is applied to the post-operative vessel and/or tumor meshes and we can compare them to their pre-operative equivalent meshes. Point-to-mesh errors from pre-operative meshes to registered post-operative meshes are computed and reported in Table 7.1 for Fig 2. Before the non-rigid registration, all the meshes are centered. The error is decreased after the non-rigid registration except for tumor 1 where the two meshes are already in good agreement before the registration. As illustrated in Figure 7.2, the point-to-mesh errors are high for the arteries as the segmented vessels differ in both segmentations of Fig 2.




Figure 7.2: Comparison between the pre-operative arterial mesh in blue and the post-operative meshes (in red: before the registration and in green: after the registration) of Fig 2. It is clear that the arterial vessels segmented on the pre-operative mesh differ from the one segmented on the post-operative mesh. However, the main part of the vessel (vertical part) is correctly registered.

Figure 7.3 shows the results of the registration algorithm on the liver meshes of Fig 2. In this case, the registration improves widely the correspondence between

the post- and the pre-operative mesh as suggested by the point-to-mesh errors (Table 7.1).




Figure 7.3: Comparison between the pre-operative arterial mesh in blue and the post-operative meshes (in red: before the registration and in green: after the registration) of Fig 2. The correspondence between the post- and the pre-operative meshes is improved.

7.2.2 Evaluation of the Smooth Vessels Trees

In order to evaluate the creation of the smooth vessels trees, DICE, Positive Predictive Value (PPV) and sensitivity scores between the original and the smoothed vessel images are computed. The values for Fig 4 are reported in Table 7.2. They show a good correlation between the original segmented vessels and the generated ones.

Table 7.2: Evaluation of the smooth vessel trees: Dice, Positive Predictive Value (PPV) and sensitivity between the original and the smoothed vessel images are computed for Fig 4.

Mesh	DICE (%)	PPV (%)	Sensitivity (%)
arteries	93.1	93.9	92.3
portal vein	91.3	85.3	98.3
hepatic veins	84.9	93.8	77.5

DICE, Positive Predictive Value (PPV) and sensitivity scores have already been defined in Chapter 4 in Section 4.4.1, but we will repeat the definition here for the sake of clarity.

Given V_m and V_s , the volume of measured (respectively simulated) necrotic area, then the DICE score ($DICE$), the sensitivity (S) and the Positive Predictive Value (PPV) are defined as:

$$DICE = \frac{2|V_m \cap V_s|}{|V_m| + |V_s|} \quad S = \frac{|V_m \cap V_s|}{|V_m|} \quad PPV = \frac{|V_m \cap V_s|}{|V_s|}$$

Figure 7.4 illustrates the inconsistent areas, which explained the errors in DICE, PPV and sensitivity for Fig 4 (Table 7.2).




Figure 7.4: (*Left*): Comparison between the arterial meshes before and after the smoothing process of Fig 4. Some vessels, which have a strong stenosis are canceled after the smoothing process. (*Right*): Comparison between the portal vein meshes before and after the smoothing process. Irregularities in the segmentation are smoothed as illustrated by the green circles.

7.3 Verification of the Parameter Estimation Framework

7.3.1 Synthetic Data Generation

In order to verify the parameter estimation framework, we consider a synthetic case on a regular cuboid domain to speed up the process (Figure 7.5). As we try to estimate the heat conductivity and the heat capacity, only the diffusion is considered in this case.

Personalized parameters found in Section 7.8 are used. We apply the typical clinical RFA protocol: during three minutes, we increase linearly the temperature at the tips of the probe, then we heat at 105 °C for six minutes, and finally no temperature is imposed for 3 more minutes as illustrated on Figure 7.6, right panel. With this protocol, a necrotic area, the delivered power and the temperature during the cooling are simulated as shown on Figure 7.6. Those data are then used as ground truth. The main parameters: \bar{d}_t and c_t are estimated by minimizing the cost function defined in Eq. 6.1. The range of parameters values used [Altrogge 2012] are reported in Table 7.3 and the other parameters are fixed to the nominal values, reported in Table 6.1.




Figure 7.5: Set-up of the synthetic case. (*Left*): The temperature map around one tip of the probe. (*Right*): In red, the necrotic area created around the probe in dark.




Figure 7.6: (*Left*): The computed delivered power curve over ablation. (*Right*): The temperature distribution over ablation on the synthetic case. During the heating period, the temperatures are imposed at the nine tips of the probe and the delivered power is simulated. During the cooling period, the temperatures are simulated.

7.3.2 Evaluation

We compare two gradient-free optimization methods available in DAKOTA¹: the Constrained Optimization BY Linear Approximations (COBYLA) and the pattern search method (PS). Those methods are sequential trust-region algorithms. Initially, the total domain of parameters is visited (Table 7.3) and then the region is contracted. Using COBYLA, we manage to obtain the estimated parameters with 0.8% error on \bar{d}_t and 0.6% on c_t in 30 minutes after 22 iterations with an error of $7.8 \cdot 10^{-5}$ of the cost function and a mean of the symmetric point-to-mesh error of 10^{-4} mm. Similarly, using PS, we manage to obtain the estimated parameters with 0.4% of error on \bar{d}_t , 0.3% on c_t in 154 minutes after 117 iterations with an error of 7.2

¹<http://dakota.sandia.gov> - multilevel framework for sensitivity analysis.

Notation	Parameter Name	Min	-	Max
\bar{d}_t	heat conductivity	0.256	0.768	$\text{W}(\text{m K})^{-1}$
c_t^U	heat capacity of undamaged tissue	18	54	$\text{J}(\text{kg K})^{-1}$

Table 7.3: Ranges of parameters values explored in the optimization framework.

10^{-6} of the cost function and a similar mean of the symmetric point-to-mesh error of 10^{-4} mm. The convergence curves of both algorithms are shown on Figure 7.7. We choose the COBYLA method to perform the personalization on the pigs as it requires less iterations of the forward model to estimate the parameters with similar accuracy as PS.




Figure 7.7: Convergence curves of the COBYLA algorithm in blue and of the pattern search algorithm in green. The two optimization algorithms are compared on a synthetic case.

7.4 Sensitivity Analysis of the Model Parameters

In order to reduce the number of parameters to estimate for the personalization of the model, a sensitivity analysis on a synthetic case is performed. We use the same synthetic case as the one used in the previous section 7.3 and described in Figure 7.5. During the heating phase, the temperature at the tips of the probe is imposed as it is shown in Figure 7.6. For this reason, we chose to show only the cooling phase of the temperature. To assess the effect of one given parameter, we manually change its value and run different simulations, all other things remaining equal. Figure 7.8 illustrates the effect of the heat capacity of undamaged cells, c_t^U on the delivered power and on the cooling temperature distribution, especially on

the time of the peak value of the delivered power.




Figure 7.8: Sensitivity analysis on the heat capacity of undamaged cells c_t^U . It affects strongly the time of the peak value of the delivered power. (*Left*): Effect on the delivered power. (*Right*): Effect on the cooling temperature.

Figure 7.9 illustrates the effect of the heat capacity of vulnerable cells c_t^V on the delivered power and on the cooling temperature distribution. This parameter has an effect both on the delivered power and on the cooling temperature distribution.




Figure 7.9: Sensitivity analysis on the heat capacity of vulnerable cells c_t^V . (*Left*): Effect on the delivered power. (*Right*): Effect on the cooling temperature.

Figure 7.10 illustrates the effect of the heat capacity of necrotic cells, c_t^N on the delivered power and on the cooling temperature distribution. This parameter does not largely affect the delivered power neither the cooling temperature.

Finally, Figure 7.11 illustrates the effect of the heat conductivity, \bar{d}_t on the delivered power and on the cooling temperature distribution. The heat conductivity has a clear impact on the cooling temperature distribution.




Figure 7.10: Sensitivity analysis on the heat capacity of necrotic cells c_t^N . (Left): Effect on the delivered power. (Right): Effect on the cooling temperature.




Figure 7.11: Sensitivity analysis on the heat conductivity \bar{d}_t . (Left): Effect on the delivered power. (Right): Effect on the cooling temperature.

We chose to estimate the heat conductivity, \bar{d}_t for its effect on the cooling temperature and the heat capacity of undamaged cells c_t^U for its effect on the time of the peak value of the delivered power.

7.5 Evaluation on Synthetic Data

We want to verify the use of our parameter estimation framework in the case where the considered liver is not healthy. To do so, we use a synthetic case on a regular cuboid domain as the available data come from healthy pigs. We consider the case of a cirrhotic liver, where it is known that there is an “oven effect”. The propagation of the temperature is more difficult in a cirrhotic liver than in a healthy liver, the cirrhotic tissue behave like a thermal insulator, preventing heating outside the

tumor targeted for the ablation [Livraghi 1999]. In order to mimic this effect, the heat conductivity is divided by two. With this value, we apply the typical clinical RFA protocol, and we obtain a “cirrhotic” necrotic area, the delivered power and the temperature during the cooling phase, which differs from those obtained with the nominal value, as illustrated on Figure 7.12, right panel. Those data are then used as ground truth in order to estimate the main parameters: \bar{d}_t and c_t by minimizing the cost function defined in Eq. 6.1.




Figure 7.12: (*Left*): The parameter estimation framework proposed is able to perfectly find the “cirrhotic” heat conductivity. (*Right*): The lesion obtained with the “cirrhotic” heat conductivity in green is enclosed within the lesion obtained with the nominal conductivity coefficient in blue.

The parameter estimation framework is used, and we are able to find exactly the cirrhotic value of the heat conductivity as illustrated on Figure 7.12, left panel. We show on a synthetic case, that our method allows to find “cirrhotic” value of the heat conductivity in the case of a diseased liver. We can now evaluate our method on the available swine data, whose liver tissue are healthy.

7.6 Evaluation on Swine Data

7.6.1 Systematic Study

During our experiments, the measured electrical power appears consistently to reach its maximum during the increase in temperature, before reaching the plateau of 105°C (Figure 7.13, left panel) in all cases.




Figure 7.13: (*Left*): The computed delivered power curves over ablation are compared with the measured one. (*Right*): The temperature distribution after ablation of tumor 4-3. During the heating phase, five temperatures are directly imposed from the five thermistors measurements (superposition of green and black curves) and the four remaining temperatures are imposed from a linear interpolation from those measurements (four green curves). After the use of personalized parameters, the error between the measurement and the computation is reduced from 18.9 to 8.8 W for the power and from 3.5 to 3.0 °C for the temperature. A zoom on the cooling phase is presented in the following Figure 7.14.

This phenomena cannot be explained by a constant heat capacity which would lead to a peak after the plateau is reached. Instead, this observation suggests that the cells reach their vulnerable state faster. Thus we update our necrosis model accordingly: after studying the ODEs of the model and a sensitivity analysis on its parameters non reported here, $\bar{\delta}$ is modified such that tissues reach very fast their vulnerable state which entails a significant change of heat capacity (Table 7.4).

The model is evaluated on twelve ablations performed in five swines (several surrogate tumors ablated for some pigs). To understand the effect of the perfusion on the necrotic lesion, the simulations are performed first without the advection term. We only take into account the cooling effect of the blood vessels, and we will account for the hepatic perfusion later by adding the advection term in a second step (Section 7.7.1).

The applied RFA protocol is not exactly the same for all ablations. Eight ablations are performed through several short cooling and heating periods, whereas the other four ablations included only one long final cooling stage after a continuous




Figure 7.14: Zoom on the cooling phase of the temperature distribution after ablation from tumor 4-3. The personalized parameters do not affect the cooling temperatures.

heating period (it is the case for tumor 4-3 as shown on Figure 7.13). For all pigs, nominal values of parameters (reported on Table 6.1) are employed. In each case, the simulated lesion was compared with the registered ground-truth. Figure 7.15 shows results for tumor 4-2.




Figure 7.15: Comparison between the simulated and the post-operative necrotic areas on tumor 4-2. (*Left*): The simulated lesion is showed around the RFA probe in the subject-specific geometry. (*Right*): Zoom on the ablation area, the simulated lesion is qualitatively close to the registered post-operative lesion. For this tumor, the point-to-mesh error is 4.6 ± 3.9 mm.

Quantitatively, symmetric point-to-mesh errors implemented like in [Zheng 2007] computed between the simulated lesion and the registered post-operative ground-truth showed good prediction of the necrotic extent (average over all ablations of 5.32 ± 3.62 mm of mean point-to-mesh error) as illustrated in red on Figure 7.16. Similarity scores are also computed and shown on Figure 7.17 in red: average over all ablations: DICE of 44.2%, Sensitivity of 47.0%, and PPV of 52.7%. We achieved errors of 25.9 W and 5.1 °C on average between measured and simulated values of cooling temperature and delivered power (Figure 7.18 in red).




Figure 7.16: Point-to-Mesh error in mm computed on the surface of the lesion for three different computations: with nominal parameters, with personalized parameters and with the advection term. For the computation with the advection term, only 6 ablations are considered. (*Left*): Mean of the point to mesh (*Right*): Standard deviation of the point to mesh error.

Values computed with nominal parameters (nominal) and without the advection term are reported in Table 7.10 for Fig 1 and Fig 4, Table 7.11 for Fig 2, Table 7.12 for Fig 3 and Table 7.13 for Fig 5. The prediction of the necrosis extent was valid up to 5mm on average which can be considered as sufficient for clinical applications. Qualitatively, as one can see on Figure 7.19 (red curves), the simulated heat power and temperature were close to the heat power and the temperature given by the RFA probe itself.




Figure 7.17: Similarity scores in % computed on the volume of the lesion for three different computations: with nominal parameters, with personalized parameters and with the advection term. For the computation with the advection term, only 6 ablations are considered. (*Left*): DICE scores (*Middle*): Sensitivity (*Right*): Positive Predictive Value.




Figure 7.18: Error in term of delivered power and temperature for three different computations: with nominal parameters, with personalized parameters and with the advection term. For the computation with the advection term, only 6 ablations are considered. (*Left*): delivered power in W (*Right*): temperature in °C.




Figure 7.19: Comparison between the simulated and the measured temperature on tumor 3-1 for 2 different computations: with nominal parameters and with personalized parameters. During the heating phase, the temperature is imposed, the power is simulated and compared to the measured one. On average, the error is reduced after the use of personalized parameters: from 35.1 W to 29.8 W. It is mainly reduced on the first heating period (from 50.3 W to 26.5 W). During the cooling phase, the simulated (non-imposed) temperature is compared to the measured one, whenever measurement is available (the RFA power was turned off during the cooling phases). On average, the error is also reduced: from 5.9 °C to 1.9 °C.




Figure 7.20: Comparison between the simulated and the measured temperature on tumor 3-1 for 2 different computations: with and without the advection term. During the heating phase, the temperature is imposed, the power is simulated and compared to the measured one. The error in term of delivered power is constant (35.0 W versus 35.1 W). During the cooling phase, the simulated (non-imposed) temperature is compared to the measured one, whenever measurement is available (the RFA power was turned off during the cooling phases). There is no real difference in term of cooling temperature error: 6.3 °C versus 5.9 °C.

7.6.2 Towards Personalization

As illustrated on the pipeline of Figure 6.2, the framework enables parameter estimation too. COBYLA is used to minimize the cost function (Eq. 6.1) as only a few forward simulations (typically 20) are required as described in Section 7.3. The data came from healthy pigs of similar age and weight, we hypothesize that the parameters would be the same for all of them. The vulnerable rate coefficient is also adjusted to match the raise in delivered power as detailed in Section 7.6.1. As the effect of the hepatic perfusion has not been clearly identified so far, we do not consider the advection in the parenchyma in this part of the study. We run the simulation without taking into account the effect of the blood flow in the parenchyma, but only the heat sink effect of the blood flow in the large visible vessels. The results of the personalization are then compared with simulations where the advection term is not accounted as well: (perso) versus (nominal).

This is a proof of concept not only for the ability to personalize the model but also for its predictive power by evaluating the simulation results on ten different tumors. Briefly, the minimization of the error between measured and simulated values of power and temperature was done only on two tumors with a long final cooling stage (2 different tumors: Fig 1 and Tumor 4-3 of Fig 4) as it was long enough to observe reliably the effect of the conductivity \bar{d}_t , yielding two sets of personalized values. In both cases, the values independently found were really close: the same value of heat capacity c_t^U was estimated, and the conductivity values were almost equal to the nominal value, as expected. The values are reported on Table 7.4 and shown on Figure 7.21.

Table 7.4: Comparisons between personalized and nominal values of estimated parameters.

Notation	Nominal	Personalized Values		Automatic optimization
	Value	on tumor 1	on tumor 4-3	
$\bar{\delta}$	3.3×10^{-3}	1×10^{-4}	1×10^{-4}	No
c_t^U	3.6×10^3	3.6×10^1	3.6×10^1	Yes
\bar{d}_t	0.512	0.614	0.512	Yes

The estimated heat capacity and the nominal conductivity were then used to simulate RFA on the ten remaining cases and errors in temperature, heat power and necrosis size were evaluated (Table 7.10 for Fig 1 and Fig 4, Table 7.11 for Fig 2, Table 7.12 for Fig 3, Table 7.13 for Fig 5, (perso)).

Small errors were obtained in those cases too, without previously having fit the parameters for those tumors. With those changes, the simulated electrical power matches much better the measured one in terms of time to peak: average error between measured power and simulated power reduces from 25.9 W to 20.4 W on average (Figure 7.18 in green, left panel). However the mean point-to-mesh error (5.99 mm on average) and the error in temperature (7.8 °C on average) do not change




Figure 7.21: Comparison of the parameters (the vulnerable rate coefficient of the cell death model $\bar{\delta}$, the heat capacity of undamaged cells c_t^U , the heat conductivity \bar{d}_t) personalized on two tumors from two different pigs against the nominal values from the literature.

significantly (in green in Figure 7.16 and Figure 7.18, right panel) as well as the similarity scores: Dice of 42.2%, Sensitivity of 43.0% and PPV of 54.6% on average (Figure 7.17 in green). Figure 7.22 and Figure 7.13 illustrate the improvement in term of delivered power after personalization on Pig 1 and tumor 4-3 (red curves versus green curves).

Although the personalization has only been performed on two pigs, we can see this improvement on tumor 3-1. Figure 7.19 shows how the first cooling phase is closer to the measurement given by the RFA probe itself with personalized parameters (green curve) rather than with the nominal parameters (red curve). The heat capacities of the vulnerable neither the necrotic cells (c_t^V , c_t^N) were modified, thus the simulations with personalized parameters and with nominal parameters of the delivered power for the following heating phases are similar. As the estimated value of \bar{d}_t after personalization is the nominal value, the personalized parameters do not have any impact on the cooling temperature, not on the tumors where the personalization has been performed (Figure 7.23), neither on the others (Figure 7.19,




Figure 7.22: Personalization of the heat capacity and the conductivity using (*Left*) the delivered power curve over ablation (*Right*) the temperature distribution after ablation from tumor 1. The error between the measurement and the computation is reduced from 19.9 to 7.5 W for the power, whereas it is constant for the cooling temperature (11.1 °C versus 10.3 °C). A zoom on the cooling phase is presented on the following Figure 7.23.




Figure 7.23: Zoom on the cooling temperature distribution after ablation from tumor 1. The personalized parameters do not affect the cooling temperatures.

Down). We do not have the measured cooling temperatures for this tumor 3-1, however we can see that this value of \bar{d}_t give accurate simulations, since there is no jump between the computed temperatures at the end of a given cooling period and the measured temperatures at the beginning of the following heating period for the

different heating/cooling stages at times: $t = 300s$, $t = 480s$, $t = 680s$ and $t = 890s$ (Figure 7.19, Down).

7.7 Sensitivity Analysis on Swine Data

7.7.1 Effect of the Advection

In order to quantify the effect of hepatic perfusion, the simulations were performed by adding the advection term for all pigs whenever possible. In some cases (Fig 2 and Fig 4), the segmentation of the inlet of vena cava was not possible, thus the impossibility to get results from the CFD solver. Nominal values of parameters (reported on Table 6.1) were employed as well.




Figure 7.24: Comparison between the simulated and the post-operative necrotic areas on Fig 1. (Left): The post-operative lesion is showed around the RFA probe in the subject-specific geometry. (Right): Zoom on the ablation area, the simulated lesions with and without advection are compared to the registered post-operative lesion. The advection has an effect on the shape of the lesion, however in this case, it does not improve the point-to-mesh error: 6.9 ± 5.2 mm versus 4.8 ± 4.4 mm.

The quantitative results are reported in Table 7.10 for Fig 1, Table 7.12 for Fig 3 and Table 7.13 for Fig 5. (nominal) stands for computation without the advection term and (advection) with the advection term. On average, the effect of the advection cannot be clearly highlighted. The mean of the point-to-mesh error over the 6 ablations computed with advection is 4.96 ± 3.50 mm (Figure 7.16 in blue), the average DICE score is 45.6%, the average sensitivity is 41.8% and the average PPV is 53.4% (Figure 7.17 in blue). On average, the error on the

simulated power is 22.1 W and 4.7°C on the temperature (Figure 7.18 in blue). Those values are comparable to the values obtained without the advection term. However, even if it is not the case globally, the advection has an effect on the computed lesion as well as the computed temperature and power depending on the ablation location. In some cases, the advection reduces the error in term of necrotic extent and temperature distribution (tumor 5-1), however it is not true in every case (Fig 1), but the advection term do change the shape of the necrotic extent (Figure 7.24). Quantitatively, this effect of the advection on the lesion shape can be seen on Table 7.14 for Fig 1 and Fig 3, Table 7.15 for Fig 5. In the cases where the advection has an effect, the sphericity of the lesion decreases and the maximum diameter of the lesion increases. The advection has an effect on the temperature distribution at the tips of the probe as well, depending on the ablation location (Figure 7.25).




Figure 7.25: Effect of the location of the ablation on the temperature dispersion. (*Left*): Fig 1, the probe is close to vessels, the tips do not cool down similarly (Figure 7.27). (*Right*): Tumor 3-1, the probe is far from vessels, the tips cool down similarly (Figure 7.20). In this case, we could not segment any post-op lesion.

If the ablation location is close the vessels (Fig 1, Figure 7.25, left panel), during the cooling phase, the measured and simulated temperature are more spread if the advection is considered (Figure 7.27), while if it is not the case (Tumor 3-1, Figure 7.25, right panel), the measured temperature cool down similarly (Figure 7.20). The perfusion has also an effect on the delivered power needed to maintain the targeted ablation temperature of 105 °C in the case of a long heating period, as illustrated on Figure 7.26, left panel. If the advection is accounted for, more power is necessary to maintain the targeted ablation temperature than if the simulation is




Figure 7.26: Effect of the advection on (*Left*) the delivered power curve over ablation (*Right*) the temperature distribution after ablation from tumor 1. The error between the measurement and the computation is reduced from $10.3\text{ }^{\circ}\text{C}$ to $9.7\text{ }^{\circ}\text{C}$ for the cooling temperature with the advection, whereas it does not really affect the delivered power (19.9 W to 20.8 W). A zoom on the cooling phase is presented on the following Figure 7.27.




Figure 7.27: Zoom on the cooling temperature distribution after ablation for tumor 1. When the advection term is taken into account, the cooling temperatures at the different tips of the probe do not have to same cooling speed (blue curves). Whereas, they cool down similarly without the advection term (red curves).

run without advection. In this case, it could be explained by the fact than the perfu-
sion tends to cool down the parenchyma and thus more power is needed to maintain

the targeted temperature. In the case of short heating and cooling stages, we do not observe this phenomena (Figure 7.20), it might be because the liver temperature drops quickly when the delivered power is decreased.

7.7.2 Effect of the Registration

One part of the evaluation of the computation is done by a comparison between the simulated necrotic extent and a lesion segmented on a post-operative image registered to the pre-operative image. The modality of the post-operative image is not always the same (Table. 7.5).

Table 7.5: Modalities of the pre- and post-operative images used. The segmentations of liver, vessels, tumor on the pre-operative image and of necrosis on post-operative image do not come from the same modality for each pig.

Pig	pre-op image	post-op image
1	CT	CT
2-1	CT	T1 + gad
2-2	CT	T2
3-1	CT-Venous	N.A
3-2	"	T2
3-3	"	"
4-1	CT	CT
4-2	"	"
4-3	"	"
5-1	CT-Port	T2
5-2	"	"
5-3	"	"

In some case, the resolution and the size of the post-operative image do not allow an accurate registration of the lesion. It is the case for tumor 2-2 for example (resolution: 0.78x0.78x4.8 mm, size: 384x384x36 mm). To check the sensitivity of the computation to the registration, a new registration of the post-operative lesion is performed for this tumor. We assume that the barycenter of the post-operative lesion is the same as the barycenter of the simulated necrotic extent, and we rigidly translate the post-operative lesion to align the two barycenters (Figure 7.28, right panel).

With this simple registration method, the results are significantly improved (Table. 7.6). The lesion registration strongly affects the error measured between the computed necrotic area and the ground truth.




Figure 7.28: Effect of the registration on Fig 2 (*Left*): The post-operative lesion after the non-rigid registration is showed in the subject-specific geometry. The point-to-mesh error is 8.4 ± 5.0 mm (*Right*): The post-op lesion after a single translation. The point-to-mesh error reduces to 3.6 ± 2.9 mm.

Table 7.6: Evaluation of the effect of the registration on Fig 2-2: on average, point-to-mesh errors are divided by 2 and DICE scores multiplied by more than 3.

Pig 2-2 registration	point-to-mesh error in mm (mean, sd, min, max, median, P_{95})	Dice (%)	Sensitivity (%)	PPV (%)
non-rigid				
(nominal-0.5)	8.52, 5.03, 0.11, 23.30, 8.37, 16.59	13.8	9.8	23.3
(nominal-0.1)	8.03, 4.85, 0.10, 22.28, 7.83, 15.58	16.8	13.1	23.5
(perso)	8.44, 5.00, 0.10, 22.99, 8.27, 16.35	14.3	10.4	22.6
translation				
(nominal-0.5)	3.87, 2.95, 0.10, 18.00, 2.98, 10.78	57.3	40.7	96.5
(nominal-0.1)	3.12, 2.65, 0.11, 17.08, 2.31, 9.38	66.3	51.7	92.1
(perso)	3.63, 2.87, 0.14, 17.68, 2.70, 10.24	60.1	43.9	94.9

7.7.3 Effect of the Probe Position

The position of the probe is known from an intra-operative anatomical MRI. In some cases, the resolution and the image size do not allow an accurate registration of the probe. It is the case for tumor 3-2 for example (resolution: $1 \times 1 \times 3.5$ mm, size: $320 \times 250 \times 30$ mm).

A new computation is performed for this tumor in order to check the sensitivity of the computations to the probe position. The same configuration was used except that we manually put the probe inside the registered post-operative necrotic area

(position 2 in Figure 7.29). With this probe position, the results are significantly improved (Table. 7.7). The probe position affects also the error measured between the computed necrotic area and the ground truth (Figure 7.29). It is equivalent to a registration error since by moving the probe, we move the computed necrosis.




Figure 7.29: Comparison between the simulated and the post-operative necrotic areas on tumor 3-2. (*Left*): The post-operative lesion is showed around 2 possible positions of the RFA probe. (*Right*): Zoom on the ablation area, the simulated lesion is closer to the registered post-operative lesion when the probe is in position 2. The point-to-mesh error is 12.6 ± 7.7 mm in Position 1 and 6.2 ± 4.0 mm in Position 2.

Table 7.7: Quantitative evaluation of the probe position on Fig 3-2: on average, point-to-mesh errors decrease significantly and DICE scores increase by a factor 2.

Pig 3-2 Probe in	Power error (W)	Temperature error (°C)	point-to-mesh error in mm (mean, sd, min, max, median, P_{95})	Dice (%)	Sensitivity (%)	PPV (%)
Position 1						
(nominal-0.5)	51.9, 43.3, 32.6, 30.7, 36.7	5.4, 1.2, 15.5, 14.4	12.13, 7.53, 0.11, 29.78, 11.51, 24.59	20.0	35.4	13.9
(nominal-0.1)	67.1, 59.9, 47.8, 46.1, 59.1	1.8, 4.8, 10.2, 8.6	12.61, 7.71, 0.06, 31.63, 12.18, 25.77	19.9	40.3	13.2
(perso)	52.0, 59.5, 48.4, 46.3, 59.2	2.7, 5.0, 9.8, 8.4	12.61, 7.70, 0.06, 31.51, 12.14, 25.59	19.9	40.3	13.2
Position 2						
(nominal-0.5)	47.9, 57.3, 45.1, 42.6, 55.6	4.0, 2.5, 12.6, 11.1	5.96, 3.88, 0.02, 17.15, 6.11, 12.42	47.8	86.6	33.0
(nominal-0.1)	65.4, 57.8, 44.3, 42.3, 55.4	3.0, 1.8, 13.1, 11.3	6.19, 4.02, 0.07, 17.42, 6.43, 12.74	46.8	89.0	31.8
(perso)	49.6, 40.2, 29.9, 28.9, 32.0	6.8, 2.4, 18.5, 17.2	5.59, 3.57, 0.08, 16.13, 5.57, 11.80	49.8	89.0	31.8

7.7.4 Cell Death Model Evaluation

As the parameters of the cell death model have been modified after parameter estimation, the accuracy of the model has to be verified. We compare the simulated necrotic extent for tumor 2-2 with the results of a lethal thermal dose model, often used in clinical settings [Sapareto 1984] (Figure 7.30). The thermal dose at each voxel is calculated according to this empirical model:

$$t_{43} = \sum_{t=0}^{t_c} C^{(43-\bar{T})} \Delta t \quad (7.1)$$

where t_{43} is the thermal dose in equivalent minutes at 43°C, \bar{T} is the average temperature during the time interval Δt , and C is a constant resulting from the Arrhenius model.

$$\begin{aligned} C &= 0.5 && \text{when } \bar{T} \geq 43^\circ\text{C} \\ C &= 0.25 && \text{when } \bar{T} < 43^\circ\text{C} \end{aligned}$$

A thermal dose of 240 (in equivalent minutes at 43°C) is considered lethal.

Quantitatively, both models give similar point-to-mesh errors and DICE scores (Table. 7.8), however the volume and maximum diameter of the necrotic extent are larger with the thermal dose model (Table. 7.9). The thermal dose model gives really spherical necrotic extents (Table. 7.9, sphericity values closer to 1 for thermal dose model). A more throughout study is necessary to understand the differences




Figure 7.30: Effect of the cell death model (*Left*): The simulated lesion computed using the cell death model is showed in the subject-specific geometry of Fig 2. (*Right*): The area where at least a thermal dose of 240 in equivalent minutes at 43 °C is shown. The post-op lesion appears in dark in both cases.

between those two models of cellular necrosis. However, the proposed three-state model enables a spatial and time-varying heat capacity (c_t^U, c_t^V, c_t^N , for undamaged, vulnerable and necrotic cells). We saw in Section 7.4 that those parameters affect the delivered power and the cooling temperature. With the thermal dose model, only two states of the cells could be considered: they are either undamaged (if $TD < 240$) or necrotic (if $TD \geq 240$). Thus only two values of the heat capacity could be used.

Table 7.8: Evaluation of the effect of the cell death model - thermal Dose at 240 on the point-to-mesh errors and DICE scores between the simulated lesion and ground truth on Fig 2-2.

Pig 2-2	point-to-mesh error in mm (mean, sd, min, max, median, P_{95})	Dice (%)	Sensitivity (%)	PPV (%)
Thermal Dose				
(nominal-0.5)	8.25, 5.71, 0.03, 21.1, 7.36, 18.87	13.9	23.2	9.9
(nominal-0.1)	9.05, 6.50, 0.06, 24.99, 7.76, 21.10	14.3	29.9	9.4
(perso)	8.68, 6.13, 0.15, 22.77, 7.61, 20.09	13.9	25.9	9.5
Cell Death				
(nominal-0.5)	8.52, 5.03, 0.11, 23.30, 8.37, 16.59	13.8	9.8	23.3
(nominal-0.1)	8.03, 4.85, 0.10, 22.28, 7.83, 15.58	16.8	13.1	23.5
(perso)	8.44, 5.00, 0.10, 22.99, 8.27, 16.35	14.3	10.4	22.6

Table 7.9: Quantitative evaluation of model prediction: Effect of the cell death model on the shape of the simulated lesion.

Pig 2-2	sphericity	volume (mm^3)	max diameter (mm)
(GT)	0.45	12853	50.09
(GT-smooth)	0.71	5356	30.43
Thermal Dose			
(nominal-0.5)	0.96	12519	34.19
(nominal-0.1)	0.93	17019	37.39
(perso)	0.94	14641	35.37
Cell Death			
(nominal-0.5)	0.90	5406	23.96
(nominal-0.1)	0.90	7193	26.32
(perso)	0.90	5933	24.78

7.8 Discussion

We have presented a multi-physics model based on multi-modal medical images, evaluated on pre-clinical data. We rely on LBM to solve the bioheat equations and to compute the blood flow in the liver. Despite possible biases in the probe location and in establishing correspondences between the post to the pre-operative images due to registration errors, our model provided new insights, and we show that parameter estimation is possible to reduce the bias introduced by the use of nominal parameters. This opens new opportunities for RFA planning and guidance.

7.8.1 Model Limitations

The impact of the segmentation process has to be taken into account as well as the model limitations (assumptions: we assume that the measured power is actually the heat power through radio-frequency, with no loss, etc).

We are not able to accurately locate the position where we measure the pressure, but we assume that there is no variation in pressure in a small neighborhood of the two outlets.

The Neumann Boundary condition applied on the border of the liver can be discussed. If the ablation location is subcapsular (tumor 5-1, tumor 4-1, tumor 3-3), the volume of the necrotic area is larger than in the other cases (Table 7.14, Table 7.15), as the heat does not dissipate through the liver capsule but reflects on the border and increases artificially the heat diffusion inside the liver as it is illustrated in Figure 7.31. However, the liver has a high regulation capacity thanks to its rich vascular system. If the heating is applied close to the border of the liver, it is not clear how the heat will dissipate, whether or not it will affect the neighboring organs.

Finally, during the RFA computation, the temperature at the tips of the probe




Figure 7.31: (*Left*): The simulated necrotic extent around the RFA probe, close to the border of the liver from tumor 3-3. The post-operative lesion is shown in brown. (*Right*): The liver extraction confirms that the necrotic lesion is subcapsular.

is imposed. It will be more accurate to directly impose the delivered power and thus simulate the cooling as well as the heating temperature distribution. It should be done in the future, but going from a Dirichlet boundary condition on the temperature at the tips of the probe to a Dirichlet boundary condition on the delivered power is not straightforward.

7.8.2 Effect of the Registration

It should be noted that the registration method used in this paper differs from the registration used in [Audigier 2015b]. In [Audigier 2015b], due to the uncertainty in the registration of the post-operative image to the pre-operative image, the necrotic lesion segmented on the post-operative image was registered rigidly to the pre-operative image by aligning its barycenter with the barycenter of the simulated necrosis which could introduce bias in the analysis. The impact of those registration processes has been evaluated on one tumor in this chapter.

7.8.3 Towards Personalization

In clinical RFA of liver tumors, we cannot assume that the biophysical parameters are the same for all patients, as assumed here for the five pigs, and parameter personalization is therefore required. Here, we propose to adjust model parameter for a better evaluation of RFA, based on the delivered electrical power during ablation and the temperature drop during cooling. Since our approach does not rely on ablated regions, the parameter estimation leads to an increased confidence in the computed temperature, in the case where no temperature map is available. By fitting the temperature and the delivered heat power, the novel approach can estimate the temperature around each tumor at any time during the ablation. This additional information could be used as surrogate to assess the amount and location of damaged tissue during the intervention (cells receiving excessive heat but without being necrosed) surrounding the ablated region.

Furthermore, the probe temperature and delivered power are information that are readily available, in real-time, from the RF system and therefore could be used for therapy guidance. Even if the computation time does not allow an interventional set-up yet, it is a first step towards the personalization of RFA computation.

The extent of the necrotic area is mainly controlled by the conductivity \bar{d}_t and the heat capacity of vulnerable cells c_t^V (not of undamaged ones c_t^U). Since \bar{d}_t was not changed after optimization, there was no significant difference in terms of necrotic area after personalization (mean of the point-to-mesh error of 5.99 mm versus 5.32 mm on average) or cooling temperature (error of 7.8 °C versus 5.1 °C), despite a better match for the measured power (error of 20.4 W versus 25.9 W) as illustrated on Figure 7.19 and Figure 7.22, and reported in Table 7.10, Table 7.11, Table 7.12, Table 7.13. Moreover the point-to-mesh errors were of the order of 5mm; it suggests that simulations with optimized \bar{d}_t and c_t^U are realistic in terms of necrotic area, power and temperature predictions which was the objective. These results confirmed the stability of the personalization framework.

7.9 Conclusion

In this chapter, we evaluated our model on a pre-clinical data. The approach is evaluated on five swines and twelve ablations. It gives promising results, but more importantly provides several lessons learned.

Despite a very comprehensive model and a complete pre-clinical study for its validation, many errors are involved, at the modeling level, but also especially at each step of the pre-processing. We manage to identify most of those errors thanks to the pre-clinical study. It would not have been possible on clinical data. The use of invasive catheter to measure the pressure could not be done on a clinical study, neither the acquisition of Phase-Contrast MRI, intra-operative MRI with the probe, injected CT scans with the three different phases, which are not part of the standard clinical workflow for RFA.

Even if model personalization is premature since the computation time has to be reduced first, we show that it can be done based on intra-operatively simple quantities (the probe temperature and delivered power) since we rely on a LBM implementation on graphics processing units (GPU): a forward simulation is faster than real time. As a first step towards personalization, we evaluate the discrepancy in terms of temperature and delivered power, and we show that key biophysical parameters can be estimated leading to promising predictions. In clinical settings, due to the large variety of diseases treated by RFA (cirrhosis, fibrosis, etc), the proposed method should be suitable to get patient-specific parameters from easily accessible data.

A study of the errors is a necessary next step before using model-based therapy in the clinical routine. A study of how the pre-processing errors (due to segmentation, post-to-pre registration, probe registration, vessels smoothing) propagate to the final computational outcome will be of great interest. It will help to understand the

modeling errors as well and be more confident on the resulting necrotic lesion for example. The evaluation could be improved by the acquisition of images of better resolution for the intra-operative MRI with the probe for example, but taking into account that with the current framework, the acquisition time is limited by the well-being of the pig. The acquisition of the Phase-Contrast MRI could be improved as well. Moreover, even if today MR thermometry does not provide such a good resolution (Figure 7.32 shows MR thermometric images acquired during the ablation of tumor 4-1), in the future, it could add more information on the spatial extent of the temperature to further improve the validation.




Figure 7.32: MR thermometric images acquired during the ablation of tumor 4-1. (*Left*): image acquired before the ablation, showing an isotropic and constant temperature around the RFA probe (liver temperature). (*Right*): image acquired during the ablation, when the RF power is decreased. It shows the temperature distribution around the RFA probe.

Table 7.10: Quantitative evaluation of model prediction. The mean errors between simulated and measured values are computed for each ablation. For the eight ablations which are performed through several cooling and heating periods, the different error values for each period are reported. For each tumor, several computations are performed, with nominal parameters and two different time-step (nominal- dt), with personalized parameters, and with advection. The parameter estimation has been performed on Pig 1 and tumor 4-3 and gave a set of personalized parameters, which are used in every case (perso). When it is not specify, the time-step is $dt = 0.1$ s. For computation with nominal parameters (respectively with personalized parameters), on average, mean errors of 25.9 W (resp. 20.4 W), 5.1 °C (resp. 7.8 °C) and 5.32 mm (resp. 5.99 mm) of mean of the point-to-mesh error are achieved. This table presents results for Pig 1 and Pig 4.

Pig	Power error (W)	Temperature error (°C)	point-to-mesh error in mm (mean, sd, min, max, median, P_{95})	Dice (%)	Sensitivity (%)	PPV (%)
1						
(nom-0.5)	21.9	8.3	5.14, 4.77, 0.14, 21.38, 3.51, 16.13	36.8	25.9	63.9
(nom-0.1)	19.9	10.3	4.79, 4.36, 0.09, 20.55, 3.58, 15.08	41.1	31.5	59.1
(perso)	7.5	11.1	5.06, 4.68, 0.19, 21.34, 3.52, 16.00	37.4	26.8	62.0
(adv)	20.8	9.7	6.90, 5.20, 0.13, 22.91, 5.53, 16.78	24.9	22.9	27.3
4-1						
(nom-0.5)	12.2, 13.3, 20.9, 31.9	11.2, 10.7, 10.6	6.29, 3.71, 0.17, 13.3, 6.24, 11.95	23.7	81.9	13.8
(nom-0.1)	16.8, 10.3, 5.1, 6.2	9.0, 7.3, 6.0	7.30, 3.91, 0.13, 14.73, 7.37, 13.19	20.2	89.0	11.4
(perso)	13.4, 8.6, 5.8, 5.7	7.2, 5.3, 4.3	6.78, 3.79, 0.16, 13.97, 6.83, 12.61	22.0	86.0	12.6
4-2						
(nom-0.5)	16.7, 19.2, 20.1	3.0, 7.8	4.85, 4.41, 0.12, 18.00, 3.31, 15.51	25.6	16.7	54.7
(nom-0.1)	24.9, 16.9, 9.4	0.5, 3.7	4.62, 3.88, 0.15, 17.23, 3.32, 14.21	30.3	21.7	50.4
(perso)	4.2, 10.0, 9.3	18.9, 4.7	4.73, 4.14, 0.12, 17.80, 3.30, 14.93	27.5	18.8	51.5
4-3						
(nom-0.5)	12.9	1.7	2.39, 1.64, 0.07, 8.82, 2.05, 5.96	70.0	61.1	81.8
(nom-0.1)	18.9	3.5	2.07, 1.32, 0.06, 7.16, 1.77, 4.78	75.6	73.8	77.5
(perso)	8.8	3.0	2.20, 1.45, 0.05, 8.05, 1.93, 5.30	73.2	67.1	80.5

Table 7.11: Quantitative evaluation of model prediction. This table presents results for Fig 2.

Pig	Power error (W)	Temperature error (°C)	point-to-mesh error in mm (mean, sd, min, max, median, P_{95})	Dice (%)	Sensitivity (%)	PPV (%)
2-1						
(nom-0.5)	7.4	2.6	2.93, 2.66, 0.18, 15.83, 2.05, 8.88	59.8	44.1	92.8
(nom-0.1)	16.0	1.5	2.46, 2.28, 0.09, 14.65, 1.64, 7.06	66.9	54.3	87.0
(perso)	10.6	1.3	2.81, 2.55, 0.12, 15.25, 1.97, 8.42	62.0	46.8	92.0
2-2						
(nom-0.5)	47.5	3.1	8.52, 5.03, 0.11, 23.30, 8.37, 16.59	13.8	9.8	23.3
(nom-0.1)	62.3	5.0	8.03, 4.85, 0.10, 22.28, 7.83, 15.58	16.8	13.1	23.5
(perso)	50.0	4.5	8.44, 5.00, 0.10, 22.99, 8.27, 16.35	14.3	10.4	22.6

Table 7.12: Quantitative evaluation of model prediction. The tumor 3-1 was ablated twice with two consecutive heating periods, thus the simulated necrosis cannot be compared with any ground truth. This table presents results for Fig 3.

Pig	Power error (W)	Temperature error (°C)	point-to-mesh error in mm (mean, sd, min, max, median, P_{95})	Dice (%)	Sensitivity (%)	PPV (%)
3-1						
(nom-0.5)	37.4, 26.0, 16.4, 19.4, 23.5	2.1				
(nom-0.1)	50.3, 38.5, 35.6, 30.4, 20.5	5.9	N.A	N.A	N.A	N.A
(perso)	26.5, 36.0, 33.9, 31.5, 20.9	1.9				
(adv)	50.4, 38.8, 35.6, 30.2, 20.1	6.3				
3-2						
(nom-0.5)	51.9, 43.3, 32.6, 30.7, 36.7	5.4, 1.2, 15.5, 14.4	12.13, 7.53, 0.11, 29.78, 11.51, 24.59	20.0	35.4	13.9
(nom-0.1)	67.1, 59.9, 47.8, 46.1, 59.1	1.8, 4.8, 10.2, 8.6	12.61, 7.71, 0.06, 31.63, 12.18, 25.77	19.9	40.3	13.2
(perso)	52.0, 59.5, 48.4, 46.3, 59.2	2.7, 5.0, 9.8, 8.4	12.61, 7.70, 0.06, 31.51, 12.14, 25.59	19.9	40.3	13.2
3-3						
(nom-0.5)	17.7, 13.7, 39.2	2.9, 5.4	2.62, 1.75, 0.11, 10.04, 2.29, 6.14	72.4	57.9	96.4
(nom-0.1)	29.6, 20.8, 19.1	1.7, 9.4	2.03, 1.36, 0.06, 8.84, 1.71, 4.66	80.0	72.0	90.1
(perso)	15.4, 19.7, 18.9	1.0, 9.8	2.20, 1.50, 0.06, 9.68, 1.86, 5.12	77.8	67.1	92.4
(adv)	29.6, 20.7, 20.2	0.8, 5.9	3.26, 2.18, 0.06, 10.95, 2.80, 7.65	68.0	65.2	71.0

Table 7.13: Quantitative evaluation of model prediction. This table presents results for Fig 5.

Pig	Power error (W)	Temperature error (°C)	point-to-mesh error in mm (mean, sd, min, max, median, P_{95})	Dice (%)	Sensitivity (%)	PPV (%)
5-1						
(nom-0.5)	9.6, 23.0, 25.2, 28.8	3.1, 10.6, 6.3, 4.8	5.06, 3.10, 0.13, 13.35, 4.68, 10.96	45.6	41.0	51.3
(nom-0.1)	16.2, 40.6, 17.7, 15.9	1.4, 6.7, 6.9, 3.9	4.89, 2.72, 0.11, 11.75, 4.62, 9.57	48.7	49.4	48.0
(perso)	3.9, 10.9, 15.8, 15.8	16.4, 21.9, 6.8, 3.7	4.91, 2.85, 0.14, 12.45, 4.53, 9.96	47.9	45.8	50.2
(adv)	16.2, 40.6, 17.7, 15.9	1.4, 6.7, 6.9, 3.9	4.89, 2.73, 0.11, 11.75, 4.59, 9.58	48.6	48.9	48.3
5-2						
(nom-0.5)	13.5, 21.2, 20.0, 25.2	6.8, 9.3, 5.3	5.34, 4.19, 0.09, 17.40, 4.28, 13.50	44.4	29.6	89.3
(nom-0.1)	19.6, 14.0, 6.8, 10.3	4.9, 6.4, 1.8	4.58, 3.80, 0.09, 16.64, 3.41, 12.45	51.0	36.9	82.6
(perso)	8.5, 13.2, 7.1, 9.9	19.1, 7.0, 1.8	4.89, 4.02, 0.08, 17.13, 3.68, 12.99	48.4	33.8	85.6
(adv)	19.6, 14.0, 6.8, 10.3	4.9, 6.4, 1.8	4.61, 3.82, 0.09, 16.64, 3.44, 12.51	50.8	36.5	83.2
5-3						
(nom-0.5)	5.7, 19.5, 14.4	6.6, 3.1	5.39, 3.77, 0.12, 18.25, 4.74, 12.95	32.4	27.7	39.2
(nom-0.1)	12.1, 14.1, 9.2	4.9, 1.7	5.15, 3.60, 0.10, 17.36, 4.46, 11.83	35.9	34.8	37.2
(perso)	2.8, 13.8, 9.0	19.4, 1.1	5.26, 3.68, 0.16, 18.08, 4.60, 12.30	34.3	31.0	38.4
(adv)	12.1, 14.1, 9.2	4.9, 1.7	5.15, 3.59, 0.10, 17.36, 4.50, 11.83	35.8	34.2	37.4

Table 7.14: Quantitative evaluation of model prediction on the shape of the computed lesion. Results for Fig 1, Fig 2 and Fig 3 are reported here.

Pig	sphericity	volume (mm^3)	max diameter (mm)
1 (GT)	0.71	10228	43.46
1 (nominal -0.5)	0.83	4137	27.48
1 (nominal -0.1)	0.84	5450	29.27
1 (perso)	0.83	4423	28.05
1 (advection)	0.82	8568	35.86
2-1 (GT)	0.77	13337	42.67
2-1 (nominal-0.5)	0.90	6344	24.41
2-1 (nominal-0.1)	0.90	8316	26.65
2-1 (perso)	0.90	6781	24.88
2-2 (GT)	0.45	12853	50.09
2-2 (nominal-0.5)	0.90	5406	23.96
2-2 (nominal-0.1)	0.90	7193	26.32
2-2 (perso)	0.90	5933	24.78
3-1 (GT)	N.A	N.A	N.A
3-1 (nominal-0.5)	0.84	13662	33.76
3-1 (nominal-0.1)	0.83	17616	37.22
3-1 (perso)	0.84	16004	36.18
3-1 (advection)	0.83	17593	38.69
3-2 (GT)	0.61	18105	57.80
3-2 (nominal-0.5)	0.87	46048	48.96
3-2 (nominal-0.1)	0.86	55386	52.44
3-2 (perso)	0.86	53008	51.79
3-3 (GT)	0.69	21722	45.18
3-3 (nominal-0.5)	0.89	13059	32.89
3-3 (nominal-0.1)	0.89	17370	35.64
3-3 (perso)	0.89	15781	34.66
3-3 (advection)	0.87	19934	39.53

Table 7.15: Quantitative evaluation of model prediction on the shape of the computed lesion. Results for Fig 4 and Fig 5 are reported here.

Pig	sphericity	volume (mm^3)	max diameter (mm)
4-1 (GT)	0.66	1849	23.94
4-1 (nominal-0.5)	0.89	10956	31.70
4-1 (nominal-0.1)	0.89	14436	34.73
4-1 (perso)	0.89	12574	33.03
4-2 (GT)	0.84	13066	35.86
4-2 (nominal - 0.5)	0.88	4001	22.98
4-2 (nominal - 0.1)	0.89	5621	24.94
4-2 (perso)	0.88	4755	23.77
4-3 (GT)	0.86	11322	35.19
4-3 (nominal-0.5)	0.86	8466	30.33
4-3 (nominal-0.1)	0.85	10778	32.65
4-3 (perso)	0.85	9441	31.50
5-1 (GT)	0.80	15849	42.59
5-1 (nominal-0.5)	0.88	12667	32.45
5-1 (nominal-0.1)	0.88	16040	35.19
5-1 (perso)	0.88	14486	33.97
5-1 (advection)	0.88	16040	35.19
5-2 (GT)	0.71	19129	51.72
5-2 (nominal-0.5)	0.80	6339	28.79
5-2 (nominal-0.1)	0.79	8562	31.62
5-2 (perso)	0.79	7552	30.43
5-2 (advection)	0.79	8398	31.21
5-3 (GT)	0.69	7432	35.89
5-3 (nominal-0.5)	0.85	5237	24.70
5-3 (nominal-0.1)	0.85	6939	26.76
5-3 (perso)	0.85	6001	25.98
5-3 (advection)	0.85	6799	26.76

Summary and Perspectives

Essentially, all models are wrong,
but some are useful.

George E. P. Box

Contents

8.1	Main Contributions	137
8.1.1	Modeling of Radiofrequency Ablation with LBM	137
8.1.2	Model Evaluation on Clinical Data	138
8.1.3	Towards Model Personalization	138
8.1.4	Pre-clinical Model Evaluation	139
8.2	Perspectives	140
8.2.1	Model Improvements	140
8.2.2	Model Validation	140
8.2.3	Model Personalization	141
8.3	Conclusion	141

Each chapter of this manuscript contains conclusions and perspectives on the performed work. We now conclude on the main achievements of this PhD work. Then, we suggest possible improvements related to this work.

8.1 Main Contributions

8.1.1 Modeling of Radiofrequency Ablation with LBM

We have presented an integrated framework for fast modeling of Radiofrequency Ablation (RFA), which computes and predicts RFA interventional outcome in terms of necrotic zone, delivered power and temperature distribution throughout the liver. Our detailed computational model of the biophysical mechanisms (heat transfer, cellular necrosis, hepatic blood flow) involved in RFA of abdominal tumors is based on patient images. We focused on modeling heat propagation and cellular necrosis based on a patient image while considering the heat sink effect of blood vessels and porous circulation in the liver. It is implemented on Graphics Processing Units (GPU) using the Lattice Boltzmann Method to take advantage of the computation capacity of the Graphics Card and thus to reach near real-time computation. Fast

computation is necessary to enable parameter estimation, to test the generalization of a biophysical model in terms of tissue properties, but also for the expected clinical applications. For interventional training, interactive therapy planning or procedure guidance, faster than real-time computations would be ideal. We have not reached this point yet, but with our LBM implementation, near real-time computation can be achieved on a standard machine.

To resume, this thesis presents novel contributions in RFA computational simulation, which can be summarized as follows:

1. A complete patient-specific geometry including hepatic venous and arterial circulation system;
2. The patient-specific modeling including simultaneously the cooling effect of large vessels and of the perfusion within the parenchyma;
3. The discretization method (LBM), which is fast and verified against an analytical solution.

8.1.2 Model Evaluation on Clinical Data

A necessary step before deploying model-based therapy planning or guidance in clinical settings is a clinical validation with extensive data on larger populations to evaluate the computational model of RFA. We showed that it will be possible due to the level of integration of the proposed framework. As a first step towards clinical application of a RFA computational model, this thesis presents a clinical evaluation on a database of ten patients of the developed RFA computational simulation. The results we obtained by comparing the real necrotic lesions from post-operative images with the simulated lesions suggest a promising correlation. The effect of the blood perfusion inside the liver parenchyma as well as the blood flow through the large vessels has also been highlighted. Finally, a first experiment of tissue parameter adjustment on one patient indicates the need of model personalization. Our system constitutes a first step towards the clinical validation of RFA modeling, which can be summarized as follows:

1. The evaluation of the proposed RFA model on a dataset of ten patients;
2. Experiments testing several hypotheses in the discussion: effect of the probe position, effect of the segmentation;
3. The effect of the parenchyma and venous blood flow on the resulting necrotic lesion (ablated tissue);
4. The importance of the personalization of the model biophysical parameters.

8.1.3 Towards Model Personalization

In clinical RFA of liver tumors, due to the large variety of diseases treated (cirrhosis, fibrosis, etc) we cannot assume that the biophysical parameters are the same for all

patients, and parameter personalization is therefore required. We showed that the proposed method could be suitable to get patient-specific parameters from easily accessible data. As we rely on the Fully Coupled LBM CFD solver, no advanced meshing techniques are required. All the computations are directly done from patient images using LBM: heat propagation and cell death modeling as well as the heat sink effect of blood vessels and porous circulation in the liver. Since we rely on an implementation on graphics processing units (GPU): a forward simulation is faster than real time. Two personalization strategies were tested and evaluated on clinical and pre-clinical data.

First, an automatic estimation of the main parameters of the model is proposed and evaluated on clinical data. In this case, the optimization algorithm minimizes the error between the ground truth and the computed necrotic lesion. Therefore, the method needs several tumors for validation. Promising results are achieved with the use of patient-specific parameters on three patients and seven tumors ablated.

Then, a second personalization strategy was presented. In this case, the optimization algorithm minimizes the error between the computed values of the probe temperature and delivered power and the actual values measured by the RFA device. It is evaluated on a pre-clinical dataset of five swines. The personalization of the sensitive tissue parameters has been performed on two surrogate tumors and its predictive power evaluated on the remaining ten surrogate tumors. We showed that key biophysical parameters can be estimated leading to promising predictions. This method is based on intra-operative simple quantities (the probe temperature and delivered power) that are readily available, in real-time, from the RF system and therefore could be used for therapy guidance.

Even if the computation time does not allow an interventional set-up yet, we showed that model personalization can be done based on intra-operative quantities. It is a first step towards the personalization of RFA computation.

8.1.4 Pre-clinical Model Evaluation

The proposed model has been evaluated on pre-clinical data from a comprehensive experimental set-up specially designed for radiofrequency ablation (RFA) model validation. To the best of our knowledge, it constitutes the largest and most comprehensible *in vivo* study reported so far. The model considered is complete and takes into account the main biophysical mechanisms (heat transfer, cellular necrosis, hepatic blood flow) involved in RFA. The pre-clinical dataset includes five swines and twelve ablations, leading to promising evaluation results, but more importantly to several lessons learned. Our pre-clinical study was challenging to establish due to the complexity of the experimental protocol. Many images from different modalities and at different times of acquisition were involved. Before accounting for all the information from the images into the RFA model, an impressive amount of pre-processing was required: segmentation, vessel smoothing, probe modeling, probe registration, image registration. Moreover, the setting of subject-specific boundary conditions for the blood flows and blood pressure was not straightforward. Many

errors are then involved, coming from the model as well as the pre-processing. However, thanks to the pre-clinical dataset used, we manage to identify most of them, that will not be possible with a clinical study. The effects of the segmentation, the probe registration, the post-operative lesion registration and the vessel smoothing have been evaluated. At the modeling level, the effects of the advection, the probe position, the cell death model as well as the sensitivity of the model biophysical parameters were highlighted.

8.2 Perspectives

8.2.1 Model Improvements

Although the implemented model includes the main important biophysical mechanisms involved in radiofrequency ablation of hepatic tumors (heat transfer, cellular necrosis, hepatic blood flow), several improvements can be performed. For instance, a weak coupling between the blood flow and the heat transfer model is currently considered. The blood flow has an influence on the temperature distribution through the WK model and through the reaction term in the Pennes model but the temperature does not affect the blood flow. This weakly coupling model allows us to speed up the calculations since the blood flow distribution is computed only once, at the beginning of the simulation. But a full coupling of blood flow and heat transfer models should give more accurate predictions, and enable us to consider coagulation and water evaporation, thus improving on the presented model.

Furthermore, we realized through the pre-clinical study that the delivered power is information that can be available easily in real-time from the RFA probe itself. In the current implementation of the model, the probe temperature is imposed through Dirichlet boundary conditions during the heating phase. We are then able to simulate completely the delivered power but to simulate the probe temperature during the cooling phases only. By imposing the delivered power instead of the probe temperature, the model will be improved and we will be able to simulate the probe temperature during the total duration of the procedure. Imposing Dirichlet boundary condition on the power in the current LBM implementation is not straightforward, but it is worth investigating this aspect. It would be a more coherent way of simulating RFA.

Finally, some interesting aspects to improve the ongoing RFA intervention have not been considered in this work, but the current RFA model opens possibilities to an automatic estimation of the optimal locations of the probe and the electrical power to deliver in order to achieve a proper tumor ablation for example.

8.2.2 Model Validation

A study of the modeling errors is a necessary next step before using model-based therapy in the clinical routine. Understanding how the pre-processing errors (due to segmentation, post-to-pre registration, probe registration, vessels smoothing) prop-

agate to the final computational outcome is necessary. It will help to make the model more realistic and to have a confidence measure on the resulting necrotic lesion for example. Especially, we saw that the segmentation has a high impact on the computed lesion as the blood flow computation through the CFD solver is strongly correlated with the vessel segmentation. The post-to-pre and the probe registrations influence the validation of the model output. Since modeling of RFA is still an open challenge, the present RFA model is a valuable asset, and future directions include a more throughout validation on a larger cohort of patients.

Moreover, the validation could be improved by the acquisition of good resolution and reliable MR thermometric images. As data acquisition techniques improve quickly, we believe that it would be possible in the near future. It could add more information on the spatial extent of the temperature even far from the RFA probe and thus further improve the validation.

8.2.3 Model Personalization

Model personalization is premature since the computation time has to be reduced first. But we showed on clinical data that it can be done in the case of multiple tumors inside the liver. Adaptation for RFA under image-guidance using intra-operative image and no post-operative image has been considered. However, a necessary step before deploying this method in clinical settings is a pre-clinical validation to evaluate the computational model of RFA and also to consider potential safety issues of the proposed application.

By evaluating the discrepancy in terms of temperature and delivered power on pre-clinical data, we showed that intra-operative simple quantities (the probe temperature and delivered power) can be used for personalization of model-based therapy. Moreover, since this approach does not rely on ablated regions, the parameter estimation leads to an increased confidence in the computed temperature, in the case where no temperature map is available. It can estimate the temperatures around each tumor at any time during the ablation. This additional information could be used as surrogate to assess the amount and location of damaged tissue during the intervention (cells receiving excessive heat but without being necrosed) surrounding the ablated region, helping the clinical monitoring during the RFA intervention.

8.3 Conclusion

Several research groups have already proposed different frameworks for the modeling of radiofrequency ablation (RFA). They have showed the potential benefits of such models but they have also pointed out the challenges involved. Through this thesis study, we showed that important advances have been made in the patient-specific modeling of RFA of hepatic tumor. The results we obtained are promising and after further future improvements, we believe that such personalized model could be used in clinical environment. Although our research was focused mainly on the RFA of hepatic tumor, the same pipeline of methods can be used for tumors located in

other organs, and could be adapted easily to other percutaneous ablative techniques like High Intensity Focus Ultrasound, Microwave Ablation or Laser Ablation for example, according to the properties of the current ablation procedure considered. However, more research is required to tackle the issues we faced at each step of the work flow (segmentation, registration, model validation and so on...). The study and validation of model simplification will be necessary before its integration in the medical practice. As data acquisition techniques and our understanding of the biophysical mechanisms involved in RFA improve quickly, such models will become more accurate and more realistic in the near future. We believe that patient-specific simulations will become a common tool in the therapy planning and guidance for a specific patient.

List of Publications

The presented work led to several published and submitted publications as first author.

Journal Papers

- [Audigier 2015a] *This paper details the presented RFA model and its evaluation on clinical data.*
Audigier, Chloé and Mansi, Tommaso and Delingette, Hervé and Rapaka, Saikiran and Mihalef, Viorel and Boctor, Emad and Choti, Michael and Kamen, Ali and Ayache, Nicholas and Comaniciu, Dorin. Efficient Lattice Boltzmann Solver for Patient-Specific Radiofrequency Ablation of Hepatic Tumors. *IEEE Transaction in Medical Imaging*, 2015, pp. 14.
- Submitted *This paper will present the pre-clinical study and the model evaluation on the pre-clinical data.*
Audigier, Chloé and Mansi, Tommaso and Delingette, Hervé and Rapaka, Saikiran and Passerini, Tiziano and Mihalef, Viorel and Jolly, Marie-Pierre, and Pop, Raoul and Diana, Michele and Soler, Luc and Kamen, Ali and Comaniciu, Dorin and Ayache, Nicholas. Comprehensive Pre-Clinical Evaluation of a Multi-physics Model of Liver TumorRadiofrequency.
- Under Preparation *This paper will present the pre-clinical protocol used for the Radiofrequency Ablation of the surrogate tumors implanted on pig liver.*

Peer-reviewed Conference Papers

- [Audigier 2013a] *This paper details the presented RFA model and its validation against analytical solution.*
Audigier, Chloé and Mansi, Tommaso and Delingette, Hervé and Rapaka, Saikiran and Mihalef, Viorel and Sharma, Puneet and Kamen, Ali and Carnegie, Daniel and Boctor, Emad and Choti, Michael and Comaniciu, Dorin and Ayache, Nicholas. Lattice Boltzmann Method for Fast Patient-Specific Simulation of Liver Tumor Ablation from CT images. *In Medical Image Computing and Computer-Assisted Intervention – MICCAI 2013* (pp. 323-330). Springer Berlin Heidelberg.
- [Audigier 2014a] *This paper presents an improved RFA model and its parameter estimation. The personalization is evaluated on clinical data.*
Audigier, Chloé and Mansi, Tommaso and Delingette, Hervé and Rapaka, Saikiran and Mihalef, Viorel, and Carnegie, Daniel and Boctor, Emad and Choti, Michael and Kamen, Ali and Comaniciu, Dorin and Ayache, Nicholas.

Parameter Estimation For Personalization of Liver Tumor Radiofrequency Ablation. *In MICCAI Workshop on Abdominal Imaging–Computational and Clinical Applications 2014 (pp. 3-12).*

- [Audigier 2015b] *This paper details the evaluation of the RFA model on pre-clinical data.*

Audigier, Chloé and Mansi, Tommaso and Delingette, Hervé and Rapaka, Saikiran and Passerini, Tiziano and Mihalef, Viorel and Pop, Raoul and Diana, Michele and Soler, Luc and Kamen, Ali and Comaniciu, Dorin and Ayache, Nicholas. Challenges to Validate Multi-physics Model of Liver Tumor Radiofrequency Ablation from Pre-clinical Data. *In MICCAI Workshop on Computational Biomechanics for Medicine X, Springer International Publishing, 2015, pp.29-40.*

Patent Application

- [Audigier 2013b] *This US patent describes the method for patient-specific modeling of RFA.*

Audigier, Chloé and Mansi, Tommaso and Mihalef, Viorel and Kamen, Ali and Comaniciu, Dorin and Sharma, Puneet and Rapaka, Saikiran and Delingette, Hervé and Ayache, Nicholas. System and Method for Patient Specific Modeling of Liver Tumor Ablation. *US Patent App. 14/071,688, 5 nov. 2013.*

- [Audigier 2014b] *This US patent describes the interactive patient specific simulation of radiofrequency ablation therapy of RFA.*

Audigier, Chloé and Mansi, Tommaso and Mihalef, Viorel and Kamen, Ali and Comaniciu, Dorin and Sharma, Puneet and Rapaka, Saikiran. System and Method for Interactive Patient Specific Simulation of Radiofrequency Ablation Therapy. *PCT/US2014/017886, 24 feb. 2014.*

- [Audigier 2015c] *This US patent describes personalized computation of tissue ablation extent based on medical images.*

Audigier, Chloé and Mansi, Tommaso and Rapaka, Saikiran and Kamen, Ali and Mihalef, Viorel and Delingette, Hervé and Ayache, Nicholas and Comaniciu, Dorin. System and Method for Personalized Computation of Tissue Ablation Extent Based on Medical Images. *US2015/0242588 A1, 27 aug. 2015.*

Invited Talks

- Chloé Audigier, Simulation of radio-frequency ablation. *Second Scientific IHU Day about Computational Simulation for Medicine, April, 3rd, 2013, IRCAD, Strasbourg.*

- Chloé Audigier, Parameter Estimation For Personalization of Liver Tumor Radiofrequency Ablation. *Siemens Corporate Research, December, 4th, 2014, Princeton, USA.*

Awards

- **Outstanding Paper Award** at the 6th International Workshop on Abdominal Imaging: Computational and Clinical Applications, 2014.

The Lattice Boltzmann Method

A.1 The Boltzmann Equation

A.1.1 Boltzmann Transport Equation

A statistical description of a system can be explained by the distribution function (probability of finding particles within a certain range of velocities at a certain range of locations at a given time) $f(\mathbf{r}, \mathbf{c}, t)$ where $f(\mathbf{r}, \mathbf{c}, t)$ is the number of molecules at time t positioned between \mathbf{r} and $\mathbf{r} + d\mathbf{r}$ which have velocities between \mathbf{c} and $\mathbf{c} + d\mathbf{c}$ [Mohamad 2011]. An external force \mathbf{F} will change the velocity of the molecule from \mathbf{c} to $\mathbf{c} + \mathbf{F}dt$ and its position from \mathbf{r} to $\mathbf{r} + \mathbf{c}dt$. The number of molecules $f(\mathbf{r}, \mathbf{c}, t)$, before applying the external force is equal to the number of molecules after the disturbance $f(\mathbf{r} + \mathbf{c}dt, \mathbf{c} + \mathbf{F}dt, t + dt)$ if no collisions take place between the molecules:

$$f(\mathbf{r} + \mathbf{c}dt, \mathbf{c} + \mathbf{F}dt, t + dt)d\mathbf{r}d\mathbf{c} - f(\mathbf{r}, \mathbf{c}, t)d\mathbf{r}d\mathbf{c} = 0$$

However, collisions can take place between the molecules. The rate of change between final and initial status of the distribution function is called collision operator, Ω . Hence:

$$f(\mathbf{r} + \mathbf{c}dt, \mathbf{c} + \mathbf{F}dt, t + dt)d\mathbf{r}d\mathbf{c} - f(\mathbf{r}, \mathbf{c}, t)d\mathbf{r}d\mathbf{c} = \Omega(f)d\mathbf{r}d\mathbf{c}dt$$

Dividing the above equation by $d\mathbf{r}d\mathbf{c}dt$ and as the limit $dt \rightarrow 0$, yields

$$\frac{df}{dt} = \Omega(f)$$

This equation states that the total rate of change of the distribution function is equal to the rate of the collision. The total rate of change can be expressed as:

$$df = \frac{\partial f}{\partial \mathbf{r}}d\mathbf{r} + \frac{\partial f}{\partial \mathbf{c}}d\mathbf{c} + \frac{\partial f}{\partial t}dt$$

Dividing by dt :

$$\frac{df}{dt} = \frac{\partial f}{\partial \mathbf{r}}\frac{d\mathbf{r}}{dt} + \frac{\partial f}{\partial \mathbf{c}}\frac{d\mathbf{c}}{dt} + \frac{\partial f}{\partial t}$$

And:

$$\frac{df}{dt} = \frac{\partial f}{\partial \mathbf{r}}\mathbf{c} + \frac{\partial f}{\partial \mathbf{c}}\mathbf{a} + \frac{\partial f}{\partial t}$$

where \mathbf{a} is $\frac{d\mathbf{c}}{dt}$, the acceleration and can be related to force \mathbf{F} by Newton's law:

$$\mathbf{a} = \frac{\mathbf{F}}{m}$$

Therefore, the Boltzmann transport equation can be written as:

$$\frac{\partial f}{\partial t} + \frac{\partial f}{\partial \mathbf{r}} \cdot \mathbf{c} + \frac{\partial f}{\partial \mathbf{c}} \cdot \frac{\mathbf{F}}{m} = \Omega(f)$$

The relations between the above equation and macroscopic quantities such as fluid temperature T , fluid velocity vector \mathbf{u} are:

$$T(\mathbf{r}, t) = \int f(\mathbf{r}, \mathbf{c}, t) d\mathbf{c}$$

$$T(\mathbf{r}, t) \mathbf{u}(\mathbf{r}, t) = \int \mathbf{c} f(\mathbf{r}, \mathbf{c}, t) d\mathbf{c}$$

A.1.2 The BGKW Approximation

It is difficult to solve Boltzmann equation because the collision term is very complicated. Bhatnagar, Gross and Krook (BGK) in 1954 introduced a simplified model for collision operator:

$$\Omega(f) = \frac{1}{\tau} (f^{eq} - f)$$

where τ is called the relaxation factor. The local equilibrium distribution function is denoted by f^{eq} , which is Maxwell-Boltzmann distribution function.

After introducing BGKW approximation, the Boltzmann equation (without external forces) can be approximated as:

$$\frac{\partial f}{\partial t} + \mathbf{c} \cdot \nabla f = \frac{1}{\tau} (f^{eq} - f)$$

In Lattice Boltzmann Method, the above equation is discretized and assumed valid along specific directions. Hence, along a specific direction, the discrete Boltzmann equation can be written:

$$\frac{\partial f_i}{\partial t} + \mathbf{c} \cdot \nabla f_i = \frac{1}{\tau} (f_i^{eq} - f_i)$$

This equation can be discretized as:

$$f_i(\mathbf{r} + \mathbf{c}_i \Delta t, t + \Delta t) = f_i(\mathbf{r}, t) + \frac{\Delta t}{\tau} [f_i^{eq}(\mathbf{r}, t) - f_i(\mathbf{r}, t)]$$

Let us consider:

$$w = \frac{\Delta t}{\tau}$$

Then:

$$f_i(\mathbf{r} + \mathbf{c}_i \Delta t, t + \Delta t) = f_i(\mathbf{r}, t) + w [f_i^{eq}(\mathbf{r}, t) - f_i(\mathbf{r}, t)]$$

A.1.3 Equilibrium Distribution Function

The key element in applying LBM for different problems is the equilibrium distribution function f^{eq} . For particles moving in a medium with macroscopic velocity \mathbf{u} , the normalized Maxwell's distribution function can be written as:

$$f^{eq} = \frac{T}{2\pi/3} e^{-\frac{3}{2}(\mathbf{c}-\mathbf{u})^2}$$

which can be written as,

$$f^{eq} = \frac{T}{2\pi/3} e^{-\frac{3}{2}c^2} e^{3(\mathbf{c}\cdot\mathbf{u}-u^2/2)}$$

where $c^2 = \mathbf{c}\cdot\mathbf{c}$ and $u^2 = \mathbf{u}\cdot\mathbf{u}$.

Using Taylor series expansion for e^x and expanded around the stationary state:

$$f^{eq} = \frac{T}{2\pi/3} e^{-\frac{3}{2}c^2} [1 + 3\mathbf{c}\cdot\mathbf{u} - \frac{3}{2}u^2 + \dots]$$

Let us write the general form of the equilibrium function as:

$$f_i^{eq} = T\omega_i[A + B\mathbf{c}_i\cdot\mathbf{u} + C(\mathbf{c}_i\cdot\mathbf{u})^2 + Du^2 + \dots]$$

where A, B, C, and D are constants and need to be determined. T stands for scalar parameter (in our case: the fluid temperature), which is equal to summation of all the distribution functions, i.e:

$$T = \sum_{i=1}^N f_i^{eq}$$

where N is the number of lattice links.

In our case, we will just consider:

$$f_i^{eq} = T\omega_i[A + B\mathbf{c}_i\cdot\mathbf{u}]$$

And let us consider:

$$\mathbf{c}_i = c\mathbf{e}_i$$

where $\mathbf{c} = d\mathbf{r}/dt$ and \mathbf{e}_i are vectors with coefficient 1 or -1.

We have the following properties (called the isotropic properties) which should be true to keep the lattices symmetries [Chai 2013]:

$$\begin{aligned} \sum_{i=1}^N \omega_i &= 1 \\ \sum_{i=1}^N \omega_i \mathbf{e}_i &= \mathbf{0} \\ \sum_{i=1}^N \omega_i \mathbf{e}_i^T \cdot \mathbf{e}_i &= c_s^2 \mathbf{I} \end{aligned}$$

where $\mathbf{0}$ and \mathbf{I} are respectively the zero matrix and identity matrix. The value of c_s vary from one configuration to another (depends on the number of lattices). The last equation can also be written as :

$$\sum_{i=1}^N \omega_i e_{i,\alpha} e_{i,\beta} = c_s^2 \delta_{\alpha,\beta}$$

where $\delta_{\alpha,\beta}$ is the Kronecker's delta.

The last isotropic property is:

$$\forall (\alpha, \beta, \gamma) \in \{1, \dots, M\} \sum_{i=1}^N \omega_i e_{i,\alpha} e_{i,\beta} e_{i,\gamma} = 0$$

The first momentum equation is:

$$T = \sum_{i=1}^N f_i^{eq}$$

By taking the sum, we get:

$$A = 1$$

The second momentum equation gives:

$$T \mathbf{u} = \sum_{i=1}^N f_i^{eq} c \mathbf{e}_i$$

So:

$$T \mathbf{u} = \sum_{i=1}^N T \omega_i [1 + B c_i \cdot \mathbf{u}] c \mathbf{e}_i = \sum_{i=1}^N T \omega_i [B c_i \cdot \mathbf{u}] c \mathbf{e}_i$$

By expanding the scalar product:

$$\mathbf{u} = B c^2 \sum_{i=1}^N \omega_i \left[\sum_{\alpha=1}^M e_{i,\alpha} u_\alpha \right] \mathbf{e}_i$$

where M is the spatial dimension of the problem. For each coordinates we get:

$$\forall \beta \in [1, \dots, M] \quad u_\beta = B c^2 \sum_{i=1}^N \omega_i \sum_{\alpha=1}^M e_{i,\alpha} u_\alpha e_{i,\beta}$$

$$\forall \beta \in [1, \dots, M] \quad u_\beta = B c^2 \sum_{\alpha=1}^M u_\alpha \sum_{i=1}^N \omega_i e_{i,\alpha} e_{i,\beta}$$

According to the last isotropic properties, we get:

$$B = \frac{1}{c_s^2 c^2}$$

Due to the last isotropic property, we have:

$$\sum_{i=0}^N f_i^{eq} \mathbf{e}_i^2 = c^2 T \sum_{i=0}^N w_i [1 + \mathbf{c}_i \cdot \mathbf{u}] \mathbf{e}_i^2 = c^2 T [c_s^2 + \sum_{i=0}^N w_i \mathbf{c}_i \cdot \mathbf{u} \mathbf{e}_i^2] = c^2 T [c_s^2 + \sum_{i=0}^N w_i \sum_{\alpha=0}^M c_{i,\alpha} u_\alpha \sum_{\beta=0}^M e_{i,\beta}^2]$$

The last term can be written as:

$$c^3 \sum_{i=0}^N w_i \sum_{\alpha=0}^M e_{i,\alpha} u_\alpha \sum_{\beta=0}^M e_{i,\beta}^2 = c^3 \sum_{\beta=0}^M \sum_{\alpha=0}^M \sum_{i=0}^N w_i e_{i,\alpha} e_{i,\beta}^2 u_\alpha = 0$$

The third momentum is:

$$T c^2 c_s^2 = \sum_{i=0}^N f_i^{eq} c^2 \mathbf{e}_i^2$$

This could also be written in a matrix, since:

$$\sum_{i=0}^N f_i^{eq} \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} = T \sum_{i=0}^N w_i [1 + \mathbf{c}_i \cdot \mathbf{u}] \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i}$$

As:

$$(\mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i})_{\alpha,\beta} = \sum_{k=0}^M (\mathbf{t}_{\mathbf{e}_i})_{\alpha,k} (\mathbf{e}_i)_{k,\beta} = e_{i,\alpha} e_{i,\beta}$$

Each element of the matrix can be written as:

$$\left(\sum_{i=0}^N f_i^{eq} \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} \right)_{\alpha,\beta} = T \sum_{i=0}^N w_i [1 + \mathbf{c}_i \cdot \mathbf{u}] e_{i,\alpha} e_{i,\beta} = T c_s^2 \delta_{\alpha,\beta} + T c \sum_{\gamma=0}^M u_\gamma \sum_{i=0}^N w_i e_{i,\gamma} e_{i,\alpha} e_{i,\beta}$$

The last term is zero, so:

$$T c^2 c_s^2 \mathbf{I} = \sum_{i=0}^N f_i^{eq} c^2 \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i}$$

A.2 Advection-Diffusion Equation

A.2.1 1D Formulation

In 1D, the advection-diffusion process can be written as,

$$\frac{\partial T}{\partial t} + u \frac{\partial T}{\partial x} = d \frac{\partial^2 T}{\partial x^2}$$

where u is the velocity and d the diffusion coefficient.

The Lattice Boltzmann equation for advection-diffusion problem is still the same:

$$f_i(x + e_i \Delta x, t + \Delta t) = f_i(x, t) + w [f_i^{eq}(x, t) - f_i(x, t)] \quad (\text{A.1})$$

with:

$$f_i^{eq}(x, t) = \omega_i T(x, t) \left[1 + \frac{e_i u}{c c_s^2} \right]$$

where $c = \frac{\Delta x}{\Delta t}$

Chapman-Enskog Expansion

The advection-diffusion equation can be scaled spatially : x is set to x/ε , where ε is a small parameter:

$$\frac{\partial T}{\partial t} + u\varepsilon \frac{\partial T}{\partial x} = d\varepsilon^2 \frac{\partial^2 T}{\partial x^2}$$

Now, if we scale time t with t/ε^2 as it is done for diffusion equation, then the first term on the left-hand side of the equation balances the diffusion term on the right hand side of the equation, hence the effect of advection term will be neglected. If we scale time t with $1/\varepsilon$, in this case there will be a balance between the first and the second term of the left-hand side of the equation, and the effect of diffusion term will be neglected. To resolve this issue, we need to introduce dual time scales, i.e:

$$\frac{\partial}{\partial t} = \varepsilon \frac{\partial}{\partial t_1} + \varepsilon^2 \frac{\partial}{\partial t_2}$$

Hence, the first term in the above equation balances the advection term and the second term balances the diffusion term.

For one-dimensional problem with N lattice velocity components ($c_i, i = 1, 2, \dots, N$), the temperature can be expressed as,

$$T(x, t) = \sum_{i=1}^N f_i(x, t)$$

Equation A.1 can be expanded using Taylor series:

$$f_i(x+c_i\Delta t, t+\Delta t) = f_i(x, t) + \frac{\partial f_i}{\partial t} \Delta t + \frac{\partial f_i}{\partial x} c_i \Delta t + \frac{\Delta t^2}{2} \left(\frac{\partial^2 f_i}{\partial t^2} + 2 \frac{\partial^2 f_i}{\partial t \partial x} c_i + \frac{\partial^2 f_i}{\partial x^2} c_i^2 \right) + O(\Delta t)^3$$

Introducing scaling, keeping term up to ε^2 we get:

$$\begin{aligned} f_i(x + ce_i\Delta t, t + \Delta t) &= f_i(x, t) + \varepsilon \frac{\partial f_i}{\partial t_1} \Delta t + \varepsilon^2 \frac{\partial f_i}{\partial t_2} \Delta t + \varepsilon \frac{\partial f_i}{\partial x} ce_i \Delta t \\ &+ \frac{\Delta t^2}{2} \left(\varepsilon^2 \frac{\partial^2 f_i}{\partial t_1^2} + 2\varepsilon^2 \frac{\partial^2 f_i}{\partial t_1 \partial x} ce_i + \varepsilon^2 \frac{\partial^2 f_i}{\partial x^2} c^2 e_i^2 \right) + O(\Delta t)^3 \end{aligned}$$

Substituting this equation in equation A.1:

$$\varepsilon \frac{\partial f_i}{\partial t_1} \Delta t + \varepsilon^2 \frac{\partial f_i}{\partial t_2} \Delta t + \varepsilon \frac{\partial f_i}{\partial x} ce_i \Delta t + \frac{\Delta t^2}{2} \left(\varepsilon^2 \frac{\partial^2 f_i}{\partial t_1^2} + 2\varepsilon^2 \frac{\partial^2 f_i}{\partial t_1 \partial x} ce_i + \varepsilon^2 \frac{\partial^2 f_i}{\partial x^2} c^2 e_i^2 \right) + O(\Delta t)^3 = w[f_i^{eq}(x, t) - f_i(x, t)]$$

The distribution function can be expanded in terms of the small parameter ε as,

$$f_i(x, t) = f_i^0(x, t) + \varepsilon f_i^1(x, t) + \varepsilon^2 f_i^2(x, t) + O(\varepsilon^3)$$

Collecting the term of order ε^0 gives:

$$f_i^0(x, t) = f_i^{eq}(x, t)$$

Since:

$$\sum_{i=1}^N f_i^{eq}(x, t) = T(x, t)$$

Then:

$$\sum_{i=1}^N f_i^0(x, t) = T(x, t)$$

As:

$$\sum_{i=1}^N f_i(x, t) = T(x, t)$$

This first equation gives:

$$\sum_{i=1}^N f_i^j(x, t) = 0 \text{ for } j \neq 0$$

Collecting the term of order ε^1 gives:

$$\frac{\partial f_i^0}{\partial t_1} \Delta t + \frac{\partial f_i^0}{\partial x} c e_i \Delta t = -w f_i^1 = -\frac{\Delta t}{\tau} f_i^1$$

Hence:

$$\frac{\partial f_i^0}{\partial t_1} + \frac{\partial f_i^0}{\partial x} c e_i = -\frac{1}{\tau} f_i^1$$

By taking the summation over all the values of i, we obtain:

$$\sum_{i=1}^N \frac{\partial f_i^0}{\partial t_1} + \sum_{i=1}^N \frac{\partial f_i^0}{\partial x} c e_i = -\frac{1}{\tau} \sum_{i=1}^N f_i^1$$

Finally:

$$\sum_{i=1}^N \frac{\partial f_i^0}{\partial t_1} + \sum_{i=1}^N \frac{\partial f_i^0}{\partial x} c e_i = 0 \tag{A.2}$$

since the right hand side is zero, and according to the momentum equations:

$$\begin{aligned} \sum_{i=1}^N f_i^0 &= T \\ \sum_{i=1}^N f_i^0 c_i &= T u \end{aligned}$$

The above equation formulate as:

$$\frac{\partial T}{\partial t_1} + \frac{\partial(uT)}{\partial x} = 0 \quad (\text{A.3})$$

which is the advection equation.

Latter we will need the derivative of f_i^1 with respect to t_1 and x :

$$-\frac{1}{\tau} \frac{\partial f_i^1}{\partial t_1} = \frac{\partial^2 f_i^0}{\partial t_1^2} + \frac{\partial^2 f_i^0}{\partial x \partial t_1} c e_i$$

and:

$$-\frac{1}{\tau} \frac{\partial f_i^1}{\partial x} = \frac{\partial^2 f_i^0}{\partial t_1 \partial x} + \frac{\partial^2 f_i^0}{\partial x^2} c e_i$$

These two equation gives:

$$-\frac{1}{\tau} \left(\frac{\partial f_i^1}{\partial t_1} + c e_i \frac{\partial f_i^1}{\partial x} \right) = \frac{\partial^2 f_i^0}{\partial t_1^2} + 2 \frac{\partial^2 f_i^0}{\partial t_1 \partial x} c e_i + \frac{\partial^2 f_i^0}{\partial x^2} c e_i^2 \quad (\text{A.4})$$

Collecting the term of order ε^2 gives:

$$\frac{\partial f_i^1}{\partial t_1} + \frac{\partial f_i^0}{\partial t_2} + \frac{\partial f_i^1}{\partial x} c e_i + \frac{\Delta t}{2} \left(\frac{\partial^2 f_i^0}{\partial t_1^2} + 2 \frac{\partial^2 f_i^0}{\partial t_1 \partial x} c e_i + \frac{\partial^2 f_i^0}{\partial x^2} c^2 e_i^2 \right) = -\frac{1}{\tau} f_i^2$$

Using equation A.4

$$\frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial^2 f_i^0}{\partial t_1^2} + 2 \frac{\partial^2 f_i^0}{\partial t_1 \partial x} c e_i + \frac{\partial^2 f_i^0}{\partial x^2} c^2 e_i^2 \right) = -\frac{1}{\tau} f_i^2$$

Hence:

$$\frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial}{\partial t_1} \left(\frac{\partial f_i^0}{\partial t_1} + \frac{\partial f_i^0}{\partial x} c_i \right) + \frac{\partial^2 f_i^0}{\partial t_1 \partial x} c e_i + \frac{\partial^2 f_i^0}{\partial x^2} c^2 e_i^2 \right) = -\frac{1}{\tau} f_i^2$$

Summing the above equation over all states:

$$\sum_{i=1}^N \frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial}{\partial t_1} \left(\sum_{i=1}^N \frac{\partial f_i^0}{\partial t_1} + \sum_{i=1}^N \frac{\partial f_i^0}{\partial x} c e_i \right) + \sum_{i=1}^N \frac{\partial^2 f_i^0}{\partial t_1 \partial x} c e_i + \sum_{i=1}^N \frac{\partial^2 f_i^0}{\partial x^2} c^2 e_i^2 \right) = -\frac{1}{\tau} \sum_{i=1}^N f_i^2$$

The right hand side is zero as well as the second term of the left hand side according to equation A.2.

$$\sum_{i=1}^N \frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\sum_{i=1}^N \frac{\partial^2 f_i^0}{\partial t_1 \partial x} c e_i + \sum_{i=1}^N \frac{\partial^2 f_i^0}{\partial x^2} c^2 e_i^2 \right) = 0$$

And:

$$\sum_{i=1}^N f_i^0 c_i^2 = T c_s^2$$

So we get:

$$\frac{\partial T}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau\right) \left(\frac{\partial^2 u T}{\partial t_1 \partial x} + c^2 c_s^2 \frac{\partial^2 T}{\partial x^2}\right) = 0$$

Finally:

$$\frac{\partial T}{\partial t_2} + c^2 c_s^2 \left(\frac{\Delta t}{2} - \tau\right) \frac{\partial^2 T}{\partial x^2} + \left(\frac{\Delta t}{2} - \tau\right) \frac{\partial^2 u T}{\partial t_1 \partial x} = 0 \quad (\text{A.5})$$

So by adding these 2 equations A.3 + A.5, we get:

$$\frac{\partial T}{\partial t} + \frac{\partial(uT)}{\partial x} + \varepsilon \left(\frac{\Delta t}{2} - \tau\right) \frac{\partial^2 u T}{\partial t_1 \partial x} = c^2 c_s^2 \left(\tau - \frac{\Delta t}{2}\right) \frac{\partial^2 T}{\partial x^2}$$

There is an error term: $\varepsilon \left(\frac{\Delta t}{2} - \tau\right) \frac{\partial^2 u T}{\partial t_1 \partial x}$. The relaxation parameter can be related to the diffusion coefficient as:

$$d = \left(\tau - \frac{\Delta t}{2}\right) c_s^2 c^2$$

A.2.2 3D Formulation

In 3D, the advection-diffusion process can be written as,

$$\frac{\partial T}{\partial t} + \mathbf{u} \cdot \nabla T = \nabla \cdot (d \nabla T)$$

where \mathbf{u} is the velocity and d the diffusion coefficient.

The Lattice Boltzmann equation for advection-diffusion problem is still the same:

$$f_i(\mathbf{p} + \mathbf{e}_i \Delta p, t + \Delta t) = f_i(\mathbf{p}, t) + w [f_i^{eq}(\mathbf{p}, t) - f_i(\mathbf{p}, t)] \quad (\text{A.6})$$

where $\mathbf{p} = (x, y, z)$ and $\Delta p = \Delta x = \Delta y = \Delta z$ with:

$$f_i^{eq}(\mathbf{p}, t) = \omega_i T(\mathbf{p}, t) \left[1 + \frac{\mathbf{e}_i \cdot \mathbf{u}}{c c_s^2}\right]$$

where $c = \frac{\Delta p}{\Delta t}$

Chapman-Enskog Expansion

The advection-diffusion equation can be scaled spatially : \mathbf{p} is set to \mathbf{p}/ε , where ε is a small parameter:

$$\frac{\partial T}{\partial t} + \varepsilon \mathbf{u} \cdot \nabla T = \varepsilon^2 \nabla \cdot (d \nabla T)$$

Now, as in 1D, we introduce dual time scales, i.e:

$$\frac{\partial}{\partial t} = \varepsilon \frac{\partial}{\partial t_1} + \varepsilon^2 \frac{\partial}{\partial t_2}$$

For three-dimensional problem with N lattice velocity components ($c_i, i = 0..N$), the temperature can be expressed as,

$$T(x, t) = \sum_{i=0}^N f_i(x, t)$$

Equation A.6 can be expanded using Taylor series:

$$f_i(\mathbf{p} + c\mathbf{e}_i\Delta t, t + \Delta t) = f_i(\mathbf{p}, t) + \frac{\partial f_i}{\partial t}\Delta t + \Delta t c\mathbf{e}_i \cdot \nabla f_i + \frac{\Delta t^2}{2} \left(\frac{\partial^2 f_i}{\partial t^2} + 2 \frac{\partial}{\partial t} (c\mathbf{e}_i \cdot \nabla f_i) + c^2 \nabla \nabla : \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i}(f_i) \right) + 0(\Delta t)^3$$

where the operator $:$ represents tensor contraction.

Introducing scaling, keeping term up to ε^2 we get:

$$f_i(\mathbf{p} + c\mathbf{e}_i\Delta t, t + \Delta t) = f_i(\mathbf{p}, t) + \varepsilon \frac{\partial f_i}{\partial t_1}\Delta t + \varepsilon^2 \frac{\partial f_i}{\partial t_2}\Delta t + \varepsilon \mathbf{e}_i \cdot \nabla f_i c\Delta t + \frac{\Delta t^2}{2} \left(\varepsilon^2 \frac{\partial^2 f_i}{\partial t_1^2} + 2\varepsilon^2 \frac{\partial}{\partial t_1} (c\mathbf{e}_i \cdot \nabla f_i) + \varepsilon^2 c^2 \nabla \nabla : f_i \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} \right) + 0(\Delta t)^3$$

Substituting this equation in equation A.6:

$$\varepsilon \frac{\partial f_i}{\partial t_1}\Delta t + \varepsilon^2 \frac{\partial f_i}{\partial t_2}\Delta t + \varepsilon \mathbf{e}_i \cdot \nabla f_i c\Delta t + \frac{\Delta t^2}{2} \left(\varepsilon^2 \frac{\partial^2 f_i}{\partial t_1^2} + 2\varepsilon^2 \frac{\partial}{\partial t_1} (c\mathbf{e}_i \cdot \nabla f_i) + \varepsilon^2 c^2 \nabla \nabla : f_i \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} \right) + 0(\Delta t)^3 = w[f_i^{eq}(\mathbf{p}, t) - f_i(\mathbf{p}, t)]$$

The distribution function can be expanded in terms of the small parameter ε as:

$$f_i(\mathbf{p}, t) = f_i^0(\mathbf{p}, t) + \varepsilon f_i^1(\mathbf{p}, t) + \varepsilon^2 f_i^2(\mathbf{p}, t) + O(\varepsilon^3)$$

Collecting the term of order ε^0 gives:

$$f_i^0(\mathbf{p}, t) = f_i^{eq}(\mathbf{p}, t)$$

Since:

$$\sum_{i=0}^N f_i^{eq}(\mathbf{p}, t) = T(\mathbf{p}, t)$$

Then:

$$\sum_{i=0}^N f_i^0(\mathbf{p}, t) = T(\mathbf{p}, t)$$

As:

$$\sum_{i=0}^N f_i(\mathbf{p}, t) = T(\mathbf{p}, t)$$

This first equation gives:

$$\sum_{i=0}^N f_i^j(\mathbf{p}, t) = 0 \text{ for } j \neq 0$$

Collecting the term of order ε^1 gives:

$$\frac{\partial f_i^0}{\partial t_1} \Delta t + \Delta t c \mathbf{e}_i \cdot \nabla f_i^0 = -w f_i^1 = -\frac{\Delta t}{\tau} f_i^1$$

Hence:

$$\frac{\partial f_i^0}{\partial t_1} + c \mathbf{e}_i \cdot \nabla f_i^0 = -\frac{1}{\tau} f_i^1$$

By taking the summation over all the values of i , we obtain:

$$\sum_{i=0}^N \frac{\partial f_i^0}{\partial t_1} + \sum_{i=0}^N \mathbf{e}_i \cdot \nabla f_i^0 c = -\frac{1}{\tau} \sum_{i=0}^N f_i^1$$

Finally:

$$\sum_{i=0}^N \frac{\partial f_i^0}{\partial t_1} + \nabla \cdot \left(\sum_{i=0}^N c \mathbf{e}_i f_i^0 \right) = 0 \quad (\text{A.7})$$

since the right hand side is zero, and according to the momentum equations:

$$\begin{aligned} \sum_{i=0}^N f_i^0 &= T \\ \sum_{i=0}^N f_i^0 \mathbf{c}_i &= T \mathbf{u} \end{aligned}$$

Indeed it is assumed here that the fluid is incompressible:

$$\nabla \cdot \mathbf{u} = 0$$

The above equation formulate as:

$$\frac{\partial T}{\partial t_1} + \mathbf{u} \cdot \nabla T = 0$$

which is the advection equation.

Latter we will need the derivative of f_i^1 with respect to t_1 and space:

$$-\frac{1}{\tau} \frac{\partial f_i^1}{\partial t_1} = \frac{\partial^2 f_i^0}{\partial t_1^2} + \frac{\partial}{\partial t_1} (c \mathbf{e}_i \cdot \nabla f_i^0)$$

and:

$$-\frac{1}{\tau} c \mathbf{e}_i \cdot \nabla f_i^1 = \frac{\partial}{\partial t_1} (c \mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i \mathbf{t}_{\mathbf{e}_i \mathbf{e}_i}$$

These two equation gives:

$$-\frac{1}{\tau} \left(\frac{\partial f_i^1}{\partial t_1} + c \mathbf{e}_i \cdot \nabla f_i^1 \right) = \frac{\partial^2 f_i^0}{\partial t_1^2} + 2 \frac{\partial}{\partial t_1} (c \mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i \mathbf{t}_{\mathbf{e}_i \mathbf{e}_i}$$

Collecting the term of order ε^2 gives:

$$\frac{\partial f_i^1}{\partial t_1} + \frac{\partial f_i^0}{\partial t_2} + c\mathbf{e}_i \cdot \nabla f_i^1 + \frac{\Delta t}{2} \left(\frac{\partial^2 f_i^0}{\partial t_1^2} + 2 \frac{\partial}{\partial t_1} (c\mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i^0 \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} \right) = -\frac{1}{\tau} f_i^2$$

Using equation A.7:

$$\frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial^2 f_i^0}{\partial t_1^2} + 2 \frac{\partial}{\partial t_1} (c\mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i^0 \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} \right) = -\frac{1}{\tau} f_i^2$$

Hence:

$$\frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial}{\partial t_1} \left(\frac{\partial f_i^0}{\partial t_1} + c\mathbf{e}_i \cdot \nabla f_i^0 \right) + \frac{\partial}{\partial t_1} (c\mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i^0 \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} \right) = -\frac{1}{\tau} f_i^2$$

Summing the above equation over all states:

$$\begin{aligned} \sum_{i=0}^N \frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial}{\partial t_1} \left(\sum_{i=0}^N \frac{\partial f_i^0}{\partial t_1} + c \sum_{i=0}^N \mathbf{e}_i \cdot \nabla f_i^0 \right) + \frac{\partial}{\partial t_1} \left(\sum_{i=0}^N c\mathbf{e}_i \cdot \nabla f_i^0 \right) + c^2 \nabla \nabla : \sum_{i=0}^N f_i^0 \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} \right) \\ = -\frac{1}{\tau} \sum_{i=0}^N f_i^2 \end{aligned}$$

The right hand side is zero as well as the second term of the left hand side according to equation A.7.

$$\sum_{i=0}^N \frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial}{\partial t_1} \left(\sum_{i=0}^N c\mathbf{e}_i \cdot \nabla f_i^0 \right) + c^2 \nabla \nabla : \sum_{i=0}^N f_i^0 \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} \right) = 0$$

And:

$$\sum_{i=0}^N f_i^0 \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} = T c_s^2 \mathbf{I}$$

So we get:

$$\frac{\partial T}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial}{\partial t_1} [\nabla \cdot (\mathbf{u}T)] + c^2 c_s^2 \nabla^2 T \right) = 0$$

Finally:

$$\frac{\partial T}{\partial t_2} + c^2 c_s^2 \left(\frac{\Delta t}{2} - \tau \right) \nabla^2 T + \left(\frac{\Delta t}{2} - \tau \right) \frac{\partial}{\partial t_1} [\nabla \cdot (\mathbf{u}T)] = 0$$

So by adding the first equation times ε and the second one times ε^2 , we get:

$$\frac{\partial T}{\partial t} + \nabla \cdot (\mathbf{u}T) = c^2 c_s^2 \left(\tau - \frac{\Delta t}{2} \right) \nabla^2 T + \varepsilon \left(\tau - \frac{\Delta t}{2} \right) \frac{\partial}{\partial t_1} [\nabla \cdot (\mathbf{u}T)]$$

There is an error term: $\varepsilon(\tau - \frac{\Delta t}{2})\frac{\partial}{\partial t_1}[\nabla \cdot (\mathbf{u}T)]$. The relaxation parameter can be related to the diffusion coefficient as:

$$d = (\tau - \frac{\Delta t}{2})c_s^2 c^2$$

Let us consider the error term:

$$E = \varepsilon(\tau - \frac{\Delta t}{2})\frac{\partial}{\partial t_1}[\nabla \cdot (\mathbf{u}T)] = \varepsilon\frac{d}{c_s^2 c^2}\frac{\partial}{\partial t_1}[\nabla \cdot (\mathbf{u}T)]$$

To interpret this term, let us consider \mathbf{u} as a constant, by taking the gradient of the advection equation we found and multiplying by \mathbf{u} , we get:

$$\varepsilon\mathbf{u} \cdot \nabla \left(\frac{\partial T}{\partial t_1} \right) + \varepsilon^2 u^2 \nabla^2 T = 0$$

Thus the error term can be written as:

$$E = -\varepsilon^2 \frac{d}{c_s^2 c^2} u^2 \nabla^2 T$$

And then the second equation becomes:

$$\frac{\partial T}{\partial t} + \nabla \cdot (uT) = d \left(1 - \frac{u^2}{c_s^2 c^2} \right) \nabla^2 T$$

Therefore, the error term amounts to an undesired correction to the diffusion coefficient, proportional to the square of the advection velocity [Chopard 2009]. Parameters d and \mathbf{u} are given by the physical problem, they are independent of the chosen time and space discretization Δp and Δt . Thus the correction to the diffusion coefficient scale as $(\frac{\Delta t}{\Delta p})^2$. If we take the limits $\Delta t \rightarrow 0$ and $\Delta p \rightarrow 0$, with $\Delta p^2/\Delta t$ constant (natural limit for a diffusion process), then the correction to the diffusion coefficient is of order $O(\Delta t)$. The scheme is thus first order accurate in time and second order accurate in space. But if we take $\Delta t \rightarrow 0$ and $\Delta p \rightarrow 0$, with $\Delta p/\Delta t$ constant, then the error is of order $O(\varepsilon^0)$ and the scheme is zeroth order accurate.

The error term can be eliminated if we consider:

$$\Omega(f_i) = \frac{1}{\tau}(f_i^{eq} - f_i) + t_i \mathbf{c}_i \cdot \delta \mathbf{j}_i$$

A.3 Implementation

We use D3Q7 configuration, and for each lattice site, the seven distribution functions are calculated according to :

$$f_i(\mathbf{r} + \mathbf{c}_i \Delta t, t + \Delta t) = f_i(\mathbf{r}, t) + w[f_i^{eq}(\mathbf{r}, t) - f_i(\mathbf{r}, t)]$$

with:

$$\begin{aligned}
 w &= \frac{\Delta t}{\tau} \\
 \tau &= \frac{\Delta t}{2} + \frac{d}{c_s^2 c^2} \\
 f_i^{eq} &= \omega_i T(\mathbf{p}, t) \left[1 + \frac{\mathbf{e}_i \cdot \mathbf{u}}{c c_s^2} \right] \\
 c &= \frac{\Delta p}{\Delta t} \\
 T(\mathbf{p}, t) &= \sum_{i=0}^6 f_i(\mathbf{p}, t)
 \end{aligned}$$

where:

$$\begin{aligned}
 c_s^2 &= \frac{1}{4} \\
 \omega_0 &= \frac{1}{4} \\
 \omega_i &= \frac{1}{8} \text{ for } i \neq 0 \\
 e_0 &= (0, 0, 0) \\
 e_1 &= (1, 0, 0) \\
 e_2 &= (-1, 0, 0) \\
 e_3 &= (0, 1, 0) \\
 e_4 &= (0, 1, 0) \\
 e_5 &= (0, -1, 0) \\
 e_6 &= (0, 0, 1) \\
 e_7 &= (0, 0, -1)
 \end{aligned}$$

A.4 Advection-Diffusion-Reaction Equation

We are interested in the 3D advection-diffusion-reaction equation:

$$\frac{\partial T}{\partial t} + \mathbf{u} \cdot \nabla T = \nabla \cdot (d \nabla T) + S(T)$$

where $S(T) = R(T - T_0)$ in our case but can be any function of T .

In this case the lattice Boltzmann equation is changed to:

$$f_i(\mathbf{p} + \mathbf{e}_i \Delta p, t + \Delta t) = f_i(\mathbf{p}, t) + w [f_i^{eq}(\mathbf{p}, t) - f_i(\mathbf{p}, t)] + \Delta t \omega_i S \quad (\text{A.8})$$

Chapman-Enskog Expansion

The advection-diffusion-reaction equation can be scaled spatially : \mathbf{p} is set to \mathbf{p}/ε , where ε is a small parameter and a dual time scale is introduced:

$$\varepsilon \frac{\partial T}{\partial t_1} + \varepsilon^2 \frac{\partial T}{\partial t_2} + \varepsilon \mathbf{u} \cdot \nabla T = \varepsilon^2 \nabla \cdot (d \nabla T) + S$$

The fonction S should also be expanded in terms of the small parameter ε as:

$$S = S_0 + \varepsilon S_1 + \varepsilon^2 S_2 + O(\varepsilon^3)$$

Introducing this expression in the above equation yields:

$$\varepsilon \frac{\partial T}{\partial t_1} + \varepsilon^2 \frac{\partial T}{\partial t_2} + \varepsilon \mathbf{u} \cdot \nabla T = \varepsilon^2 \nabla \cdot (d \nabla T) + S_0 + \varepsilon S_1 + \varepsilon^2 S_2$$

Clearly, it can be noted that $S_0 = 0$.

For three-dimensional problem with N lattice velocity components ($c_i, i = 0..N$), the temperature can be expressed as:

$$T(x, t) = \sum_{i=0}^N f_i(x, t)$$

Equation A.8 can be expanded using Taylor series:

$$\begin{aligned} f_i(\mathbf{p} + c\mathbf{e}_i \Delta t, t + \Delta t) &= f_i(\mathbf{p}, t) + \frac{\partial f_i}{\partial t} \Delta t + \Delta t c \mathbf{e}_i \cdot \nabla f_i \\ &+ \frac{\Delta t^2}{2} \left(\frac{\partial^2 f_i}{\partial t^2} + 2 \frac{\partial}{\partial t} (c \mathbf{e}_i \cdot \nabla f_i) + c^2 \nabla \nabla : \mathbf{t}_{\mathbf{e}_i \mathbf{e}_i} (f_i) \right) + O(\Delta t)^3 \end{aligned}$$

where the operator $:$ represents tensor contraction.

Introducing scaling, keeping term up to ε^2 we get:

$$\begin{aligned} f_i(\mathbf{p} + c\mathbf{e}_i \Delta t, t + \Delta t) &= f_i(\mathbf{p}, t) + \varepsilon \frac{\partial f_i}{\partial t_1} \Delta t + \varepsilon^2 \frac{\partial f_i}{\partial t_2} \Delta t + \varepsilon c \mathbf{e}_i \cdot \nabla f_i c \Delta t \\ &+ \frac{\Delta t^2}{2} \left(\varepsilon^2 \frac{\partial^2 f_i}{\partial t_1^2} + 2 \varepsilon^2 \frac{\partial}{\partial t_1} (c \mathbf{e}_i \cdot \nabla f_i) + \varepsilon^2 c^2 \nabla \nabla : f_i \mathbf{t}_{\mathbf{e}_i \mathbf{e}_i} \right) + O(\Delta t)^3 \end{aligned}$$

Substituting this equation in equation A.8:

$$\begin{aligned} \varepsilon \frac{\partial f_i}{\partial t_1} \Delta t + \varepsilon^2 \frac{\partial f_i}{\partial t_2} \Delta t + \varepsilon c \mathbf{e}_i \cdot \nabla f_i c \Delta t + \frac{\Delta t^2}{2} \left(\varepsilon^2 \frac{\partial^2 f_i}{\partial t_1^2} + 2 \varepsilon^2 \frac{\partial}{\partial t_1} (c \mathbf{e}_i \cdot \nabla f_i) + \varepsilon^2 c^2 \nabla \nabla : f_i \mathbf{t}_{\mathbf{e}_i \mathbf{e}_i} \right) + O(\Delta t)^3 \\ = w [f_i^{eq}(\mathbf{p}, t) - f_i(\mathbf{p}, t)] + \Delta t \omega_i (\varepsilon S_1 + \varepsilon^2 S_2) \end{aligned}$$

The distribution function can be expanded in terms of the small parameter ε as:

$$f_i(\mathbf{p}, t) = f_i^0(\mathbf{p}, t) + \varepsilon f_i^1(\mathbf{p}, t) + \varepsilon^2 f_i^2(\mathbf{p}, t) + O(\varepsilon^3)$$

Collecting the term of order ε^0 gives:

$$f_i^0(\mathbf{p}, t) = f_i^{eq}(\mathbf{p}, t)$$

Since:

$$\sum_{i=0}^N f_i^{eq}(\mathbf{p}, t) = T(\mathbf{p}, t)$$

Then:

$$\sum_{i=0}^N f_i^0(\mathbf{p}, t) = T(\mathbf{p}, t)$$

As:

$$\sum_{i=0}^N f_i(\mathbf{p}, t) = T(\mathbf{p}, t)$$

This first equation gives:

$$\sum_{i=0}^N f_i^j(\mathbf{p}, t) = 0 \text{ for } j \neq 0$$

Collecting the term of order ε^1 gives:

$$\frac{\partial f_i^0}{\partial t_1} \Delta t + \Delta t c \mathbf{e}_i \cdot \nabla f_i^0 = -w f_i^1 + \Delta t \omega_i S_1 = -\frac{\Delta t}{\tau} f_i^1 + \Delta t \omega_i S_1$$

Hence:

$$\frac{\partial f_i^0}{\partial t_1} + c \mathbf{e}_i \cdot \nabla f_i^0 = -\frac{1}{\tau} f_i^1 + \omega_i S_1$$

By taking the summation over all the values of i , we obtain:

$$\sum_{i=0}^N \frac{\partial f_i^0}{\partial t_1} + \sum_{i=0}^N \mathbf{e}_i \cdot \nabla f_i^0 c = -\frac{1}{\tau} \sum_{i=0}^N f_i^1 + S_1$$

Finally:

$$\sum_{i=0}^N \frac{\partial f_i^0}{\partial t_1} + \nabla \cdot \left(\sum_{i=0}^N c \mathbf{e}_i f_i^0 \right) = S_1 \quad (\text{A.9})$$

Since the first term of right hand side is zero, and according to the momentum equations:

$$\begin{aligned} \sum_{i=0}^N f_i^0 &= T \\ \sum_{i=0}^N f_i^0 \mathbf{c}_i &= T \mathbf{u} \end{aligned}$$

Indeed it is assumed here that the fluid is incompressible:

$$\nabla \cdot \mathbf{u} = 0$$

The above equation formulate as:

$$\frac{\partial T}{\partial t_1} + \mathbf{u} \cdot \nabla T = S_1$$

which is the advection equation.

Latter we will need the derivative of f_i^1 with respect to t_1 and space:

$$-\frac{1}{\tau} \frac{\partial f_i^1}{\partial t_1} = \frac{\partial^2 f_i^0}{\partial t_1^2} + \frac{\partial}{\partial t_1} (\mathbf{c} \mathbf{e}_i \cdot \nabla f_i^0)$$

and:

$$-\frac{1}{\tau} \mathbf{c} \mathbf{e}_i \cdot \nabla f_i^1 = \frac{\partial}{\partial t_1} (\mathbf{c} \mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i \mathbf{t} \mathbf{e}_i \cdot \mathbf{e}_i$$

These two equation gives:

$$-\frac{1}{\tau} \left(\frac{\partial f_i^1}{\partial t_1} + \mathbf{c} \mathbf{e}_i \cdot \nabla f_i^1 \right) = \frac{\partial^2 f_i^0}{\partial t_1^2} + 2 \frac{\partial}{\partial t_1} (\mathbf{c} \mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i \mathbf{t} \mathbf{e}_i \cdot \mathbf{e}_i \quad (\text{A.10})$$

Collecting the term of order ε^2 gives:

$$\frac{\partial f_i^1}{\partial t_1} + \frac{\partial f_i^0}{\partial t_2} + \mathbf{c} \mathbf{e}_i \cdot \nabla f_i^1 + \frac{\Delta t}{2} \left(\frac{\partial^2 f_i^0}{\partial t_1^2} + 2 \frac{\partial}{\partial t_1} (\mathbf{c} \mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i \mathbf{t} \mathbf{e}_i \cdot \mathbf{e}_i \right) = -\frac{1}{\tau} f_i^2 + \omega_i S_2$$

Using equation A.10:

$$\frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial^2 f_i^0}{\partial t_1^2} + 2 \frac{\partial}{\partial t_1} (\mathbf{c} \mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i \mathbf{t} \mathbf{e}_i \cdot \mathbf{e}_i \right) = -\frac{1}{\tau} f_i^2 + \omega_i S_2$$

Hence:

$$\frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial}{\partial t_1} \left(\frac{\partial f_i^0}{\partial t_1} + \mathbf{c} \mathbf{e}_i \cdot \nabla f_i^0 \right) + \frac{\partial}{\partial t_1} (\mathbf{c} \mathbf{e}_i \cdot \nabla f_i^0) + c^2 \nabla \nabla : f_i \mathbf{t} \mathbf{e}_i \cdot \mathbf{e}_i \right) = -\frac{1}{\tau} f_i^2 + \omega_i S_2$$

Summing the above equation over all states:

$$\begin{aligned} \sum_{i=0}^N \frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial}{\partial t_1} \left(\sum_{i=0}^N \frac{\partial f_i^0}{\partial t_1} + c \sum_{i=0}^N \mathbf{e}_i \cdot \nabla f_i^0 \right) + \frac{\partial}{\partial t_1} \left(\sum_{i=0}^N \mathbf{c} \mathbf{e}_i \cdot \nabla f_i^0 \right) + c^2 \nabla \nabla : \sum_{i=0}^N f_i \mathbf{t} \mathbf{e}_i \cdot \mathbf{e}_i \right) \\ = -\frac{1}{\tau} \sum_{i=0}^N f_i^2 + S_2 \end{aligned}$$

The first term of the right hand side is zero and according to equation A.9:

$$\sum_{i=0}^N \frac{\partial f_i^0}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau \right) \left(\frac{\partial S_1}{\partial t_1} + \frac{\partial}{\partial t_1} \left(\sum_{i=0}^N \mathbf{c} \mathbf{e}_i \cdot \nabla f_i^0 \right) + c^2 \nabla \nabla : \sum_{i=0}^N f_i \mathbf{t} \mathbf{e}_i \cdot \mathbf{e}_i \right) = S_2$$

And:

$$\sum_{i=0}^N f_i^0 \mathbf{t}_{\mathbf{e}_i \cdot \mathbf{e}_i} = T c_s^2 \mathbf{I}$$

So we get:

$$\frac{\partial T}{\partial t_2} + \left(\frac{\Delta t}{2} - \tau\right) \left(\frac{\partial S_1}{\partial t_1} + \frac{\partial}{\partial t_1} [\nabla \cdot (\mathbf{u}T)] + c^2 c_s^2 \nabla^2 T\right) = S_2$$

Finally:

$$\frac{\partial T}{\partial t_2} + c^2 c_s^2 \left(\frac{\Delta t}{2} - \tau\right) \nabla^2 T + \left(\frac{\Delta t}{2} - \tau\right) \frac{\partial}{\partial t_1} [\nabla \cdot (\mathbf{u}T) + S_1] = S_2$$

So by adding the first equation times ε and the second one times ε^2 , we get:

$$\frac{\partial T}{\partial t} + \mathbf{u} \cdot \nabla T = c^2 c_s^2 \left(\tau - \frac{\Delta t}{2}\right) \nabla^2 T + \varepsilon \left(\tau - \frac{\Delta t}{2}\right) \frac{\partial}{\partial t_1} [\nabla \cdot (\mathbf{u}T) + \varepsilon S_1] + R$$

There is an error term: $\varepsilon^2 \left(\tau - \frac{\Delta t}{2}\right) \frac{\partial}{\partial t_1} [\nabla \cdot (\mathbf{u}T + S_1)]$. The relaxation parameter can be related to the diffusion coefficient as:

$$d = \left(\tau - \frac{\Delta t}{2}\right) c_s^2 c^2$$

A.5 Boundary Condition

We use Neumann Boundary condition, with an assumption of zero flux through the boundary. Thus, if \mathbf{r} is on the boundary and $\mathbf{r} + c\mathbf{e}_\alpha \Delta t$ is outside the domain, then:

for $\Delta \geq \frac{1}{2}$

$$f_{\bar{\alpha}}(\mathbf{r}, t + \Delta t) = \frac{1}{2\Delta} [\hat{f}_\alpha(\mathbf{r}, t + \Delta t)] + \frac{2\Delta - 1}{2\Delta} [\hat{f}_{\bar{\alpha}}(\mathbf{r}, t + \Delta t)]$$

for $\Delta < \frac{1}{2}$

$$f_{\bar{\alpha}}(\mathbf{r}, t + \Delta t) = 2\Delta [\hat{f}_\alpha(\mathbf{r}, t + \Delta t)] + (1 - 2\Delta) [\hat{f}_\alpha(\mathbf{r} + c\mathbf{e}_{\bar{\alpha}} \Delta t, t + \Delta t)]$$

where $\mathbf{e}_{\bar{\alpha}} = -\mathbf{e}_\alpha$ and $\hat{f}_i(\mathbf{r}, t) = f_i(\mathbf{r}, t) + w[f_i^{eq}(\mathbf{r}, t) - f_i(\mathbf{r}, t)]$

Bibliography

- [Adams 2014] Brian M Adams, WJ Bohnhoff, KR Dalbey, JP Eddy, MS Eldred, DM Gay, K Haskell, Patricia D Hough and LP Swiler. *DAKOTA, a multilevel parallel object-oriented framework for design optimization, parameter estimation, uncertainty quantification, and sensitivity analysis: version 6.2 user's manual*. Sandia National Laboratories, Tech. Rep. SAND2010-2183, 2014. (Cited on pages [71](#) and [72](#).)
- [Alliez 2011] P. Alliez, L. Rineau, S. Tayeb, J. Tournois and M. Yvinec. *3D Mesh Generation*. CGAL User and Reference Manual, vol. 3.9, 2011. (Cited on page [32](#).)
- [Altrogge 2012] Inga Altrogge, Tobias Preusser, Tim Kroger, Sabrina Haase, Torben Patz and Robert Michael Kirby. *Sensitivity analysis for the optimization of radiofrequency ablation in the presence of material parameter uncertainty*. International Journal for Uncertainty Quantification, vol. 2, no. 3, 2012. (Cited on pages [23](#), [24](#), [25](#), [26](#), [73](#) and [104](#).)
- [Audigier 2013a] Chloé Audigier, Tommaso Mansi, Hervé Delingette, Saikiran Rapaka, Viorel Mihalef, Puneet Sharma, Ali Kamen, Daniel Carnegie, Emad Boctor, Michael Choti, Ali Kamen, Dorin Comaniciu and Nicholas Ayache. *Lattice-Boltzmann Method For Fast Patient-Specific Simulation of Liver Tumor Ablation from CT Images*. In Medical Image Computing and Computer-Assisted Intervention—MICCAI 2013, pages 323–330. Springer, 2013. (Cited on pages [5](#), [23](#), [25](#), [29](#), [34](#), [36](#), [53](#), [69](#), [71](#), [97](#) and [143](#).)
- [Audigier 2013b] Chloé Audigier, Tommaso Mansi, Viorel Mihalef, Ali Kamen, Dorin Comaniciu, Puneet Sharma, Saikiran Rapaka, Hervé Delingette, Nicholas Ayache *et al.* *System and Method for Patient Specific Modeling of Liver Tumor Ablation*, November 5 2013. US Patent App. 14/071,688. (Cited on page [144](#).)
- [Audigier 2014a] Chloé Audigier, Tommaso Mansi, Hervé Delingette, Saikiran Rapaka, Viorel Mihalef, Daniel Carnegie, Emad Boctor, Michael Choti, Ali Kamen, Dorin Comaniciu and Nicholas Ayache. *Parameter Estimation For Personalization of Liver Tumor Radiofrequency Ablation*. In MICCAI Workshop on Abdominal Imaging—Computational and Clinical Applications, 2014. (Cited on pages [5](#), [26](#), [67](#), [69](#) and [143](#).)
- [Audigier 2014b] Chloé Audigier, Tommaso Mansi, Viorel Mihalef, Ali Kamen, Dorin Comaniciu, Puneet Sharma and Saikiran Rapaka. *System and method for interactive patient specific simulation of radiofrequency ablation therapy*, September 4 2014. WO Patent App. PCT/US2014/017,886. (Cited on page [144](#).)

- [Audigier 2015a] Chloé Audigier, Tommaso Mansi, Hervé Delingette, Saikiran Rapaka, Viorel Mihalef, Emad Boctor, Michael Choti, Ali Kamen, Nicholas Ayache and Dorin Comaniciu. *Efficient Lattice Boltzmann Solver for Patient-Specific Radiofrequency Ablation of Hepatic Tumors*. IEEE TMI, 2015. (Cited on pages 5, 23, 24, 29, 34, 53, 86 and 143.)
- [Audigier 2015b] Chloé Audigier, Tommaso Mansi, Hervé Delingette, Saikiran Rapaka, Viorel Mihalef, Raoul Pop, Michele Diana, Luc Soler, Ali Kamen, Dorin Comaniciu and Nicholas Ayache. *Challenges to Validate Multi-physics Model of Liver Tumor Radiofrequency Ablation from Pre-clinical Data*. In MICCAI Workshop on Computational Biomechanics for Medicine X, 2015. (Cited on pages 81, 99, 128 and 144.)
- [Audigier 2015c] Chloe Audigier, Tommaso Mansi, Saikiran Rapaka, Ali Kamen, Viorel Mihalef, Herve Delingette, Nicholas Ayache and Dorin Comaniciu. *System and Method for Personalized Computation of Tissue Ablation Extent Based on Medical Images*, August 27 2015. US Patent 20,150,242,588. (Cited on page 144.)
- [Avants 2009] Brian B Avants, Nick Tustison and Gang Song. *Advanced normalization tools (ANTs)*. Insight J, 2009. (Cited on pages 55 and 56.)
- [Baegert 2007] Claire Baegert, Caroline Villard, Pascal Schreck and Luc Soler. *Precise determination of regions of interest for hepatic RFA planning*. In Medical Imaging, pages 650923–650923. International Society for Optics and Photonics, 2007. (Cited on pages 26 and 27.)
- [Berjano 2006] Enrique J Berjano. *Theoretical modeling for radiofrequency ablation: state-of-the-art and challenges for the future*. Biomedical engineering online, vol. 5, no. 1, page 24, 2006. (Cited on pages 3, 13, 20 and 23.)
- [Bilchik 2001] Anton J Bilchik, Thomas F Wood and David P Allegra. *Radiofrequency ablation of unresectable hepatic malignancies: lessons learned*. The Oncologist, vol. 6, no. 1, pages 24–33, 2001. (Cited on page 2.)
- [Bouzidi 2001] M'hamed Bouzidi, Mouaouia Firdaouss and Pierre Lallemand. *Momentum transfer of a Boltzmann-lattice fluid with boundaries*. Physics of Fluids, vol. 13, page 3452, 2001. (Cited on page 44.)
- [Boxenbaum 1980] Harold Boxenbaum. *Interspecies variation in liver weight, hepatic blood flow, and antipyrine intrinsic clearance: extrapolation of data to benzodiazepines and phenytoin*. Journal of pharmacokinetics and biopharmaceutics, vol. 8, no. 2, pages 165–176, 1980. (Cited on page 93.)
- [Breen 2002] MS Breen, X Chen, DL Wilson and GM Saidel. *Modeling cellular thermal damage from radio-frequency ablation*. In EMBS/BMES Conference, 2002., volume 1, pages 715–vol. IEEE, 2002. (Cited on page 39.)

- [Brinkman 1949] HC Brinkman. *A calculation of the viscous force exerted by a flowing fluid on a dense swarm of particles*. Applied Scientific Research, vol. 1, no. 1, pages 27–34, 1949. (Cited on pages 32 and 35.)
- [Bruix 2002] Jordi Bruix and Josep M Llovet. *Prognostic prediction and treatment strategy in hepatocellular carcinoma*. Hepatology, vol. 35, no. 3, pages 519–524, 2002. (Cited on page 1.)
- [Chai 2013] Zhenhua Chai and TS Zhao. *Lattice Boltzmann model for the convection-diffusion equation*. Physical Review E, vol. 87, no. 6, page 063309, 2013. (Cited on page 149.)
- [Chang 2004] Isaac A Chang and Uyen D Nguyen. *Thermal modeling of lesion growth with radiofrequency ablation devices*. Biomedical engineering online, vol. 3, no. 1, page 27, 2004. (Cited on pages 21, 22 and 24.)
- [Chato 1980] JC Chato. *Heat transfer to blood vessels*. Journal of Biomechanical Engineering, vol. 102, no. 2, pages 110–118, 1980. (Cited on page 19.)
- [Chen 1998] Shiyi Chen and Gary D Doolen. *Lattice Boltzmann method for fluid flows*. Annual review of fluid mechanics, vol. 30, no. 1, pages 329–364, 1998. (Cited on pages 30, 36, 42 and 43.)
- [Chen 2009a] Chun-Cheng R Chen, Michael Miga, Robert L Galloway Jr et al. *Optimizing electrode placement using finite-element models in radiofrequency ablation treatment planning*. Biomedical Engineering, IEEE Transactions on, vol. 56, no. 2, pages 237–245, 2009. (Cited on pages 21 and 26.)
- [Chen 2009b] Xin Chen and Gerald M. Saidel. *Mathematical Modeling of Thermal Ablation in Tissue Surrounding a Large Vessel*. J Biomech, vol. 131, 2009. (Cited on pages 23, 24 and 25.)
- [Cheng 2002] Hai-Ling Margaret Cheng and Donald B Plewes. *Tissue thermal conductivity by magnetic resonance thermometry and focused ultrasound heating*. Journal of Magnetic Resonance Imaging, vol. 16, no. 5, pages 598–609, 2002. (Cited on page 65.)
- [Chopard 2009] B Chopard, JL Falcone and J Latt. *The lattice Boltzmann advection-diffusion model revisited*. The European Physical Journal-Special Topics, vol. 171, no. 1, pages 245–249, 2009. (Cited on page 159.)
- [Cornelis 2011] François Cornelis, Nicolas Grenier, Chrit T Moonen and Bruno Quesson. *In vivo characterization of tissue thermal properties of the kidney during local hyperthermia induced by MR-guided high-intensity focused ultrasound*. NMR in Biomedicine, vol. 24, no. 7, pages 799–806, 2011. (Cited on page 27.)

- [Courant 1928] Richard Courant, Kurt Friedrichs and Hans Lewy. *Über die partiellen Differenzgleichungen der mathematischen Physik*. Mathematische Annalen, vol. 100, no. 1, pages 32–74, 1928. (Cited on page 49.)
- [Crezee 1990] J Crezee and JJW Lagendijk. *Experimental verification of bioheat transfer theories: measurement of temperature profiles around large artificial vessels in perfused tissue*. Physics in Medicine and Biology, vol. 35, no. 7, page 905, 1990. (Cited on page 24.)
- [Criminisi 2008] Antonio Criminisi, Toby Sharp and Andrew Blake. *Geos: Geodesic image segmentation*. In Computer Vision–ECCV 2008, pages 99–112. Springer, 2008. (Cited on page 45.)
- [Culjat 2010] Martin O Culjat, David Goldenberg, Priyamvada Tewari and Rahul S Singh. *A review of tissue substitutes for ultrasound imaging*. Ultrasound in medicine & biology, vol. 36, no. 6, pages 861–873, 2010. (Cited on page 84.)
- [Curley 1997] SA Curley, BS Davidson, RY Fleming, F Izzo, LC Stephens, P Tinkey and D Cromeens. *Laparoscopically guided bipolar radiofrequency ablation of areas of porcine liver*. Surgical endoscopy, vol. 11, no. 7, pages 729–733, 1997. (Cited on page 1.)
- [Curley 2003] Steven A Curley. *Radiofrequency ablation of malignant liver tumors*. Annals of Surgical Oncology, vol. 10, no. 4, pages 338–347, 2003. (Cited on page 11.)
- [Dawson 1993] S Ponce Dawson, S Chen and GD Doolen. *Lattice Boltzmann computations for reaction-diffusion equations*. The Journal of Chemical Physics, vol. 98, no. 2, pages 1514–1523, 1993. (Cited on page 42.)
- [De Senneville 2007] Baudouin Denis De Senneville, Charles Mougnot and Chrit TW Moonen. *Real-time adaptive methods for treatment of mobile organs by MRI-controlled high-intensity focused ultrasound*. Magnetic Resonance in Medicine, vol. 57, no. 2, pages 319–330, 2007. (Cited on page 27.)
- [De Senneville 2013] Baudouin Denis De Senneville, Sébastien Roujol, Silke Hey, Chrit Moonen and Mario Ries. *Extended Kalman filtering for continuous volumetric MR-temperature imaging*. Medical Imaging, IEEE Transactions on, vol. 32, no. 4, pages 711–718, 2013. (Cited on page 27.)
- [d’Humières 2002] Dominique d’Humières. *Multiple-relaxation-time lattice Boltzmann models in three dimensions*. Philosophical Transactions of the Royal Society of London. Series A: Mathematical, Physical and Engineering Sciences, vol. 360, no. 1792, pages 437–451, 2002. (Cited on page 43.)
- [Diller 2008] KR Diller, JT Oden, C Bajaj, JC Browne, J Hazle, I Babuska, J Bass, L Bidaut, L Demkowicz, A Elliott *et al.* *Computational infrastructure for the*

- real-time patient-specific treatment of cancer*. Advances in Numerical Heat Transfer, vol. 3, 2008. (Cited on page 20.)
- [Doss 1982] James D Doss. *Calculation of electric fields in conductive media*. Medical physics, vol. 9, no. 4, pages 566–573, 1982. (Cited on page 17.)
- [Dunmire 2013] Barbrina Dunmire, John C Kucewicz, Stuart B Mitchell, Lawrence A Crum and K Michael Sekins. *Characterizing an agar/gelatin phantom for image guided dosing and feedback control of high-intensity focused ultrasound*. Ultrasound in medicine & biology, vol. 39, no. 2, pages 300–311, 2013. (Cited on page 84.)
- [El-Serag 2003] Hashem B. El-Serag, Jessica A. Davila, Nancy J. Petersen and Katherine A. McGlynn. *The Continuing Increase in the Incidence of Hepatocellular Carcinoma in the United States: An Update*. Ann Intern Med, vol. 139, no. 10, pages 817–823, 2003. (Cited on pages 1, 51 and 67.)
- [Faure 2012] François Faure, Christian Duriez, Hervé Delingette, Jérémie Allard, Benjamin Gilles, Stéphanie Marchesseau, Hugo Talbot, Hadrien Courteuisse, Guillaume Bousquet, Igor Peterlik *et al.* *Sofa: A multi-model framework for interactive physical simulation*. In Soft Tissue Biomechanical Modeling for Computer Assisted Surgery, pages 283–321. Springer, 2012. (Cited on page 96.)
- [Feng 2008] Yusheng Feng, J Tinsley Oden and Marissa Nichole Rylander. *A two-state cell damage model under hyperthermic conditions: theory and in vitro experiments*. Journal of biomechanical engineering, vol. 130, no. 4, page 041016, 2008. (Cited on page 22.)
- [Feng 2011] Yusheng Feng and David Fuentes. *Model-based planning and real-time predictive control for laser-induced thermal therapy*. International Journal of Hyperthermia, vol. 27, no. 8, pages 751–761, 2011. (Cited on page 22.)
- [Ferlay 2015] Jacques Ferlay, Isabelle Soerjomataram, Rajesh Dikshit, Sultan Eser, Colin Mathers, Marise Rebelo, Donald Maxwell Parkin, David Forman and Freddie Bray. *Cancer incidence and mortality worldwide: sources, methods and major patterns in GLOBOCAN 2012*. International Journal of Cancer, vol. 136, no. 5, pages E359–E386, 2015. (Cited on page 1.)
- [Garrean 2008] Sean Garrean, Justin Hering, Abdul Saied, W Scott Helton and N Joseph Espot. *Radiofrequency ablation of primary and metastatic liver tumors: a critical review of the literature*. The American Journal of Surgery, vol. 195, no. 4, pages 508–520, 2008. (Cited on page 2.)
- [Guenette 2010] Jeffrey P Guenette and Damian E Dupuy. *Radiofrequency ablation of colorectal hepatic metastases*. Journal of surgical oncology, vol. 102, no. 8, pages 978–987, 2010. (Cited on page 14.)

- [Guetter 2011] Christoph Guetter, Hui Xue, Christophe Chedf'Hotel and Jens Guehring. *Efficient symmetric and inverse-consistent deformable registration through interleaved optimization*. In Biomedical Imaging: From Nano to Macro, 2011 IEEE International Symposium on, pages 590–593. IEEE, 2011. (Cited on page 93.)
- [Gulsun 2006] M. A. Gulsun and H Tek. *3D Modeling of Coronary Arteries*. In Computer Vision for Intravascular and Intracardiac Imaging, MICCAI Workshop, 2006. (Cited on page 93.)
- [Guo 2002] Zhaoli Guo and TS Zhao. *Lattice-Boltzmann model for incompressible flows through porous media*. Physical Review E, vol. 66, no. 3, page 036304, 2002. (Cited on pages 35 and 36.)
- [Hall 2015] Sheldon K Hall, Ean H Ooi and Stephen J Payne. *Cell death, perfusion and electrical parameters are critical in models of hepatic radiofrequency ablation*. International Journal of Hyperthermia, no. ahead-of-print, pages 1–13, 2015. (Cited on pages 23, 25 and 26.)
- [Henriques Jr 1947] FC Henriques Jr and AR Moritz. *Studies of thermal injury: I. The conduction of heat to and through skin and the temperatures attained therein. A theoretical and an experimental investigation*. The American journal of pathology, vol. 23, no. 4, page 530, 1947. (Cited on page 21.)
- [Hildebrand 2006] P Hildebrand, T Leibecke, M Kleemann, L Mirow, M Birth, HP Bruch and C Bürk. *Influence of operator experience in radiofrequency ablation of malignant liver tumours on treatment outcome*. European Journal of Surgical Oncology (EJSO), vol. 32, no. 4, pages 430–434, 2006. (Cited on page 2.)
- [Horng 2007] Tzyy-Leng Horng, Win-Li Lin, Chihng-Tsung Liauh and Tzu-Ching Shih. *Effects of pulsatile blood flow in large vessels on thermal dose distribution during thermal therapy*. Medical physics, vol. 34, no. 4, pages 1312–1320, 2007. (Cited on page 94.)
- [Ishihara 1995] Yasutoshi Ishihara, Arturo Calderon, Hidehiro Watanabe, Kazuya Okamoto, Yoshinori Suzuki, Kagayaki Kuroda and Yutaka Suzuki. *A precise and fast temperature mapping using water proton chemical shift*. Magnetic Resonance in Medicine, vol. 34, no. 6, pages 814–823, 1995. (Cited on page 27.)
- [Jain 2000] Mudit K Jain and Patrick D Wolf. *A three-dimensional finite element model of radiofrequency ablation with blood flow and its experimental validation*. Annals of biomedical engineering, vol. 28, no. 9, pages 1075–1084, 2000. (Cited on pages 24 and 25.)
- [Jiang 2010] Y. Jiang, S. Mulier, W. Chong, M.C. Diel Rambo, F. Chen, G. Marchal and Y. Ni. *Formulation of 3D finite elements for hepatic radiofrequency*

- ablation. *IJMIC*, vol. 9, no. 3, pages 225–235, 2010. (Cited on pages 23, 24 and 25.)
- [Khaled 2003] A-Ra Khaled and K Vafai. *The role of porous media in modeling flow and heat transfer in biological tissues*. Internal journal of heat and mass transfer, vol. 46, no. 26, pages 4989–5003, 2003. (Cited on page 19.)
- [Khlebnikov 2010] Rostislav Khlebnikov and Judith Muehl. *Effects of needle placement inaccuracies in hepatic radiofrequency tumor ablation*. In Engineering in Medicine and Biology Society (EMBC), 2010 Annual International Conference of the IEEE, pages 716–721. IEEE, 2010. (Cited on page 18.)
- [Kim 2006] Young-sun Kim, Hyunchul Rhim, On Koo Cho, Byung Hee Koh and Yongsoo Kim. *Intrahepatic recurrence after percutaneous radiofrequency ablation of hepatocellular carcinoma: analysis of the pattern and risk factors*. European journal of radiology, vol. 59, no. 3, pages 432–441, 2006. (Cited on pages 2 and 12.)
- [Klinger 1974] H.G. Klinger. *Heat transfer in perfused biological tissue I: General theory*. B Math Biol, vol. 36, no. 0, pages 403 – 415, 1974. (Cited on page 20.)
- [Kröger 2010] T. Kröger, T. Pätz, I. Altrogge, A. Schenk, KS Lehmann, BB Frericks, J.P. Ritz, H.O. Peitgen and T. Preusser. *Fast estimation of the vascular cooling in RFA based on numerical simulation*. Open Biomed Eng J, vol. 4, pages 16–26, 2010. (Cited on page 23.)
- [Labonté 1994] Sylvain Labonté. *Numerical model for radio-frequency ablation of the endocardium and its experimental validation*. Biomedical Engineering, IEEE Transactions on, vol. 41, no. 2, pages 108–115, 1994. (Cited on pages 19 and 24.)
- [Lehmann 2009] Kai S Lehmann, JP Ritz, Steffi Valdeig, Verena Knappe, Andrea Schenk, A Weihusen, C Rieder, C Holmer, U Zurbuchen, P Hoffmann *et al.* *Ex situ quantification of the cooling effect of liver vessels on radiofrequency ablation*. Langenbeck’s Archives of Surgery, vol. 394, no. 3, pages 475–481, 2009. (Cited on page 24.)
- [Lencioni 2005] R Lencioni, C Della Pina and C Bartolozzi. *Percutaneous image-guided radiofrequency ablation in the therapeutic management of hepatocellular carcinoma*. Abdominal imaging, vol. 30, no. 4, pages 401–408, 2005. (Cited on page 12.)
- [Lepetit-Coiffé 2010] Matthieu Lepetit-Coiffé, Hervé Laumonier, Olivier Seror, Bruno Quesson, Musa-Bahazid Sesay, Chrit TW Moonen, Nicolas Grenier and Hervé Trillaud. *Real-time monitoring of radiofrequency ablation of liver tumors using thermal-dose calculation by MR temperature imaging: initial results in nine patients, including follow-up*. European radiology, vol. 20, no. 1, pages 193–201, 2010. (Cited on page 13.)

- [Linte 2013a] Cristian A Linte, Jon J Camp, David R Holmes III, Maryam E Rettmann and Richard A Robb. *Modeling of Radiofrequency Ablation Lesions for Image-Guided Arrhythmia Therapy: A Preliminary ex vivo Demonstration*. In *Augmented Environments for Computer-Assisted Interventions*, pages 22–33. Springer, 2013. (Cited on page 25.)
- [Linte 2013b] Cristian A Linte, Jon J Camp, David R Holmes III, Maryam E Rettmann and Richard A Robb. *Toward Online Modeling for Lesion Visualization and Monitoring in Cardiac Ablation Therapy*. In *Medical Image Computing and Computer-Assisted Intervention—MICCAI 2013*, pages 9–17. Springer, 2013. (Cited on pages 20 and 21.)
- [Livraghi 1999] Tito Livraghi, S Nahum Goldberg, Sergio Lazzaroni, Franca Meloni, Luigi Solbiati and G Scott Gazelle. *Small hepatocellular carcinoma: treatment with radio-frequency ablation versus ethanol injection*. *Radiology*, vol. 210, no. 3, pages 655–661, 1999. (Cited on page 109.)
- [Lubner 2010] Meghan G Lubner, Christopher L Brace, J Louis Hinshaw and Fred T Lee. *Microwave tumor ablation: mechanism of action, clinical results, and devices*. *Journal of Vascular and Interventional Radiology*, vol. 21, no. 8, pages S192–S203, 2010. (Cited on page 15.)
- [Mahjoob 2009] Shabi Mahjoob and Kambiz Vafai. *Analytical characterization and production of an isothermal surface for biological and electronic applications*. *Journal of Heat Transfer*, vol. 131, no. 5, page 52604, 2009. (Cited on page 19.)
- [Mai 2010] Hao-Chueh Mai, Kuen-Hau Lin, Cheng-Hsiu Yang and Chao-An Lin. *A thermal lattice Boltzmann model for flows with viscous heat dissipation*. *Computer Modeling in Engineering & Sciences(CMES)*, vol. 61, no. 1, pages 45–62, 2010. (Cited on page 41.)
- [Marchesseau 2013] Stéphanie Marchesseau, Hervé Delingette, Maxime Sermesant, R Cabrera-Lozoya, Catalina Tobon-Gomez, Philippe Moireau, RM Figueras i Ventura, Karim Lekadir, Alfredo Hernandez, Mireille Garreau *et al.* *Personalization of a cardiac electromechanical model using reduced order unscented Kalman filtering from regional volumes*. *Medical image analysis*, vol. 17, no. 7, pages 816–829, 2013. (Cited on page 25.)
- [Minami 2011] Yasunori Minami and Masatoshi Kudo. *Radiofrequency ablation of hepatocellular carcinoma: a literature review*. *International journal of hepatology*, vol. 2011, 2011. (Cited on page 11.)
- [Mitchell 1968] John W Mitchell and Glen E Myers. *An analytical model of the counter-current heat exchange phenomena*. *Biophysical journal*, vol. 8, no. 8, pages 897–911, 1968. (Cited on page 19.)

- [Mohamad 2011] Abdulmajeed A Mohamad. Lattice boltzmann method: fundamentals and engineering applications with computer codes. Springer Science & Business Media, 2011. (Cited on page 147.)
- [Moritz 1947] A Re Moritz and FC Henriques Jr. *Studies of Thermal Injury: II. The Relative Importance of Time and Surface Temperature in the Causation of Cutaneous Burns*. The American journal of pathology, vol. 23, no. 5, page 695, 1947. (Cited on page 21.)
- [Nakayama 2001] Akira Nakayama, Fujio Kuwahara, Masazumi Sugiyama and Guoliang Xu. *A two-energy equation model for conduction and convection in porous media*. International Journal of Heat and Mass Transfer, vol. 44, no. 22, pages 4375 – 4379, 2001. (Cited on page 19.)
- [Nakayama 2008] A. Nakayama and F. Kuwahara. *A general bioheat transfer model based on the theory of porous media*. International Journal of Heat and Mass Transfer, vol. 51, no. 11-12, pages 3190 – 3199, 2008. (Cited on pages 18 and 19.)
- [Narayanan 2013] Govindarajan Narayanan, Tatiana Froud, Rekhaven Suthar and Katuska Barbery. *Irreversible electroporation of hepatic malignancy*. In Seminars in interventional radiology, volume 30, page 67. Thieme Medical Publishers, 2013. (Cited on page 17.)
- [Neumann 1982] Eberhard Neumann, M Schaefer-Ridder, Y Wang and PH Hofschneider. *Gene transfer into mouse lyoma cells by electroporation in high electric fields*. The EMBO journal, vol. 1, no. 7, page 841, 1982. (Cited on page 16.)
- [ONEill 2011] David ONEill, Tingying Peng, Philipp Stiegler, Ursula Mayrhauser, Sonja Koestenbauer, Karlheinz Tscheliessnigg and Stephen Payne. *A Three-State Mathematical Model of Hyperthermic Cell Death*. Ann Biomed Eng, vol. 39, pages 570–579, 2011. (Cited on pages 20, 21, 22, 39, 40, 41, 53, 55, 70, 74 and 86.)
- [Pan 2006] Chongxun Pan, Li-Shi Luo and Cass T Miller. *An evaluation of lattice Boltzmann schemes for porous medium flow simulation*. Computers & fluids, vol. 35, no. 8, pages 898–909, 2006. (Cited on page 36.)
- [Panescu 1995] Dorin Panescu, James G Wayne, Sidney D Fleischman, Mark S Mirotznik, David K Swanson and John G Webster. *Three-dimensional finite element analysis of current density and temperature distributions during radio-frequency ablation*. Biomedical Engineering, IEEE Transactions on, vol. 42, no. 9, pages 879–890, 1995. (Cited on pages 17, 18, 20 and 24.)
- [Patterson 1998] Emma J Patterson, Charles H Scudamore, David A Owen, Alex G Nagy and Andrzej K Buczkowski. *Radiofrequency ablation of porcine liver*

- in vivo: effects of blood flow and treatment time on lesion size.* Annals of surgery, vol. 227, no. 4, page 559, 1998. (Cited on page 24.)
- [Pätz 2010] Torben Pätz, Tim Kröger and Tobias Preusser. *Simulation of radiofrequency ablation including water evaporation.* In World Congress on Medical Physics and Biomedical Engineering, September 7-12, 2009, Munich, Germany, pages 1287–1290. Springer, 2010. (Cited on page 24.)
- [Payne 2010] Stephen J Payne, T Peng and DP O’Neill. *Mathematical modeling of thermal ablation.* Critical Reviews in Biomedical Engineering, vol. 38, no. 1, 2010. (Cited on pages 23 and 25.)
- [Payne 2011] Stephen Payne, Ronan Flanagan, Mika Pollari, Tuomas Alhonorro, Claire Bost, David O’Neill, Tingying Peng and Philipp Stiegler. *Image-based multi-scale modelling and validation of radio-frequency ablation in liver tumours.* Philos T Roy Soc A, vol. 369, no. 1954, pages 4233–4254, 2011. (Cited on pages 5, 18, 19, 23, 24, 25, 29, 35, 40 and 95.)
- [Pearce 1995] John Pearce and Sharon Thomsen. *Rate process analysis of thermal damage.* In Optical-thermal response of laser-irradiated tissue, pages 561–606. Springer, 1995. (Cited on page 21.)
- [Peng 2011] T Peng, DP O’Neill and SJ Payne. *A two-equation coupled system for determination of liver tissue temperature during thermal ablation.* International Journal of Heat and Mass Transfer, vol. 54, no. 9, pages 2100–2109, 2011. (Cited on pages 20, 39 and 51.)
- [Pennes 1948] Harry H Pennes. *Analysis of tissue and arterial blood temperatures in the resting human forearm.* Journal of applied physiology, vol. 1, no. 2, pages 93–122, 1948. (Cited on pages 19, 24 and 40.)
- [Podhajsky 2009] Ronald J Podhajsky, Ming Yi and Roop L Mahajan. *Differential and directional effects of perfusion on electrical and thermal conductivities in liver.* In Engineering in Medicine and Biology Society, 2009. EMBC 2009. Annual International Conference of the IEEE, pages 4295–4298. IEEE, 2009. (Cited on page 65.)
- [Pompili 2010] Maurizio Pompili, Claudio Maurizio Pacella, Giampiero Francica, Mario Angelico, Giuseppe Tisone, Paolo Craboledda, Erica Nicolardi, Gian Ludovico Rapaccini and Giovanni Gasbarrini. *Percutaneous laser ablation of hepatocellular carcinoma in patients with liver cirrhosis awaiting liver transplantation.* European journal of radiology, vol. 74, no. 3, pages e6–e11, 2010. (Cited on page 16.)
- [Pop 2010] Mihaela Pop, Sean RH Davidson, Mark Gertner, Michael AS Jewett, Michael D Sherar and Michael C Kolios. *A theoretical model for RF ablation of kidney tissue and its experimental validation.* In Biomedical Simulation, pages 119–129. Springer, 2010. (Cited on page 21.)

- [Quesson 2000] Bruno Quesson, Jacco A de Zwart and Chrit TW Moonen. *Magnetic resonance temperature imaging for guidance of thermotherapy*. *Journal of Magnetic Resonance Imaging*, vol. 12, no. 4, pages 525–533, 2000. (Cited on page 27.)
- [Ralovich 2012] K. Ralovich, L. Itu, V. Mihalef, P. Sharma, R. Ionasec, D. Vitinovski, W. Krawtschuk, A. Everett, R. Ringel and N. Navab. *Hemodynamic Assessment of Pre- and Post-operative Aortic Coarctation from MRI*. In MICCAI 2012, volume 7511 of *LNCS*, pages 486–493. Springer, 2012. (Cited on page 32.)
- [Rapaka 2012] S Rapaka, T Mansi, B Georgescu, M Pop, G Wright, A Kamen and Dorin Comaniciu. *LBM-EP: Lattice-Boltzmann Method for Fast Cardiac Electrophysiology Simulation from 3D Images*. In MICCAI 2012, volume 7511 of *LNCS*, pages 33–40, 2012. (Cited on pages 23, 42, 43 and 44.)
- [Relan 2011] Jatin Relan, Phani Chinchapatnam, Maxime Sermesant, Kawal Rhode, Matt Ginks, Hervé Delingette, C Aldo Rinaldi, Reza Razavi and Nicholas Ayache. *Coupled personalization of cardiac electrophysiology models for prediction of ischaemic ventricular tachycardia*. *Interface Focus*, page rsfs20100041, 2011. (Cited on page 25.)
- [Rieder 2011] Christian Rieder, Tim Kroeger, Christian Schumann and Horst K Hahn. *GPU-based real-time approximation of the ablation zone for radiofrequency ablation*. *Visualization and Computer Graphics, IEEE Transactions on*, vol. 17, no. 12, pages 1812–1821, 2011. (Cited on page 23.)
- [Rieder 2013] Christian Rieder. *Interactive Visualization for Assistance of Needle-Based Interventions*. PhD thesis, Bremen, Jacobs Univ., Diss., 2013, 2013. (Cited on page 2.)
- [Roetzel 1998] Wilfried Roetzel and Yimin Xuan. *Transient response of the human limb to an external stimulist*. *Internal journal of heat and mass transfer*, vol. 41, no. 1, pages 229–239, 1998. (Cited on page 19.)
- [Sapareto 1984] Stephen A Sapareto and William C Dewey. *Thermal dose determination in cancer therapy*. *International Journal of Radiation Oncology* Biology* Physics*, vol. 10, no. 6, pages 787–800, 1984. (Cited on pages 13, 21 and 125.)
- [Schenk Jr 1962] Worthington G Schenk Jr, John C McDonald, Khlar McDonald and Theodore Drapanas. *Direct measurement of hepatic blood flow in surgical patients: with related observations on hepatic flow dynamics in experimental animals*. *Ann Surg*, vol. 156, no. 3, pages 463–469, 1962. (Cited on page 66.)
- [Schramm 2006] Wolfgang Schramm, Deshan Yang and D Haemmerich. *Contribution of direct heating, thermal conduction and perfusion during radiofrequency and microwave ablation*. In *Engineering in Medicine and Biology*

- Society, 2006. EMBS'06. 28th Annual International Conference of the IEEE, pages 5013–5016. IEEE, 2006. (Cited on page 18.)
- [Schumann 2015] Christian Schumann, Christian Rieder, Sabrina Haase, Katrin Teichert, Philipp Süß, Peter Isfort, Philipp Bruners and Tobias Preusser. *Interactive multi-criteria planning for radiofrequency ablation*. International journal of computer assisted radiology and surgery, pages 1–11, 2015. (Cited on pages 22, 23 and 26.)
- [Schwarzmaier 1998] Hans-Joachim Schwarzmaier, Ilya V Yaroslavsky, Anna N Yaroslavsky, Volkhard Fiedler, Frank Ulrich and Thomas Kahn. *Treatment planning for MRI-guided laser-induced interstitial thermotherapy of brain tumors - The role of blood perfusion*. Journal of Magnetic Resonance Imaging, vol. 8, no. 1, pages 121–127, 1998. (Cited on page 21.)
- [Seitel 2011] Alexander Seitel, Markus Engel, Christof M Sommer, Boris A Radeleff, Caroline Essert-Villard, Claire Baegert, Markus Fangerau, Klaus H Fritzsche, Kwong Yung, Hans-Peter Meinzer *et al.* *Computer-assisted trajectory planning for percutaneous needle insertions*. Medical physics, vol. 38, no. 6, pages 3246–3259, 2011. (Cited on pages 26 and 27.)
- [Seror 2008] Olivier Seror, Matthieu Lepetit-Coiffé, Brigitte Le Bail, Baudouin Denis De Senneville, Hervé Trillaud, Chrit Moonen and Bruno Quesson. *Real time monitoring of radiofrequency ablation based on MR thermometry and thermal dose in the pig liver in vivo*. European radiology, vol. 18, no. 2, pages 408–416, 2008. (Cited on page 14.)
- [Shrivastava 2009] Devashish Shrivastava and J Thomas Vaughan. *A generic bio-heat transfer thermal model for a perfused tissue*. Journal of biomechanical engineering, vol. 131, no. 7, page 074506, 2009. (Cited on page 19.)
- [Talbot 2014] Hugo Talbot, Myriam Lekkal, Rémi BÉssard-Duparc, Stéphane Cotinet *et al.* *Interactive Planning of Cryotherapy Using Physically-Based Simulation*. In MMVR 21-Medicine Meets Virtual Reality-2014, 2014. (Cited on pages 16 and 20.)
- [Tiesler 2011] Hanne Tiesler. *Identification of Material Parameters from Temperature Measurements in Radio Frequency Ablation*. PhD thesis, PhD thesis, University of Bremen, 2011, <http://elib.suub.uni-bremen.de/peid>, 2011. (Cited on page 63.)
- [Tungjitkusolmun 2002] Supan Tungjitkusolmun, S Tyler Staelin, Dieter Haemmerich, Jang-Zern Tsai, Hong Cao, John G Webster, Fred T Lee Jr, David M Mahvi and Vicken R Vorperian. *Three-dimensional finite-element analyses for radio-frequency hepatic tumor ablation*. Biomedical Engineering, IEEE Transactions on, vol. 49, no. 1, pages 3–9, 2002. (Cited on pages 17, 23 and 24.)

- [Van Beers 2001] Bernard E Van Beers, Isabelle Leconte, Roland Materne, Anne M Smith, Jacques Jamart and Yves Horsmans. *Hepatic perfusion parameters in chronic liver disease: dynamic CT measurements correlated with disease severity*. American Journal of Roentgenology, vol. 176, no. 3, pages 667–673, 2001. (Cited on page 57.)
- [Villard 2003] Caroline Villard, Luc Soler, Nicolas Papier, Vincent Agnus, Afshin Gangi, Didier Mutter and Jacques Marescaux. *RF-Sim: a treatment planning tool for radiofrequency ablation of hepatic tumors*. In Information Visualization, 2003. IV 2003. Proceedings. Seventh International Conference on, pages 561–566. IEEE, 2003. (Cited on page 26.)
- [Villard 2004] Caroline Villard, Luc Soler, Afshin Gangi, Didier Mutter and Jacques Marescaux. *Toward realistic radiofrequency ablation of hepatic tumors 3D simulation and planning*. In Medical Imaging 2004, pages 586–595. International Society for Optics and Photonics, 2004. (Cited on page 26.)
- [Villard 2005] Caroline Villard, Claire Baegert, Pascal Schreck, Luc Soler and Afshin Gangi. *Optimal trajectories computation within regions of interest for hepatic RFA planning*. In Medical Image Computing and Computer-Assisted Intervention—MICCAI 2005, pages 49–56. Springer, 2005. (Cited on page 26.)
- [Weiss 2013] Matthew J Weiss and Christopher L Wolfgang. *Irreversible electroporation: A novel pancreatic cancer therapy*. Current problems in cancer, vol. 37, no. 5, pages 262–265, 2013. (Cited on page 16.)
- [Wulff 1974] W Wulff. *The energy conservation equation for living tissue*. Biomedical Engineering, IEEE Transactions on, no. 6, pages 494–495, 1974. (Cited on page 20.)
- [Yang 2007] Deshan Yang, Mark C Converse, David M Mahvi and John G Webster. *Expanding the bioheat equation to include tissue internal water evaporation during heating*. Biomedical Engineering, IEEE Transactions on, vol. 54, no. 8, pages 1382–1388, 2007. (Cited on page 24.)
- [Yoshida 2010] Hiroaki Yoshida and Makoto Nagaoka. *Multiple-relaxation-time lattice Boltzmann model for the convection and anisotropic diffusion equation*. J Comput Phys, vol. 229, no. 20, pages 7774 – 7795, 2010. (Cited on pages 42, 43 and 44.)
- [Yushkevich 2006] Paul A. Yushkevich, Joseph Piven, Heather Cody Hazlett, Rachel Gimpel Smith, Sean Ho, James C. Gee and Guido Gerig. *User-Guided 3D Active Contour Segmentation of Anatomical Structures: Significantly Improved Efficiency and Reliability*. Neuroimage, vol. 31, no. 3, pages 1116–1128, 2006. (Cited on page 45.)
- [Zhang 2006] CX Zhang, S Zhang, Z Zhang and YZ Chen. *Effects of large blood vessel locations during high intensity focused ultrasound therapy for hepatic*

tumors: a finite element study. In Engineering in Medicine and Biology Society, 2005. IEEE-EMBS 2005. 27th Annual International Conference of the, pages 209–212. IEEE, 2006. (Cited on page 15.)

[Zheng 2007] Yefeng Zheng, Adrian Barbu, Bogdan Georgescu, Michael Scheuering and Dorin Comaniciu. *Fast automatic heart chamber segmentation from 3D CT data using marginal space learning and steerable features*. In Computer Vision, 2007. ICCV 2007. IEEE 11th International Conference on, pages 1–8. IEEE, 2007. (Cited on pages 71 and 112.)

Computational Modeling of Radiofrequency Ablation for the Planning and Guidance of Abdominal Tumor Treatment

Abstract: The outcome of radiofrequency ablation (RFA) treatment for abdominal tumors is challenged by lack of patient-specific planning. In particular, the presence of blood vessels, time- and spatial-varying thermal conductivity makes the prediction of the extent of ablated tissue difficult. This results in possible incomplete treatments and increased risks of recurrence and may compromise the overall outcome. By providing predictive tools, biophysical models can help clinicians to plan and guide the procedure for an effective treatment. For such models to be useful in clinical routine, the computation has to be fast and accurate.

We first present a detailed computational model of the biophysical mechanisms involved in RFA of hepatic tumors such as heat diffusion, cellular necrosis and hepatic blood flow. This model simulates the extent of ablated tissue given the probe location and some biological parameters such as tissue conductivity or heat capacity. The model is based on medical images, from which patient-specific models of the liver, visible vessels and tumors are build. A new approach for solving these partial differential equations based the Lattice Boltzmann Method is introduced. The model is implemented with a material speed-up thanks to the graphics card to reach near real-time computation. The resulting simulations are thoroughly verified against an analytical solution.

In a second step, we aim at validating the model. It is first evaluated against clinical data of patients who underwent RFA of liver tumors. The importance of taking into account the blood flow and of estimating the parameters is highlighted. Then, a comprehensive pre-clinical experiment combining multi-modal, pre- and post-operative anatomical and functional images, as well as the interventional monitoring of the temperature and delivered power is presented. This enables an end-to-end pre-clinical validation framework that considers the most comprehensive data set for model validation.

Then, we automatically estimate patient-specific parameters using inverse problem algorithms, to better predict the ablated tissue. This personalization strategy has been tested and evaluated on seven ablations from three clinical cases. From the pre-clinical study, we can go further in the personalization by comparing the simulated temperature and delivered power with the actual temperature and delivered power measured during the procedure. This second personalization strategy has been tested and evaluated on twelve ablations from five pre-clinical cases. These contributions have led to promising results, and open new perspectives in RFA guidance and planning.
